

ABX Minotrol 16



LOT MX 091 L
 Rev 1

CONTROL L

(Exp.) 2011-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							1.9	2.0	1.9	± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							2.38	2.30	2.38	± 0.15
HB HGB	g/dl							5.9	5.9	5.9	± 0.4
	g/l							59	59	59	± 4
	mmol/l							3.66	3.66	3.66	± 0.25
HT HCT	%							16.4	16.6	16.4	± 2.0
	l/l							0.164	0.166	0.164	± 0.020
VGM MCV	µm ³ ; fl							69	72	69	± 4
TGMH MCH	pg							24.8	25.7	24.8	± 2.0
	fmol							1.54	1.59	1.54	± 0.12
CCMH MCHC	g/dl							35.9	35.6	35.9	± 3.0
	g/l							359	356	359	± 30
	mmol/l							22.31	22.13	22.31	± 1.86
IDR RDW	%							14.8	14.5	16.0	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 73	* 70	73	± 20 ± 22
VPM MPV	µm ³ ; fl							9.3	9.5	9.2	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 7.0
MONO	#							N/A	N/A	N/A	± 0.20
	%							N/A	N/A	N/A	± 6.0
GRANULO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.15 BACK / VERSO 9930046-A

ABX Minotrol 16



LOT MX 091 N
 Rev 1

CONTROL N

(Exp.) 2011-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							7.7	7.7	7.9	± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							4.61	4.62	4.65	± 0.18
HB HGB	g/dl							13.5	13.5	13.5	± 0.5
	g/l							135	135	135	± 5
	mmol/l							8.38	8.38	8.38	± 0.31
HT HCT	%							37.3	37.9	37.7	± 2.5
	l/l							0.373	0.379	0.377	± 0.025
VGM MCV	µm ³ ; fl							81	82	81	± 4
TGMH MCH	pg							29.3	29.2	29.0	± 2.0
	f mol							1.82	1.81	1.80	± 0.12
CCMH MCHC	g/dl							36.2	35.6	35.8	± 3.0
	g/l							362	356	358	± 30
	mmol/l							22.45	22.13	22.26	± 1.86
IDR RDW	%							13.3	14.0	15.0	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 250	* 252	250	± 40 *± 40
VPM MPV	µm ³ ; fl							8.4	8.4	8.3	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 0.50
	%							N/A	N/A	N/A	± 6.0
MONO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 5.0
GRANULO	#							N/A	N/A	N/A	± 0.60
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.15 BACK / VERSO 9930047-A



ABX Minotrol 16



LOT MX 091 H
 Rev 1

CONTROL H

(Exp.) 2011-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX 120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							20.8	20.6	20.7	± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							5.61	5.64	5.72	± 0.20
HB HGB	g/dl							18.0	17.9	17.8	± 0.6
	g/l							180	179	178	± 6
	mmol/l							11.18	11.12	11.05	± 0.37
HT HCT	%							48.8	48.5	49.2	± 3.0
	l/l							0.488	0.485	0.492	± 0.030
VGM MCV	µm ³ ; fl							87	86	86	± 4
TGMH MCH	pg							32.1	31.7	31.1	± 2.0
	f mol							1.99	1.97	1.93	± 0.12
CCMH MCHC	g/dl							36.9	36.9	36.2	± 3.0
	g/l							369	369	362	± 30
	mmol/l							22.90	22.92	22.47	± 1.86
IDR RDW	%							12.5	13.1	14.3	± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l							* 515	* 524	515	± 55 *± 55
VPM MPV	µm ³ ; fl							8.4	8.4	8.3	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 1.00
	%							N/A	N/A	N/A	± 5.0
MONO	#							N/A	N/A	N/A	± 0.80
	%							N/A	N/A	N/A	± 4.0
GRANULO	#							N/A	N/A	N/A	± 1.40
	%							N/A	N/A	N/A	± 7.0

Ref: FORM-0830 Rev.15 BACK / VERSO 9930048-A

* For ABX Pentra 60 Pentra 80 instruments only