

## ABX Minotrol 16



**LOT**

**MX**

Rev

**CONTROL** **L**



(Exp.) **2012-04-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l							<b>1.9</b>	<b>1.9</b>	<b>2.0</b>	± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l							<b>2.32</b>	<b>2.29</b>	<b>2.32</b>	± 0.15
HB HGB	g/dl							<b>5.8</b>	<b>5.8</b>	<b>6.0</b>	± 0.4
	g/l							<b>58</b>	<b>58</b>	<b>60</b>	± 4
	mmol/l							<b>3.60</b>	<b>3.60</b>	<b>3.73</b>	± 0.25
HT HCT	%							<b>16.2</b>	<b>16.7</b>	<b>16.5</b>	± 2.0
	l/l							<b>0.162</b>	<b>0.167</b>	<b>0.165</b>	± 0.020
VGM MCV	µm <sup>3</sup> ; fl							<b>70</b>	<b>73</b>	<b>71</b>	± 4
TGMH MCH	pg							<b>25.0</b>	<b>25.3</b>	<b>25.9</b>	± 2.0
	f mol							<b>1.55</b>	<b>1.57</b>	<b>1.61</b>	± 0.12
CCMH MCHC	g/dl							<b>35.7</b>	<b>34.7</b>	<b>36.4</b>	± 3.0
	g/l							<b>357</b>	<b>347</b>	<b>364</b>	± 30
	mmol/l							<b>22.18</b>	<b>21.55</b>	<b>22.62</b>	± 1.86
IDR RDW	%							<b>14.5</b>	<b>14.5</b>	<b>16.5</b>	± 3.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l							<b>* 80</b>	<b>* 75</b>	<b>75</b>	± 25   ± 28
VPM MPV	µm <sup>3</sup> ; fl							<b>9.5</b>	<b>9.5</b>	<b>9.2</b>	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 9.0
MONO	#							N/A	N/A	N/A	± 0.20
	%							N/A	N/A	N/A	± 7.0
GRANULO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 8.0

Ref. FORM-0830 Rev.16 BACK / VERSO 9930054-A

\* For ABX Pentra 60 Pentra 80 instruments only

## ABX Minotrol 16



**LOT** MX 012 N  
 Rev 1

**CONTROL** N

(Exp.) 2012-04-05  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l							7.3	7.3	7.7	± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l							4.76	4.80	4.79	± 0.18
HB HGB	g/dl							13.5	13.5	13.6	± 0.5
	g/l							135	135	136	± 5
	mmol/l							8.38	8.38	8.45	± 0.31
HT HCT	%							38.1	38.4	38.3	± 2.5
	l/l							0.381	0.384	0.383	± 0.025
VGM MCV	µm <sup>3</sup> ; fl							80	80	80	± 4
TGMH MCH	pg							28.4	28.1	28.4	± 2.0
	f mol							1.76	1.75	1.76	± 0.12
CCMH MCHC	g/dl							35.5	35.2	35.5	± 3.0
	g/l							355	352	355	± 30
	mmol/l							22.02	21.83	22.04	± 1.86
IDR RDW	%							13.5	14.0	16.0	± 3.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l							* 255	* 255	257	± 40   *± 40
VPM MPV	µm <sup>3</sup> ; fl							8.7	8.7	8.4	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 0.50
	%							N/A	N/A	N/A	± 6.0
MONO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 5.0
GRANULO	#							N/A	N/A	N/A	± 0.60
	%							N/A	N/A	N/A	± 7.0

Ref: FORM-0830 Rev.16 BACK / VERSO 9930055-A

\* For ABX Pentra 60 Pentra 80 instruments only

## ABX Minotrol 16



**LOT** MX 012 H

**CONTROL** H

(Exp.) 2012-04-05  
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l							20.4	20.0	21.0	± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l							5.65	5.75	5.77	± 0.20
HB HGB	g/dl							18.1	18.2	18.0	± 0.6
	g/l							181	182	180	± 6
	mmol/l							11.24	11.30	11.18	± 0.37
HT HCT	%							49.7	50.0	49.6	± 3.0
	l/l							0.497	0.500	0.496	± 0.030
VGM MCV	µm <sup>3</sup> ; fl							88	87	86	± 4
TGMH MCH	pg							32.0	31.7	31.2	± 2.0
	f mol							1.99	1.97	1.94	± 0.12
CCMH MCHC	g/dl							36.4	36.4	36.3	± 3.0
	g/l							364	364	363	± 30
	mmol/l							22.61	22.59	22.53	± 1.86
IDR RDW	%							12.5	13.5	15.0	± 3.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l							* 505	* 510	510	± 45   *± 55
VPM MPV	µm <sup>3</sup> ; fl							8.1	8.2	8.0	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 1.00
	%							N/A	N/A	N/A	± 5.0
MONO	#							N/A	N/A	N/A	± 0.80
	%							N/A	N/A	N/A	± 4.0
GRANULO	#							N/A	N/A	N/A	± 1.40
	%							N/A	N/A	N/A	± 7.0

Ref: FORM-0830 Rev.16 BACK / VERSO 9930056-A

\* For ABX Pentra 60 Pentra 80 instruments only