

ABX Minotrol 16



LOT MX 032 L
 Rev 1

CONTROL L

(Exp.) 2012-06-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF 120 - DX 120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							2.0	2.0	2.1	± 0.4
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l							2.36	2.31	2.36	± 0.15
HB HGB	g/dl							6.0	6.0	6.1	± 0.4
	g/l							60	60	61	± 4
	mmol/l							3.73	3.73	3.79	± 0.25
HT HCT	%							16.8	17.3	17.0	± 2.0
	l/l							0.168	0.173	0.170	± 0.020
VGM MCV	µm ³ ; fl							71	75	72	± 4
TGMH MCH	pg							25.4	26.0	25.8	± 2.0
	f mol							1.58	1.61	1.61	± 0.12
CCMH MCHC	g/dl							35.8	34.6	35.9	± 3.0
	g/l							358	346	359	± 30
	mmol/l							22.24	21.51	22.29	± 1.86
IDR RDW	%							14.5	14.5	17.0	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 82	* 80	78	± 20 *± 22
VPM MPV	µm ³ ; fl							9.8	9.6	9.2	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 9.0
MONO	#							N/A	N/A	N/A	± 0.20
	%							N/A	N/A	N/A	± 7.0
GRANULO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 8.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.16 BACK / VERSO 9930054-A

ABX Minotrol 16



LOT MX 032 N
 Rev 1

CONTROL N

(Exp.) 2012-06-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							7.4	7.5	8.0	± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							4.59	4.60	4.60	± 0.18
HB HGB	g/dl							13.4	13.4	13.5	± 0.5
	g/l							134	134	135	± 5
	mmol/l							8.32	8.32	8.38	± 0.31
HT HCT	%							37.2	37.7	37.3	± 2.5
	l/l							0.372	0.377	0.373	± 0.025
VGM MCV	µm ³ ; fl							81	82	81	± 4
TGMH MCH	pg							29.2	29.1	29.3	± 2.0
	fmol							1.81	1.81	1.82	± 0.12
CCMH MCHC	g/dl							36.0	35.5	36.2	± 3.0
	g/l							360	355	362	± 30
	mmol/l							22.38	22.06	22.50	± 1.86
IDR RDW	%							13.5	14.0	16.0	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 275	* 277	273	± 40 *± 40
VPM MPV	µm ³ ; fl							8.4	8.4	8.2	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 0.50
	%							N/A	N/A	N/A	± 6.0
MONO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 5.0
GRANULO	#							N/A	N/A	N/A	± 0.60
	%							N/A	N/A	N/A	± 7.0

Ref: FORM-0830 Rev.16 BACK / VERSO 9930055-A

* For ABX Pentra 60 Pentra 80 instruments only



ABX Minotrol 16



LOT MX 032 H
 Rev 1

CONTROL H

(Exp.) 2012-06-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							20.8	20.3	21.6	± 1.6
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l							5.50	5.59	5.64	± 0.20
HB HGB	g/dl							18.1	18.2	18.0	± 0.6
	g/l							181	182	180	± 6
	mmol/l							11.24	11.30	11.18	± 0.37
HT HCT	%							49.0	49.8	49.1	± 3.0
	l/l							0.490	0.498	0.491	± 0.030
VGM MCV	µm ³ ; fl							89	89	87	± 4
TGMH MCH	pg							32.9	32.6	31.9	± 2.0
	f mol							2.04	2.02	1.98	± 0.12
CCMH MCHC	g/dl							37.0	36.6	36.7	± 3.0
	g/l							370	366	367	± 30
	mmol/l							22.96	22.72	22.78	± 1.86
IDR RDW	%							12.5	13.5	15.5	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 505	* 510	506	± 45 *± 55
VPM MPV	µm ³ ; fl							8.5	8.6	8.5	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 1.00
	%							N/A	N/A	N/A	± 5.0
MONO	#							N/A	N/A	N/A	± 0.80
	%							N/A	N/A	N/A	± 4.0
GRANULO	#							N/A	N/A	N/A	± 1.40
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.16 BACK / VERSO 9930056-A