

**HORIBA ABX SAS**  
 HORIBA ABX SAS  
 Parc Euromédecine  
 Rue du Caducée - BP 7290  
 34184 MONTPELLIER Cedex 4  
 FRANCE  
 Tél.: 33 (0) 4 67 14 15 16  
 Fax.: 33 (0) 4 67 14 15 17

**LOT** MX 072 L  
 Rev 1

**CONTROL** L

(Exp.) 2012-10-05  
 (YYYY - MM - DD)

		ABX Lysebio										
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		DX NEXUS DF NEXUS	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1						1.9	1.9	2.1	± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.43						2.40	2.36	2.43	± 0.15	
HB HGB	g/dl	6.0						6.0	6.0	6.0	± 0.4	
	g/l	60						60	60	60	± 4	
	mmol/l	3.73						3.73	3.73	3.73	± 0.25	
HT HCT	%	17.3						16.8	17.2	17.3	± 2.0	
	l/l	0.173						0.168	0.172	0.173	± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	71						70	73	71	± 4	
TGMH MCH	pg	24.7						25.0	25.4	24.7	± 2.0	
	fmol	1.53						1.55	1.58	1.53	± 0.12	
CCMH MCHC	g/dl	34.8						35.7	34.8	34.8	± 3.0	
	g/l	348						357	348	348	± 30	
	mmol/l	21.60						22.18	21.63	21.60	± 1.86	
IDR RDW	%	18.5						14.5	15.0	18.5	± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72						* 73	* 70	72	± 20	*± 22
VPM MPV	µm <sup>3</sup> ; fl	9.2						9.6	9.6	9.2	± 2.0	
LYMPHO	#	N/A						N/A	N/A	N/A	± 0.40	
	%	N/A						N/A	N/A	N/A	± 9.0	
MONO	#	N/A						N/A	N/A	N/A	± 0.20	
	%	N/A						N/A	N/A	N/A	± 7.0	
GRANULO	#	N/A						N/A	N/A	N/A	± 0.40	
	%	N/A						N/A	N/A	N/A	± 8.0	


\* For ABX Pentra 60 Pentra 80 instruments only

Ref: TEMP-0830 Rev.17 BACK / VERSO 9930067-A

**HORIBA ABX SAS**  
 HORIBA ABX SAS  
 Parc Euromédecine  
 Rue du Caducée - BP 7290  
 34184 MONTPELLIER Cedex 4  
 FRANCE  
 Tél.: 33 (0) 4 67 14 15 16  
 Fax.: 33 (0) 4 67 14 15 17

**LOT** MX 072 N  
 Rev 1

**CONTROL** N

 (Exp.) 2012-10-05  
 (YYYY - MM - DD)

		ABX Lysebio										
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		DX NEXUS DF NEXUS	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	8.0						7.5	7.3	8.0	± 0.8	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.74						4.67	4.70	4.74	± 0.18	
HB HGB	g/dl	13.6						13.6	13.7	13.6	± 0.5	
	g/l	136						136	137	136	± 5	
	mmol/l	8.45						8.45	8.51	8.45	± 0.31	
HT HCT	%	37.9						37.4	38.1	37.9	± 2.5	
	l/l	0.379						0.374	0.381	0.379	± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	80						80	81	80	± 4	
TGMH MCH	pg	28.7						29.1	29.1	28.7	± 2.0	
	fmol	1.78						1.81	1.81	1.78	± 0.12	
CCMH MCHC	g/dl	35.9						36.4	36.0	35.9	± 3.0	
	g/l	359						364	360	359	± 30	
	mmol/l	22.27						22.61	22.35	22.27	± 1.86	
IDR RDW	%	17.5						13.5	14.5	17.5	± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	255						* 250	* 254	255	± 40	* ± 40
VPM MPV	µm <sup>3</sup> ; fl	8.5						8.7	8.7	8.5	± 2.0	
LYMPHO	#	N/A						N/A	N/A	N/A	± 0.50	
	%	N/A						N/A	N/A	N/A	± 6.0	
MONO	#	N/A						N/A	N/A	N/A	± 0.40	
	%	N/A						N/A	N/A	N/A	± 5.0	
GRANULO	#	N/A						N/A	N/A	N/A	± 0.60	
	%	N/A						N/A	N/A	N/A	± 7.0	

\* For ABX Pentra 60 Pentra 80 instruments only

Ref: TEMP-0830 Rev.17 BACK / VERSO 9930068-A

**HORIBA ABX SAS**  
 HORIBA ABX SAS  
 Parc Euromédecine  
 Rue du Caducée - BP 7290  
 34184 MONTPELLIER Cedex 4  
 FRANCE  
 Tél.: 33 (0) 4 67 14 15 16  
 Fax.: 33 (0) 4 67 14 15 17

**LOT** MX 072 H  
 Rev 1

**CONTROL** H

 (Exp.) **2012-10-05**  
 (YYYY - MM - DD)

		ABX Lysebio										
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		DX NEXUS DF NEXUS	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	21.8						20.9	20.4	21.8	± 1.6	
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.72						5.58	5.66	5.72	± 0.20	
HB HGB	g/dl	18.0						18.2	18.2	18.0	± 0.6	
	g/l	180						182	182	180	± 6	
	mmol/l	11.18						11.30	11.30	11.18	± 0.37	
HT HCT	%	49.8						49.7	49.8	49.8	± 3.0	
	l/l	0.498						0.497	0.498	0.498	± 0.030	
VGM MCV	µm <sup>3</sup> ; fl	87						89	88	87	± 4	
TGMH MCH	pg	31.5						32.6	32.2	31.5	± 2.0	
	fmol	1.95						2.03	2.00	1.95	± 0.12	
CCMH MCHC	g/dl	36.2						36.6	36.5	36.2	± 3.0	
	g/l	362						366	365	362	± 30	
	mmol/l	22.46						22.76	22.69	22.46	± 1.86	
IDR RDW	%	16.5						12.5	13.5	16.5	± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	528						* 525	* 540	528	± 45	* ± 55
VPM MPV	µm <sup>3</sup> ; fl	8.1						8.4	8.4	8.1	± 2.0	
LYMPHO	#	N/A						N/A	N/A	N/A	± 1.00	
	%	N/A						N/A	N/A	N/A	± 5.0	
MONO	#	N/A						N/A	N/A	N/A	± 0.80	
	%	N/A						N/A	N/A	N/A	± 4.0	
GRANULO	#	N/A						N/A	N/A	N/A	± 1.40	
	%	N/A						N/A	N/A	N/A	± 7.0	

\* For ABX Pentra 60 Pentra 80 instruments only