

**LOT** MX 075 L  
Rev 1

**CONTROL** L

 (Exp.)

2015-10-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1		2.1			2.0	1.9	2.0	2.1	± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.45		2.41			2.38	2.43	2.38	2.45	± 0.15
HB HGB	g/dl	6.0		6.1			5.9	5.9	5.9	6.0	± 0.4
	g/l	60		61			59	59	59	60	± 4
	mmol/l	3.73		3.79			3.66	3.66	3.66	3.73	± 0.25
HT HCT	%	16.9		16.4			16.7	16.5	16.7	16.9	± 2.0
	l/l	0.169		0.164			0.167	0.165	0.167	0.169	± 0.020
VGM MCV	µm <sup>3</sup> ; fl	69		68			70	68	70	69	± 4
TGMH MCH	pg	24.5		25.3			24.8	24.3	24.8	24.5	± 2.0
	fmol	1.52		1.57			1.54	1.51	1.54	1.52	± 0.12
CCMH MCHC	g/dl	35.5		37.2			35.4	35.7	35.4	35.5	± 3.0
	g/l	355		372			354	357	354	355	± 30
	mmol/l	22.04		23.12			21.99	22.17	21.99	22.04	± 1.86
IDR RDW	%	18.5		13.6			15.8	15.8	15.8	18.5	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70		75			* 60	* 65	* 60	70	± 20 *± 22
VPM MPV	µm <sup>3</sup> ; fl	10.2		8.4			8.7	9.4	8.7	10.2	± 2.0
LYMPHO	#	N/A		1.3			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		62.5			N/A	N/A	N/A	N/A	± 8.0
MONO	#	N/A		0.3			N/A	N/A	N/A	N/A	± 0.20
	%	N/A		12.0			N/A	N/A	N/A	N/A	± 6.0
GRANULO	#	N/A		0.5			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		25.5			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only

LOT		MX 075 N		CONTROL		N		2015-10-05		(Exp.)		
Rev 1												(YYYY-MM-DD)
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio										TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA		
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.9		7.6			7.6	7.5	7.6	7.9	± 0.8	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.65		4.57			4.63	4.62	4.63	4.65	± 0.18	
HB HGB	g/dl	13.4		13.3			13.5	13.5	13.5	13.4	± 0.5	
	g/l	134		133			135	135	135	134	± 5	
	mmol/l	8.32		8.26			8.38	8.38	8.38	8.32	± 0.31	
HT HCT	%	37.2		36.6			37.0	37.0	37.0	37.2	± 2.5	
	l/l	0.372		0.366			0.370	0.370	0.370	0.372	± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	80		80			80	80	80	80	± 4	
TGMH MCH	pg	28.8		29.1			29.2	29.2	29.2	28.8	± 2.0	
	fmol	1.79		1.81			1.81	1.81	1.81	1.79	± 0.12	
CCMH MCHC	g/dl	36.0		36.4			36.4	36.5	36.4	36.0	± 3.0	
	g/l	360		364			364	365	364	360	± 30	
	mmol/l	22.37		22.59			22.63	22.68	22.63	22.37	± 1.86	
IDR RDW	%	16.0		12.8			14.6	13.8	14.6	16.0	± 3.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	260		255			* 250	* 250	* 250	260	± 40   *± 40	
VPM MPV	µm <sup>3</sup> ; fl	9.6		8.6			9.7	9.8	9.7	9.6	± 2.0	
LYMPHO	#	N/A		2.5			N/A	N/A	N/A	N/A	± 0.40	
	%	N/A		33.0			N/A	N/A	N/A	N/A	± 5.0	
MONO	#	N/A		0.6			N/A	N/A	N/A	N/A	± 0.30	
	%	N/A		8.5			N/A	N/A	N/A	N/A	± 4.0	
GRANULO	#	N/A		4.4			N/A	N/A	N/A	N/A	± 0.60	
	%	N/A		58.5			N/A	N/A	N/A	N/A	± 6.0	

\* For ABX Pentra 60 Pentra 80 instruments only

1219930084-B  
BACK / VERSO  
Ret: TEMP-0830 Rev.37

**LOT** MX 075 H  
Rev 1

**CONTROL** H

 (Exp.)

2015-10-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	20.8		19.8			20.8	20.7	20.8	20.8	± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.75		5.62			5.72	5.66	5.72	5.75	± 0.20
HB HGB	g/dl	17.8		17.7			18.2	18.1	18.2	17.8	± 0.6
	g/l	178		177			182	181	182	178	± 6
	mmol/l	11.05		10.99			11.30	11.24	11.30	11.05	± 0.37
HT HCT	%	48.3		48.9			49.2	49.2	49.2	48.3	± 3.0
	l/l	0.483		0.489			0.492	0.492	0.492	0.483	± 0.030
VGM MCV	µm <sup>3</sup> ; fl	84		87			86	87	86	84	± 4
TGMH MCH	pg	31.0		31.5			31.8	32.0	31.8	31.0	± 2.0
	fmol	1.92		1.96			1.98	1.99	1.98	1.92	± 0.12
CCMH MCHC	g/dl	36.9		36.2			37.0	36.8	37.0	36.9	± 3.0
	g/l	369		362			370	368	370	369	± 30
	mmol/l	22.89		22.48			22.98	22.83	22.98	22.89	± 1.86
IDR RDW	%	15.5		11.9			13.8	12.8	13.8	15.5	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	540		510			* 525	* 525	* 525	540	± 45   *± 55
VPM MPV	µm <sup>3</sup> ; fl	9.3		8.3			9.3	9.2	9.3	9.3	± 2.0
LYMPHO	#	N/A		3.1			N/A	N/A	N/A	N/A	± 1.00
	%	N/A		15.5			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		1.1			N/A	N/A	N/A	N/A	± 0.80
	%	N/A		5.5			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		15.6			N/A	N/A	N/A	N/A	± 1.40
	%	N/A		79.0			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only