

**LOT** MX 115 L  
Rev 1

**CONTROL** L

 (Exp.)

**2016-02-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE	
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80		120 - 120 RETIC DF120 - DX120
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.2		2.1			2.1	2.0	2.1	2.2	± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.41		2.35			2.33	2.35	2.33	2.41	± 0.15
HB HGB	g/dl	5.9		5.9			5.7	5.7	5.7	5.9	± 0.4
	g/l	59		59			57	57	57	59	± 4
	mmol/l	3.66		3.66			3.54	3.54	3.54	3.66	± 0.25
HT HCT	%	16.9		16.2			16.5	16.2	16.5	16.9	± 2.0
	l/l	0.169		0.162			0.165	0.162	0.165	0.169	± 0.020
VGM MCV	µm <sup>3</sup> ; fl	70		69			71	69	71	70	± 4
TGMH MCH	pg	24.5		25.1			24.5	24.3	24.5	24.5	± 2.0
	fmol	1.52		1.56			1.52	1.51	1.52	1.52	± 0.12
CCMH MCHC	g/dl	35.0		36.4			34.5	35.2	34.5	35.0	± 3.0
	g/l	350		364			345	352	345	350	± 30
	mmol/l	21.72		22.60			21.40	21.83	21.40	21.72	± 1.86
IDR RDW	%	17.0		13.0			15.5	16.0	15.5	17.0	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	68		71			* 57	* 58	* 57	68	± 20 *± 22
VPM MPV	µm <sup>3</sup> ; fl	10.0		8.3			8.8	9.0	8.8	10.0	± 2.0
LYMPHO	#	N/A		1.3			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		63.0			N/A	N/A	N/A	N/A	± 8.0
MONO	#	N/A		0.2			N/A	N/A	N/A	N/A	± 0.20
	%	N/A		8.5			N/A	N/A	N/A	N/A	± 6.0
GRANULO	#	N/A		0.6			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		28.5			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only

LOT		MX 115 N		CONTROL N						2016-02-05 (YYYY-MM-DD)	
Rev 1											
		ABX Lysebio									
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	8.2		8.0			7.9	7.8	7.9	8.2	± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.53		4.47			4.52	4.50	4.52	4.53	± 0.18
HB HGB	g/dl	13.4		13.2			13.4	13.3	13.4	13.4	± 0.5
	g/l	134		132			134	133	134	134	± 5
	mmol/l	8.32		8.20			8.32	8.26	8.32	8.32	± 0.31
HT HCT	%	36.2		36.2			36.6	36.5	36.6	36.2	± 2.5
	l/l	0.362		0.362			0.366	0.365	0.366	0.362	± 0.025
VGM MCV	µm <sup>3</sup> ; fl	80		81			81	81	81	80	± 4
TGMH MCH	pg	29.6		29.5			29.6	29.6	29.6	29.6	± 2.0
	fmol	1.84		1.83			1.84	1.84	1.84	1.84	± 0.12
CCMH MCHC	g/dl	37.0		36.5			36.6	36.5	36.6	37.0	± 3.0
	g/l	370		365			366	365	366	370	± 30
	mmol/l	22.96		22.64			22.73	22.66	22.73	22.96	± 1.86
IDR RDW	%	16.0		12.5			15.0	14.0	15.0	16.0	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	255		245			* 252	* 248	* 252	255	± 40   *± 40
VPM MPV	µm <sup>3</sup> ; fl	9.0		8.0			9.1	9.3	9.1	9.0	± 2.0
LYMPHO	#	N/A		2.7			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		33.5			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		0.5			N/A	N/A	N/A	N/A	± 0.30
	%	N/A		6.5			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		4.8			N/A	N/A	N/A	N/A	± 0.60
	%	N/A		60.0			N/A	N/A	N/A	N/A	± 6.0

\* For ABX Pentra 60 Pentra 80 instruments only

1219930084-B  
BACK / VERSO  
Ret: TEMP-0830 Rev.37

**LOT** MX 115 H  
Rev 1

**CONTROL** H

 (Exp.)

2016-02-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	21.2		20.1			20.7	21.0	20.7	21.2	± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.77		5.62			5.77	5.66	5.77	5.77	± 0.20
HB HGB	g/dl	18.3		17.9			18.6	18.5	18.6	18.3	± 0.6
	g/l	183		179			186	185	186	183	± 6
	mmol/l	11.36		11.12			11.55	11.49	11.55	11.36	± 0.37
HT HCT	%	49.6		49.5			50.2	49.8	50.2	49.6	± 3.0
	l/l	0.496		0.495			0.502	0.498	0.502	0.496	± 0.030
VGM MCV	µm <sup>3</sup> ; fl	86		88			87	88	87	86	± 4
TGMH MCH	pg	31.7		31.9			32.2	32.7	32.2	31.7	± 2.0
	fmol	1.97		1.98			2.00	2.03	2.00	1.97	± 0.12
CCMH MCHC	g/dl	36.9		36.2			37.1	37.1	37.1	36.9	± 3.0
	g/l	369		362			371	371	371	369	± 30
	mmol/l	22.90		22.48			23.01	23.07	23.01	22.90	± 1.86
IDR RDW	%	14.5		12.0			13.5	12.5	13.5	14.5	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	540		490			* 535	* 522	* 535	540	± 45   *± 55
VPM MPV	µm <sup>3</sup> ; fl	8.7		8.0			8.9	9.0	8.9	8.7	± 2.0
LYMPHO	#	N/A		3.6			N/A	N/A	N/A	N/A	± 1.00
	%	N/A		18.0			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		1.1			N/A	N/A	N/A	N/A	± 0.80
	%	N/A		5.5			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		15.4			N/A	N/A	N/A	N/A	± 1.40
	%	N/A		76.5			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only