

LOT MX 401 L
Rev 1

CONTROL L

 (Exp.)

2016-12-05
(YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-----------------|----------------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP CRP 200 | 8P | 18P | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 1.9 | 2.0 | 2.0 | 2.0 | 1.9 | | | | | ± 0.4 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.32 | 2.32 | 2.33 | 2.33 | 2.30 | | | | | ± 0.15 |
| HB HGB | g/dl | 6.0 | 6.0 | 5.9 | 5.9 | 6.0 | | | | | ± 0.4 |
| | g/l | 60 | 60 | 59 | 59 | 60 | | | | | ± 4 |
| | mmol/l | 3.73 | 3.73 | 3.66 | 3.66 | 3.73 | | | | | ± 0.25 |
| HT HCT | % | 16.2 | 16.0 | 16.3 | 16.3 | 16.1 | | | | | ± 2.0 |
| | l/l | 0.162 | 0.160 | 0.163 | 0.163 | 0.161 | | | | | ± 0.020 |
| VGM MCV | µm ³ ; fl | 70 | 69 | 70 | 70 | 70 | | | | | ± 4 |
| TGMH MCH | pg | 25.9 | 25.9 | 25.3 | 25.3 | 26.1 | | | | | ± 2.0 |
| | fmol | 1.61 | 1.61 | 1.57 | 1.57 | 1.62 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.9 | 37.5 | 36.2 | 36.2 | 37.3 | | | | | ± 3.0 |
| | g/l | 369 | 375 | 362 | 362 | 373 | | | | | ± 30 |
| | mmol/l | 22.94 | 23.28 | 22.46 | 22.46 | 23.14 | | | | | ± 1.86 |
| IDR RDW | % | 13.7 | 13.9 | N/A | 14.0 | 13.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 80 | 74 | 71 | 76 | 75 | | | | | ± 20 |
| VPM MPV | µm ³ ; fl | 8.9 | 8.9 | N/A | 8.8 | 8.0 | | | | | ± 2.0 |
| LYMPHO | # | 1.22 | 1.28 | N/A | 1.04 | 1.20 | | | | | ± 0.40 |
| | % | 64.0 | 64.0 | N/A | 52.0 | 63.0 | | | | | ± 8.0 |
| MONO | # | 0.20 | 0.21 | N/A | 0.30 | 0.20 | | | | | ± 0.20 |
| | % | 9.5 | 10.5 | N/A | 15.0 | 10.5 | | | | | ± 6.0 |
| GRANULO | # | 0.50 | 0.51 | N/A | 0.66 | 0.50 | | | | | ± 0.40 |
| | % | 26.5 | 25.5 | N/A | 33.0 | 26.5 | | | | | ± 7.0 |

LOT MX 401 N
Rev 1

CONTROL N

 (Exp.)

2016-12-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|--|-----------------|----------------|--------|--------|----------------------------|----------------|--------|--------|--------|-------------------------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | PENTRA | |
| | | ES60 Care ST | CRP CRP 200 | 8P | 18P | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 7.7 | 7.7 | 7.7 | 7.6 | 7.7 | | | | | ± 0.8 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 4.55 | 4.50 | 4.53 | 4.48 | 4.50 | | | | | ± 0.18 |
| HB HGB | g/dl | 13.1 | 13.1 | 13.0 | 13.1 | 13.1 | | | | | ± 0.5 |
| | g/l | 131 | 131 | 130 | 131 | 131 | | | | | ± 5 |
| | mmol/l | 8.14 | 8.14 | 8.07 | 8.14 | 8.14 | | | | | ± 0.31 |
| HT HCT | % | 36.4 | 36.0 | 36.2 | 35.8 | 36.0 | | | | | ± 2.5 |
| | l/l | 0.364 | 0.360 | 0.362 | 0.358 | 0.360 | | | | | ± 0.025 |
| VGM MCV | µm ³ ; fl | 80 | 80 | 80 | 80 | 80 | | | | | ± 4 |
| TGMH MCH | pg | 28.8 | 29.1 | 28.7 | 29.2 | 29.1 | | | | | ± 2.0 |
| | fmol | 1.79 | 1.81 | 1.78 | 1.82 | 1.81 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.0 | 36.4 | 35.9 | 36.6 | 36.4 | | | | | ± 3.0 |
| | g/l | 360 | 364 | 359 | 366 | 364 | | | | | ± 30 |
| | mmol/l | 22.35 | 22.60 | 22.28 | 22.70 | 22.60 | | | | | ± 1.86 |
| IDR RDW | % | 13.8 | 13.2 | N/A | 13.5 | 12.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 260 | 253 | 253 | 255 | 254 | | | | | ± 40 |
| VPM MPV | µm ³ ; fl | 9.1 | 8.9 | N/A | 9.2 | 8.2 | | | | | ± 2.0 |
| LYMPHO | # | 2.62 | 2.54 | N/A | 2.13 | 2.54 | | | | | ± 0.40 |
| | % | 34.0 | 33.0 | N/A | 28.0 | 33.0 | | | | | ± 5.0 |
| MONO | # | 0.58 | 0.65 | N/A | 0.61 | 0.62 | | | | | ± 0.30 |
| | % | 7.5 | 8.5 | N/A | 8.0 | 8.0 | | | | | ± 4.0 |
| GRANULO | # | 4.50 | 4.50 | N/A | 4.86 | 4.54 | | | | | ± 0.60 |
| | % | 58.5 | 58.5 | N/A | 64.0 | 59.0 | | | | | ± 6.0 |

Ref: TEMP-0830 Rev.37 FRONT / RECTO 1219930084-B

LOT
MX 401 H
Rev 1

CONTROL H

 (Exp.)

2016-12-05
(YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-----------------|----------------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP CRP 200 | 8P | 18P | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 19.7 | 19.7 | 19.4 | 19.6 | 19.6 | | | | | ± 1.6 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 5.74 | 5.67 | 5.69 | 5.65 | 5.66 | | | | | ± 0.20 |
| HB HGB | g/dl | 17.8 | 17.8 | 17.6 | 17.8 | 17.8 | | | | | ± 0.6 |
| | g/l | 178 | 178 | 176 | 178 | 178 | | | | | ± 6 |
| | mmol/l | 11.05 | 11.05 | 10.93 | 11.05 | 11.05 | | | | | ± 0.37 |
| HT HCT | % | 49.9 | 49.9 | 50.1 | 49.7 | 49.8 | | | | | ± 3.0 |
| | l/l | 0.499 | 0.499 | 0.501 | 0.497 | 0.498 | | | | | ± 0.030 |
| VGM MCV | µm ³ ; fl | 87 | 88 | 88 | 88 | 88 | | | | | ± 4 |
| TGMH MCH | pg | 31.0 | 31.4 | 30.9 | 31.5 | 31.4 | | | | | ± 2.0 |
| | fmol | 1.93 | 1.95 | 1.92 | 1.96 | 1.95 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 35.6 | 35.7 | 35.1 | 35.8 | 35.7 | | | | | ± 3.0 |
| | g/l | 356 | 357 | 351 | 358 | 357 | | | | | ± 30 |
| | mmol/l | 22.14 | 22.15 | 21.83 | 22.23 | 22.19 | | | | | ± 1.86 |
| IDR RDW | % | 13.5 | 12.7 | N/A | 13.0 | 12.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 510 | 500 | 525 | 530 | 500 | | | | | ± 45 |
| VPM MPV | µm ³ ; fl | 8.4 | 8.1 | N/A | 8.7 | 7.7 | | | | | ± 2.0 |
| LYMPHO | # | 3.11 | 2.96 | N/A | 2.45 | 2.94 | | | | | ± 1.00 |
| | % | 15.8 | 15.0 | N/A | 12.5 | 15.0 | | | | | ± 5.0 |
| MONO | # | 0.95 | 1.08 | N/A | 0.80 | 0.98 | | | | | ± 0.80 |
| | % | 4.8 | 5.5 | N/A | 4.0 | 5.0 | | | | | ± 4.0 |
| GRANULO | # | 15.64 | 15.66 | N/A | 16.37 | 15.68 | | | | | ± 1.40 |
| | % | 79.4 | 79.5 | N/A | 83.5 | 80.0 | | | | | ± 7.0 |