



LOT		MX 411 L		CONTROL		L		(Exp.)				2018-08-05 (YYYY - MM - DD)			
		Rev 1		ABX Lyse											
PARAMETRES PARAMETERS		UNITES UNITS		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE		
				ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60							
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0	2.1	2.0	2.0						± 0.4		
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.23	2.25	2.27	2.26	2.24						± 0.15		
HB	HGB	g/dl	5.9	5.9	6.0	5.9	5.9						± 0.4		
		g/l	59	59	60	59	59						± 4		
		mmol/l	3.66	3.66	3.73	3.66	3.66						± 0.25		
HT	HCT	%	15.8	16.0	16.1	16.3	15.9						± 2.0		
		l/l	0.158	0.160	0.161	0.163	0.159						± 0.020		
VGM	MCV	µm ³ ; fl	71	71	71	72	71						± 4		
TGMH	MCH	pg	26.5	26.2	26.4	26.1	26.3						± 2.0		
		fmol	1.64	1.63	1.64	1.62	1.64						± 0.12		
CCMH	MCHC	g/dl	37.3	36.9	37.2	36.3	37.1						± 3.0		
		g/l	373	369	372	363	371						± 30		
		mmol/l	23.14	22.94	23.12	22.52	23.04						± 1.86		
IDR	RDW	%	14.0	13.5	N/A	14.0	13.5						± 3.0		
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	80	76	73	80	79						± 20		
VPM	MPV	µm ³ ; fl	9.0	8.9	N/A	9.0	8.1						± 2.0		
LYMPHO	#	1.31	1.32	N/A	1.06	1.29							± 0.40		
	%	65.5	66.0	N/A	53.0	64.5							± 8.0		
MONO	#	0.20	0.20	N/A	0.31	0.20							± 0.20		
	%	8.5	9.0	N/A	15.5	10.0							± 6.0		
GRANULO	#	0.52	0.50	N/A	0.63	0.51							± 0.40		
	%	26.0	25.0	N/A	31.5	25.5							± 7.0		

LOT		MX 411 N		CONTROL		N		 (Exp.)		2018-08-05 (YYYY-MM-DD)	
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	7.5	7.5	7.3	7.6	7.6					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.62	4.62	4.64	4.62	4.62					± 0.18
HB HGB	g/dl	13.4	13.4	13.5	13.5	13.5					± 0.5
	g/l	134	134	135	135	135					± 5
	mmol/l	8.32	8.32	8.38	8.38	8.38					± 0.31
HT HCT	%	37.0	37.4	37.6	37.4	37.4					± 2.5
	l/l	0.370	0.374	0.376	0.374	0.374					± 0.025
VGM MCV	µm ³ ; fl	80	81	81	81	81					± 4
TGMH MCH	pg	29.0	29.0	29.1	29.2	29.2					± 2.0
	fmol	1.80	1.80	1.81	1.81	1.81					± 0.12
CCMH MCHC	g/dl	36.3	35.8	35.9	36.1	36.1					± 3.0
	g/l	363	358	359	361	361					± 30
	mmol/l	22.51	22.24	22.31	22.40	22.40					± 1.86
IDR RDW	%	13.5	13.5	N/A	14.0	13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	270	264	259	268	266					± 40
VPM MPV	µm ³ ; fl	8.4	8.4	N/A	8.8	7.8					± 2.0
LYMPHO	#	2.55	2.51	N/A	1.98	2.47					± 0.40
	%	34.0	33.5	N/A	26.0	32.5					± 5.0
MONO	#	0.53	0.56	N/A	0.72	0.61					± 0.30
	%	7.0	7.5	N/A	9.5	8.0					± 4.0
GRANULO	#	4.43	4.43	N/A	4.90	4.52					± 0.60
	%	59.0	59.0	N/A	64.5	59.5					± 6.0

Ref: TEMP-0830 Rev.38 FRONT / RECTO 1219930084-B



LOT		MX 411 H		CONTROL		H		 (Exp.)		2018-08-05 (YYYY - MM - DD)			
		Rev 1		ABX Lyse									
PARAMETRES PARAMETERS		UNITES UNITS		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
				ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.4	20.2	19.8	20.0	20.4						± 1.6
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.80	5.78	5.75	5.79	5.79						± 0.20
HB	HGB	g/dl	18.0	18.1	18.1	18.2	18.1						± 0.6
		g/l	180	181	181	182	181						± 6
		mmol/l	11.18	11.24	11.24	11.30	11.24						± 0.37
HT	HCT	%	51.0	50.9	50.6	51.0	51.0						± 3.0
		l/l	0.510	0.509	0.506	0.510	0.510						± 0.030
VGM	MCV	µm ³ ; fl	88	88	88	88	88						± 4
TGMH	MCH	pg	31.0	31.3	31.5	31.4	31.3						± 2.0
		fmol	1.93	1.94	1.95	1.95	1.94						± 0.12
CCMH	MCHC	g/dl	35.3	35.6	35.8	35.7	35.5						± 3.0
		g/l	353	356	358	357	355						± 30
		mmol/l	21.90	22.10	22.21	22.18	22.06						± 1.86
IDR	RDW	%	13.0	13.0	N/A	13.5	12.5						± 3.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	494	495	495	495	495						± 45
VPM	MPV	µm ³ ; fl	8.6	8.6	N/A	9.0	8.0						± 2.0
LYMPHO	#	3.26	3.13	N/A	2.50	3.16							± 1.00
	%	16.0	15.5	N/A	12.5	15.5							± 5.0
MONO	#	1.12	1.11	N/A	0.80	1.02							± 0.80
	%	5.5	5.5	N/A	4.0	5.0							± 4.0
GRANULO	#	16.01	15.96	N/A	16.70	16.22							± 1.40
	%	78.5	79.0	N/A	83.5	79.5							± 7.0