

**LOT** MX 412 L  
Rev 1

**CONTROL** L

 (Exp.)

2018-10-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.1	2.2	2.1	2.1					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.27	2.28	2.29	2.30	2.27					± 0.15
HB HGB	g/dl	6.0	6.0	6.0	6.0	6.0					± 0.4
	g/l	60	60	60	60	60					± 4
	mmol/l	3.73	3.73	3.73	3.73	3.73					± 0.25
HT HCT	%	16.1	16.2	16.3	16.3	16.1					± 2.0
	l/l	0.161	0.162	0.163	0.163	0.161					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	71	71	71	71	71					± 4
TGMH MCH	pg	26.4	26.3	26.2	26.1	26.4					± 2.0
	fmol	1.64	1.63	1.63	1.62	1.64					± 0.12
CCMH MCHC	g/dl	37.2	37.1	36.9	36.7	37.2					± 3.0
	g/l	372	371	369	367	372					± 30
	mmol/l	23.12	23.02	22.92	22.82	23.12					± 1.86
IDR RDW	%	14.0	13.5	N/A	14.0	13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	80	80	71	80	79					± 20
VPM MPV	µm <sup>3</sup> ; fl	8.7	8.3	N/A	8.7	7.8					± 2.0
LYMPHO	#	1.29	1.37	N/A	1.05	1.32					± 0.40
	%	64.5	65.0	N/A	50.0	63.0					± 8.0
MONO	#	0.20	0.20	N/A	0.34	0.21					± 0.20
	%	8.5	9.0	N/A	16.0	10.0					± 6.0
GRANULO	#	0.54	0.55	N/A	0.71	0.57					± 0.40
	%	27.0	26.0	N/A	34.0	27.0					± 7.0

**LOT** MX 412 N  
Rev 1

**CONTROL** N

 (Exp.)

**2018-10-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.4	7.4	7.4	7.5	7.6					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.55	4.57	4.62	4.59	4.59					± 0.18
HB HGB	g/dl	13.3	13.3	13.4	13.4	13.3					± 0.5
	g/l	133	133	134	134	133					± 5
	mmol/l	8.26	8.26	8.32	8.32	8.26					± 0.31
HT HCT	%	36.4	36.6	37.0	36.7	36.7					± 2.5
	l/l	0.364	0.366	0.370	0.367	0.367					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	80	80	80	80	80					± 4
TGMH MCH	pg	29.2	29.1	29.0	29.2	29.0					± 2.0
	fmol	1.82	1.81	1.80	1.81	1.80					± 0.12
CCMH MCHC	g/dl	36.5	36.4	36.3	36.5	36.2					± 3.0
	g/l	365	364	363	365	362					± 30
	mmol/l	22.69	22.59	22.51	22.66	22.49					± 1.86
IDR RDW	%	14.0	13.5	N/A	14.0	13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	260	265	260	270	263					± 40
VPM MPV	µm <sup>3</sup> ; fl	7.8	7.8	N/A	8.2	7.4					± 2.0
LYMPHO	#	2.52	2.44	N/A	1.95	2.43					± 0.40
	%	34.0	33.0	N/A	26.0	32.0					± 5.0
MONO	#	0.52	0.59	N/A	0.75	0.61					± 0.30
	%	7.0	8.0	N/A	10.0	8.0					± 4.0
GRANULO	#	4.37	4.37	N/A	4.80	4.56					± 0.60
	%	59.0	59.0	N/A	64.0	60.0					± 6.0

Ref: TEMP-0830 Rev.36 FRONT / RECTO 1219930084-B

LOT		MX 412 H		CONTROL H						(Exp.)	2018-10-05 (YYYY-MM-DD)
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	20.5	20.4	19.9	20.4	20.5					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.60	5.60	5.60	5.64	5.62					± 0.20
HB HGB	g/dl	18.0	17.8	17.9	18.1	17.9					± 0.6
	g/l	180	178	179	181	179					± 6
	mmol/l	11.18	11.05	11.12	11.24	11.12					± 0.37
HT HCT	%	49.8	49.8	49.8	50.2	50.0					± 3.0
	l/l	0.498	0.498	0.498	0.502	0.500					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	89	89	89	89	89					± 4
TGMH MCH	pg	32.1	31.8	32.0	32.1	31.9					± 2.0
	fmol	2.00	1.97	1.98	1.99	1.98					± 0.12
CCMH MCHC	g/dl	36.1	35.7	35.9	36.1	35.8					± 3.0
	g/l	361	357	359	361	358					± 30
	mmol/l	22.43	22.18	22.30	22.39	22.22					± 1.86
IDR RDW	%	13.0	13.0	N/A	13.5	12.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	525	525	520	540	520					± 45
VPM MPV	µm <sup>3</sup> ; fl	8.0	7.9	N/A	8.5	7.5					± 2.0
LYMPHO	#	3.08	2.96	N/A	2.35	2.87					± 1.00
	%	15.0	14.5	N/A	11.5	14.0					± 5.0
MONO	#	1.13	1.22	N/A	0.92	1.13					± 0.80
	%	5.5	6.0	N/A	4.5	5.5					± 4.0
GRANULO	#	16.30	16.22	N/A	17.14	16.50					± 1.40
	%	79.5	79.5	N/A	84.0	80.5					± 7.0