

LOT MX 434 L
 Rev 1

CONTROL L

(Exp.)

2022-06-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.23	2.26			2.23					± 0.17
HB HGB	g/dl	6.0	6.1			6.0					± 0.5
	g/l	60	61			60					± 5
	mmol/l	3.73	3.79			3.73					± 0.31
HT HCT	%	16.1	16.3			16.1					± 2.0
	l/l	0.161	0.163			0.161					± 0.020
VGM MCV	µm ³ ; fl	72	72			72					± 4
TGMH MCH	pg	26.9	27.0			26.9					± 2.0
	fmol	1.67	1.68			1.67					± 0.12
CCMH MCHC	g/dl	37.4	37.5			37.4					± 3.0
	g/l	374	375			374					± 30
	mmol/l	23.21	23.28			23.21					± 1.86
IDR RDW	%	15.0	13.5			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	80	79			78					± 22
VPM MPV	µm ³ ; fl	8.7	8.5			8.3					± 2.0
LYMPHO	#	1.23	1.26			1.26					± 0.40
	%	61.5	63.0			63.0					± 8.0
MONO	#	0.23	0.25			0.22					± 0.20
	%	11.5	12.5			11.0					± 6.0
GRANULO	#	0.54	0.49			0.52					± 0.40
	%	27.0	24.5			26.0					± 7.0

LOT MX 434 N
Rev 1

CONTROL N

 (Exp.)

2022-06-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.8	7.8			7.8					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.55	4.61			4.59					± 0.26
HB HGB	g/dl	13.4	13.6			13.6					± 0.7
	g/l	134	136			136					± 7
	mmol/l	8.32	8.45			8.45					± 0.43
HT HCT	%	37.3	37.3			37.6					± 2.7
	l/l	0.373	0.373			0.376					± 0.027
VGM MCV	µm ³ ; fl	82	81			82					± 4
TGMH MCH	pg	29.5	29.5			29.6					± 2.0
	fmol	1.83	1.83			1.84					± 0.12
CCMH MCHC	g/dl	35.9	36.4			36.1					± 3.0
	g/l	359	364			361					± 30
	mmol/l	22.30	22.62			22.44					± 1.86
IDR RDW	%	14.0	13.0			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	259	260			258					± 40
VPM MPV	µm ³ ; fl	8.4	7.9			7.8					± 2.0
LYMPHO	#	2.57	2.65			2.57					± 0.40
	%	33.0	34.0			33.0					± 5.0
MONO	#	0.55	0.62			0.51					± 0.30
	%	7.0	8.0			6.5					± 4.0
GRANULO	#	4.68	4.52			4.72					± 0.60
	%	60.0	58.0			60.5					± 6.0

Ref: TEMP-0830 Rev.41 FRONT / RECTO 130090307



LOT

MX 434 H

CONTROL H

(Exp.)

2022-06-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.3	20.2			20.5					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.64	5.66			5.70					± 0.32
HB HGB	g/dl	18.2	18.3			18.3					± 0.9
	g/l	182	183			183					± 9
	mmol/l	11.30	11.36			11.36					± 0.56
HT HCT	%	50.2	50.9			50.9					± 3.4
	l/l	0.502	0.509			0.509					± 0.034
VGM MCV	µm ³ ; fl	89	90			89					± 4
TGMH MCH	pg	32.3	32.3			32.1					± 2.0
	fmol	2.00	2.01			1.99					± 0.12
CCMH MCHC	g/dl	36.3	35.9			36.1					± 3.0
	g/l	363	359			361					± 30
	mmol/l	22.52	22.31			22.40					± 1.86
IDR RDW	%	13.0	13.0			12.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	518	515			515					± 65
VPM MPV	µm ³ ; fl	8.3	7.9			7.8					± 2.0
LYMPHO	#	3.15	3.23			3.08					± 1.00
	%	15.5	16.0			15.0					± 5.0
MONO	#	1.02	1.11			1.03					± 0.80
	%	5.0	5.5			5.0					± 4.0
GRANULO	#	16.14	15.86			16.40					± 1.40
	%	79.5	78.5			80.0					± 7.0