

LOT		MX 428 L		CONTROL		L		(Exp.) <b>2021-06-05</b> (YYYY - MM - DD)			
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1	2.0			2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.38	2.34			2.38					± 0.15
HB HGB	g/dl	6.0	5.9			6.1					± 0.4
	g/l	69	59			61					± 4
	mmol/l	3.73	3.66			3.79					± 0.25
HT HCT	%	16.4	15.9			16.4					± 2.0
	l/l	0.164	0.159			0.164					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	69	68			69					± 4
TGMH MCH	pg	25.2	25.2			25.6					± 2.0
	fmol	1.57	1.57			1.59					± 0.12
CCMH MCHC	g/dl	36.5	37.1			37.1					± 3.0
	g/l	365	371			371					± 30
	mmol/l	22.69	23.03			23.07					± 1.86
IDR RDW	%	14.5	14.5			13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	77	81			77					± 20
VPM MPV	µm <sup>3</sup> ; fl	9.1	8.6			8.1					± 2.0
LYMPHO	#	1.40	1.40			1.34					± 0.40
	%	66.5	70.0			67.0					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	8.5	9.0			8.5					± 6.0
GRANULO	#	0.53	0.42			0.49					± 0.40
	%	25.0	21.0			24.5					± 7.0

**LOT** MX 428 N  
 Rev 1

**CONTROL** N

(Exp.)

**2021-06-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.7	7.5			7.7					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.65	4.60			4.61					± 0.18
HB HGB	g/dl	13.4	13.2			13.5					± 0.5
	g/l	134	132			135					± 5
	mmol/l	8.32	8.20			8.38					± 0.31
HT HCT	%	37.2	36.8			37.3					± 2.5
	l/l	0.372	0.368			0.373					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	80	80			81					± 4
TGMH MCH	pg	28.8	28.7			29.3					± 2.0
	fmol	1.79	1.78			1.82					± 0.12
CCMH MCHC	g/dl	36.0	35.9			36.2					± 3.0
	g/l	360	359			362					± 30
	mmol/l	22.37	22.28			22.45					± 1.86
IDR RDW	%	13.7	13.8			13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	252	248			250					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.2	8.1			7.5					± 2.0
LYMPHO	#	2.66	2.63			2.62					± 0.40
	%	34.5	35.0			34.0					± 5.0
MONO	#	0.54	0.60			0.58					± 0.30
	%	7.0	8.0			7.5					± 4.0
GRANULO	#	4.50	4.28			4.50					± 0.60
	%	58.5	57.0			58.5					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060513



**LOT**

**MX 428 H**  
 Rev 1

**CONTROL H**

(Exp.)

**2021-06-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.8</b>	<b>19.8</b>			<b>20.5</b>					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.68</b>	<b>5.55</b>			<b>5.63</b>					± 0.20
HB HGB	g/dl	<b>17.8</b>	<b>17.5</b>			<b>17.9</b>					± 0.6
	g/l	<b>178</b>	<b>175</b>			<b>179</b>					± 6
	mmol/l	<b>11.05</b>	<b>10.87</b>			<b>11.12</b>					± 0.37
HT HCT	%	<b>50.0</b>	<b>48.8</b>			<b>50.1</b>					± 3.0
	l/l	<b>0.500</b>	<b>0.488</b>			<b>0.501</b>					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>88</b>	<b>88</b>			<b>89</b>					± 4
TGMH MCH	pg	<b>31.3</b>	<b>31.5</b>			<b>31.8</b>					± 2.0
	fmol	<b>1.95</b>	<b>1.96</b>			<b>1.97</b>					± 0.12
CCMH MCHC	g/dl	<b>35.6</b>	<b>35.8</b>			<b>35.7</b>					± 3.0
	g/l	<b>356</b>	<b>358</b>			<b>357</b>					± 30
	mmol/l	<b>22.11</b>	<b>22.25</b>			<b>22.18</b>					± 1.86
IDR RDW	%	<b>13.2</b>	<b>13.2</b>			<b>12.0</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>498</b>	<b>490</b>			<b>493</b>					± 45
VPM MPV	µm <sup>3</sup> ; fl	<b>8.1</b>	<b>8.1</b>			<b>7.5</b>					± 2.0
LYMPHO	#	<b>3.43</b>	<b>3.17</b>			<b>3.28</b>					± 1.00
	%	<b>16.5</b>	<b>16.0</b>			<b>16.0</b>					± 5.0
MONO	#	<b>0.94</b>	<b>1.09</b>			<b>0.92</b>					± 0.80
	%	<b>4.5</b>	<b>5.5</b>			<b>4.5</b>					± 4.0
GRANULO	#	<b>16.43</b>	<b>15.54</b>			<b>16.30</b>					± 1.40
	%	<b>79.0</b>	<b>78.5</b>			<b>79.5</b>					± 7.0