



LOT		MX 093 L		CONTROL		L		 (Exp.)		2013-12-05 (YYYY - MM - DD)											
		Rev 1		ABX Lyse																	
PARAMETRES PARAMETERS		UNITES UNITS		MICROS		MICROS		MICROS		MICROS		ARGOS - HELIOS		PENTRA		PENTRA		PENTRA		TOLERANCES TOLERANCE	
				ES60 Care ST		CRP CRP 200		8P		18P		60 8P - 60 18P ADVIA 60				60 60 C+ ES60		80 XL80			
GB WBC		10 ⁹ /mm ³ ; 10 ⁹ /l		1.9		2.0		2.0		2.0		2.0		1.9		2.0				± 0.4	
GR RBC		10 ⁶ /mm ³ ; 10 ¹² /l		2.46		2.41		2.46		2.41		2.43		2.46		2.41				± 0.15	
HB HGB		g/dl		6.2		6.1		6.1		6.0		6.1		6.0		6.0				± 0.4	
		g/l		62		61		61		60		61		60		60				± 4	
		mmol/l		3.85		3.79		3.79		3.73		3.79		3.73		3.73				± 0.25	
HT HCT		%		16.7		16.4		17.0		16.6		16.8		16.7		16.9				± 2.0	
		l/l		0.167		0.164		0.170		0.166		0.168		0.167		0.169				± 0.020	
VGM MCV		µm ³ ; fl		68		68		69		69		69		68		70				± 4	
TGMH MCH		pg		25.2		25.3		24.8		24.9		25.1		24.4		24.9				± 2.0	
		fmol		1.57		1.57		1.54		1.55		1.56		1.51		1.55				± 0.12	
CCMH MCHC		g/dl		37.1		37.2		35.9		36.1		36.4		35.9		35.6				± 3.0	
		g/l		371		372		359		361		364		359		356				± 30	
		mmol/l		23.02		23.12		22.32		22.41		22.59		22.27		22.09				± 1.86	
IDR RDW		%		14.2		13.6		N/A		13.3		14.0		15.2		15.4				± 3.0	
PLAQ. PLTS		10 ³ /mm ³ ; 10 ⁹ /l		75		70		67		70		73		* 62		* 62				± 20 *± 22	
VPM MPV		µm ³ ; fl		9.0		9.0		N/A		9.1		8.3		9.0		9.6				± 3.0	
LYMPHO		#		1.22		1.27		N/A		0.98		1.28		N/A		N/A				± 0.40	
		%		64.0		63.5		N/A		49.0		64.0		N/A		N/A				± 6.0	
MONO		#		0.20		0.22		N/A		0.34		0.22		N/A		N/A				± 0.20	
		%		10.5		11.0		N/A		17.0		11.0		N/A		N/A				± 5.0	
GRANULO		#		0.48		0.51		N/A		0.68		0.50		N/A		N/A				± 0.40	
		%		25.5		25.5		N/A		34.0		25.0		N/A		N/A				± 6.0	

* For ABX Pentra 60 Pentra 80 instruments only



LOT

MX 093 N

Rev 1

CONTROL

N



2013-12-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.6	7.5	7.5	7.4	7.6		7.7	7.7		± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.60	4.52	4.60	4.53	4.57		4.56	4.56		± 0.18
HB HGB	g/dl	13.7	13.4	13.4	13.6	13.5		13.6	13.5		± 0.5
	g/l	137	134	134	136	135		136	135		± 5
	mmol/l	8.51	8.32	8.32	8.45	8.38		8.45	8.38		± 0.31
HT HCT	%	37.3	37.1	37.7	37.1	37.5		36.9	36.9		± 2.5
	l/l	0.373	0.371	0.377	0.371	0.375		0.369	0.369		± 0.025
VGM MCV	µm ³ ; fl	81	82	82	82	82		81	81		± 4
TGMH MCH	pg	29.8	29.6	29.1	30.0	29.5		29.8	29.6		± 2.0
	fmol	1.85	1.84	1.81	1.86	1.83		1.85	1.84		± 0.12
CCMH MCHC	g/dl	36.8	36.2	35.5	36.6	36.0		36.8	36.5		± 3.0
	g/l	368	362	355	366	360		368	365		± 30
	mmol/l	22.83	22.45	22.06	22.74	22.37		22.87	22.70		± 1.86
IDR RDW	%	13.8	13.4	N/A	13.0	13.6		13.5	14.8		± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	265	255	260	258	258		* 255	* 260		± 40 *± 40
VPM MPV	µm ³ ; fl	8.4	8.3	N/A	8.4	7.7		9.6	9.7		± 2.0
LYMPHO	#	2.47	2.40	N/A	1.74	2.47		N/A	N/A		± 0.40
	%	32.5	32.0	N/A	23.5	32.5		N/A	N/A		± 5.0
MONO	#	0.57	0.68	N/A	0.74	0.57		N/A	N/A		± 0.30
	%	7.5	9.0	N/A	10.0	7.5		N/A	N/A		± 4.0
GRANULO	#	4.56	4.43	N/A	4.92	4.56		N/A	N/A		± 0.60
	%	60.0	59.0	N/A	66.5	60.0		N/A	N/A		± 6.0

* For ABX Pentra 60 Pentra 80 instruments only

9930079-B
FRONT / RECTO
Ref TEMP-0830 Rev.32



LOT

MX 093 H

Rev 1

CONTROL

H



(Exp.)

2013-12-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA		
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.2	19.8	19.9	19.7	20.1		20.7	20.5		± 1.6	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.73	5.64	5.70	5.67	5.68		5.64	5.67		± 0.20	
HB HGB	g/dl	17.9	17.6	17.5	18.0	17.8		17.7	17.7		± 0.6	
	g/l	179	176	175	180	178		177	177		± 6	
	mmol/l	11.12	10.93	10.87	11.18	11.05		10.99	10.99		± 0.37	
HT HCT	%	50.4	50.2	50.7	50.5	50.6		49.1	48.8		± 3.0	
	l/l	0.504	0.502	0.507	0.505	0.506		0.491	0.488		± 0.030	
VGM MCV	µm ³ ; fl	88	89	89	89	89		87	86		± 4	
TGMH MCH	pg	31.2	31.2	30.7	31.7	31.3		31.4	31.2		± 2.0	
	fmol	1.94	1.94	1.91	1.97	1.95		1.95	1.94		± 0.12	
CCMH MCHC	g/dl	35.5	35.1	34.5	35.7	35.2		36.1	36.3		± 3.0	
	g/l	355	351	345	357	352		361	363		± 30	
	mmol/l	22.04	21.77	21.42	22.15	21.87		22.40	22.54		± 1.86	
IDR RDW	%	13.4	13.5	N/A	12.4	13.0		12.5	13.8		± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	515	495	518	520	495		* 504	* 515		± 45	*± 55
VPM MPV	µm ³ ; fl	8.2	8.0	N/A	8.1	7.6		9.4	9.5		± 2.0	
LYMPHO	#	3.23	3.07	N/A	2.27	3.12		N/A	N/A		± 1.00	
	%	16.0	15.5	N/A	11.5	15.5		N/A	N/A		± 5.0	
MONO	#	1.11	1.19	N/A	0.89	1.11		N/A	N/A		± 0.80	
	%	5.5	6.0	N/A	4.5	5.5		N/A	N/A		± 4.0	
GRANULO	#	15.86	15.54	N/A	16.55	15.88		N/A	N/A		± 1.40	
	%	78.5	78.5	N/A	84.0	79.0		N/A	N/A		± 7.0	

* For ABX Pentra 60 Pentra 80 instruments only