



LOT

MX 114 L

Rev 1

CONTROL

L



2015-02-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.0	2.0	2.0	2.0				± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.29	2.27	2.30	2.30	2.30				± 0.15
HB HGB	g/dl	6.0	6.0	5.9	5.9	6.0				± 0.4
	g/l	60	60	59	59	60				± 4
	mmol/l	3.73	3.73	3.66	3.66	3.73				± 0.25
HT HCT	%	16.0	15.7	16.1	16.1	16.1				± 2.0
	l/l	0.160	0.157	0.161	0.161	0.161				± 0.020
VGM MCV	µm ³ ; fl	70	69	70	70	70				± 4
TGMH MCH	pg	26.2	26.4	25.7	25.7	26.1				± 2.0
	fmol	1.63	1.64	1.59	1.59	1.62				± 0.12
CCMH MCHC	g/dl	37.4	38.3	36.6	36.6	37.3				± 3.0
	g/l	374	383	366	366	373				± 30
	mmol/l	23.24	23.79	22.76	22.76	23.14				± 1.86
IDR RDW	%	14.0	13.5	N/A	14.0	14.0				± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	85	80	75	79	85				± 20
VPM MPV	µm ³ ; fl	8.5	8.4	N/A	8.6	7.6				± 2.0
LYMPHO	#	1.34	1.31	N/A	1.02	1.28				± 0.40
	%	64.0	65.5	N/A	51.0	64.0				± 8.0
MONO	#	0.20	0.20	N/A	0.33	0.20				± 0.20
	%	9.0	9.5	N/A	16.5	9.0				± 6.0
GRANULO	#	0.57	0.50	N/A	0.65	0.54				± 0.40
	%	27.0	25.0	N/A	32.5	27.0				± 7.0



LOT		MX 114 N		CONTROL N		 (Exp.)		2015-02-05 (YYYY - MM - DD)												
		Rev 1																		
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse										TOLERANCES TOLERANCE								
		MICROS		MICROS		MICROS		MICROS		MICROS			ARGOS - HELIOS		PENTRA		PENTRA		PENTRA	
		ES60 Care ST		CRP CRP 200		8P		18P		60 8P - 60 18P ADVIA 60										
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.8	7.7	7.6	7.5	7.7													± 0.8	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.63	4.59	4.65	4.62	4.62													± 0.18	
HB HGB	g/dl	13.5	13.4	13.4	13.5	13.5													± 0.5	
	g/l	135	134	134	135	135													± 5	
	mmol/l	8.38	8.32	8.32	8.38	8.38													± 0.31	
HT HCT	%	37.5	37.2	37.7	37.4	37.4													± 2.5	
	l/l	0.375	0.372	0.377	0.374	0.374													± 0.025	
VGM MCV	µm ³ ; fl	81	81	81	81	81													± 4	
TGMH MCH	pg	29.2	29.2	28.8	29.2	29.2													± 2.0	
	fmol	1.81	1.81	1.79	1.81	1.81													± 0.12	
CCMH MCHC	g/dl	36.0	36.0	35.6	36.1	36.1													± 3.0	
	g/l	360	360	356	361	361													± 30	
	mmol/l	22.35	22.38	22.09	22.40	22.40													± 1.86	
IDR RDW	%	13.5	13.0	N/A	13.0	13.5													± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	258	255	260	256	255													± 40	
VPM MPV	µm ³ ; fl	8.2	7.9	N/A	8.5	7.4													± 2.0	
LYMPHO	#	2.61	2.54	N/A	1.95	2.54													± 0.40	
	%	33.5	33.0	N/A	26.0	33.0													± 5.0	
MONO	#	0.55	0.62	N/A	0.71	0.58													± 0.30	
	%	7.0	8.0	N/A	9.5	7.5													± 4.0	
GRANULO	#	4.64	4.54	N/A	4.84	4.58													± 0.60	
	%	59.5	59.0	N/A	64.5	59.5													± 6.0	

9930084-B
FRONT / RECTO
Ref: TEMP-0830 Rev.36



LOT

MX 114 H

Rev 1

CONTROL

H



2015-02-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.7	20.6	20.2	20.3	20.7					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.61	5.55	5.61	5.62	5.60					± 0.20
HB HGB	g/dl	17.9	17.7	17.6	17.9	17.8					± 0.6
	g/l	179	177	176	179	178					± 6
	mmol/l	11.12	10.99	10.93	11.12	11.05					± 0.37
HT HCT	%	49.9	49.4	49.9	50.0	49.8					± 3.0
	l/l	0.499	0.494	0.499	0.500	0.498					± 0.030
VGM MCV	µm ³ ; fl	89	89	89	89	89					± 4
TGMH MCH	pg	31.9	31.9	31.4	31.9	31.8					± 2.0
	fmol	1.98	1.98	1.95	1.98	1.97					± 0.12
CCMH MCHC	g/dl	35.9	35.8	35.3	35.8	35.7					± 3.0
	g/l	359	358	353	358	357					± 30
	mmol/l	22.26	22.25	21.89	22.22	22.18					± 1.86
IDR RDW	%	13.0	12.8	N/A	12.5	13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	533	525	550	540	530					± 45
VPM MPV	µm ³ ; fl	7.9	7.7	N/A	8.2	7.2					± 2.0
LYMPHO	#	3.31	3.30	N/A	2.74	3.31					± 1.00
	%	16.0	16.0	N/A	13.5	16.0					± 5.0
MONO	#	0.83	0.82	N/A	0.81	0.83					± 0.80
	%	4.0	4.0	N/A	4.0	4.0					± 4.0
GRANULO	#	16.56	16.48	N/A	16.75	16.56					± 1.40
	%	80.0	80.0	N/A	82.5	80.0					± 7.0