



LOT **MX 430** **L**
 Rev 1

CONTROL **L**



2021-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.30			2.35					± 0.15
HB HGB	g/dl	6.0	6.0			6.0					± 0.4
	g/l	60	60			60					± 4
	mmol/l	3.73	3.73			3.73					± 0.25
HT HCT	%	16.5	16.1			16.5					± 2.0
	l/l	0.165	0.161			0.165					± 0.020
VGM MCV	µm ³ ; fl	70	70			70					± 4
TGMH MCH	pg	25.5	26.1			25.5					± 2.0
	fmol	1.59	1.62			1.59					± 0.12
CCMH MCHC	g/dl	36.5	37.3			36.5					± 3.0
	g/l	365	373			365					± 30
	mmol/l	22.65	23.14			22.65					± 1.86
IDR RDW	%	14.5	14.0			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	77	75			78					± 20
VPM MPV	µm ³ ; fl	9.3	9.3			8.8					± 2.0
LYMPHO	#	1.36	1.41			1.35					± 0.40
	%	68.0	70.5			67.5					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	7.0	7.5			7.5					± 6.0
GRANULO	#	0.50	0.44			0.50					± 0.40
	%	25.0	22.0			25.0					± 7.0

LOT **MX 430** **N**
 Rev 1

CONTROL **N**

(Exp.)

2021-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.9	7.8			8.0					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.60	4.55			4.62					± 0.18
HB HGB	g/dl	13.5	13.5			13.6					± 0.5
	g/l	135	135			136					± 5
	mmol/l	8.38	8.38			8.45					± 0.31
HT HCT	%	37.7	37.3			37.4					± 2.5
	l/l	0.377	0.373			0.374					± 0.025
VGM MCV	µm ³ ; fl	82	82			81					± 4
TGMH MCH	pg	29.3	29.7			29.4					± 2.0
	fmol	1.82	1.84			1.83					± 0.12
CCMH MCHC	g/dl	35.8	36.2			36.3					± 3.0
	g/l	358	362			363					± 30
	mmol/l	22.23	22.47			22.57					± 1.86
IDR RDW	%	14.5	14.0			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	258	253			260					± 40
VPM MPV	µm ³ ; fl	8.4	8.3			7.8					± 2.0
LYMPHO	#	2.92	2.96			2.92					± 0.40
	%	37.0	38.0			36.5					± 5.0
MONO	#	0.47	0.55			0.48					± 0.30
	%	6.0	7.0			6.0					± 4.0
GRANULO	#	4.50	4.29			4.60					± 0.60
	%	57.0	55.0			57.5					± 6.0

Ref: TEMP-0830 Rev.41 FRONT / RECTO 130090307

LOT		MX 430 H		CONTROL H						(Exp.) 2021-10-05 (YYYY - MM - DD)	
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.7	20.3			21.0					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.61	5.55			5.64					± 0.20
HB HGB	g/dl	17.8	17.7			18.0					± 0.6
	g/l	178	177			180					± 6
	mmol/l	11.05	10.99			11.18					± 0.37
HT HCT	%	49.9	49.4			50.2					± 3.0
	l/l	0.499	0.494			0.502					± 0.030
VGM MCV	µm ³ ; fl	89	89			89					± 4
TGMH MCH	pg	31.7	31.9			31.9					± 2.0
	fmol	1.97	1.98			1.98					± 0.12
CCMH MCHC	g/dl	35.7	35.8			35.9					± 3.0
	g/l	357	358			359					± 30
	mmol/l	22.14	22.25			22.27					± 1.86
IDR RDW	%	13.5	13.5			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	501	490			500					± 45
VPM MPV	µm ³ ; fl	8.0	8.0			7.5					± 2.0
LYMPHO	#	3.52	3.45			3.36					± 1.00
	%	17.0	17.0			16.0					± 5.0
MONO	#	1.04	1.32			0.95					± 0.80
	%	5.0	6.5			4.5					± 4.0
GRANULO	#	16.15	15.53			16.70					± 1.40
	%	78.0	76.5			79.5					± 7.0