

LOT		MX 413 L		CONTROL L						(Exp.)	2018-12-05 (YYYY-MM-DD)
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.1	2.1	2.1	2.1					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.27	2.29	2.31	2.30	2.27					± 0.15
HB HGB	g/dl	5.9	5.9	5.8	5.9	5.9					± 0.4
	g/l	59	59	58	59	59					± 4
	mmol/l	3.66	3.66	3.60	3.66	3.66					± 0.25
HT HCT	%	15.7	15.8	16.2	16.1	15.9					± 2.0
	l/l	0.157	0.158	0.162	0.161	0.159					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	69	69	70	70	70					± 4
TGMH MCH	pg	26.0	25.8	25.1	25.7	26.0					± 2.0
	fmol	1.61	1.60	1.56	1.59	1.61					± 0.12
CCMH MCHC	g/dl	37.7	37.3	35.9	36.6	37.1					± 3.0
	g/l	377	373	359	366	371					± 30
	mmol/l	23.39	23.19	22.27	22.76	23.06					± 1.86
IDR RDW	%	14.5	13.5	N/A	14.5	13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	77	80	69	76	77					± 20
VPM MPV	µm <sup>3</sup> ; fl	9.0	8.8	N/A	9.1	8.2					± 2.0
LYMPHO	#	1.31	1.41	N/A	1.07	1.35					± 0.40
	%	65.5	67.0	N/A	51.0	64.5					± 8.0
MONO	#	0.20	0.20	N/A	0.36	0.21					± 0.20
	%	8.5	9.5	N/A	17.0	10.0					± 6.0
GRANULO	#	0.52	0.49	N/A	0.67	0.54					± 0.40
	%	26.0	23.5	N/A	32.0	25.5					± 7.0

**LOT** MX 413 N  
 Rev 1

**CONTROL** N

(Exp.)

**2018-12-05**  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.5	7.5	7.5	7.5	7.6					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.53	4.54	4.57	4.58	4.55					± 0.18
HB HGB	g/dl	13.3	13.4	13.4	13.5	13.5					± 0.5
	g/l	133	134	134	135	135					± 5
	mmol/l	8.26	8.32	8.32	8.38	8.38					± 0.31
HT HCT	%	36.2	36.8	36.6	36.6	36.9					± 2.5
	l/l	0.362	0.368	0.366	0.366	0.369					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	80	81	80	80	81					± 4
TGMH MCH	pg	29.4	29.5	29.3	29.5	29.7					± 2.0
	fmol	1.82	1.83	1.82	1.83	1.84					± 0.12
CCMH MCHC	g/dl	36.7	36.4	36.7	36.8	36.6					± 3.0
	g/l	367	364	367	368	366					± 30
	mmol/l	22.79	22.63	22.76	22.88	22.75					± 1.86
IDR RDW	%	14.5	13.5	N/A	14.5	13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	262	262	253	260	260					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.6	8.4	N/A	9.0	7.9					± 2.0
LYMPHO	#	2.63	2.70	N/A	2.10	2.62					± 0.40
	%	35.0	36.0	N/A	28.0	34.5					± 5.0
MONO	#	0.45	0.45	N/A	0.71	0.53					± 0.30
	%	6.0	6.0	N/A	9.5	7.0					± 4.0
GRANULO	#	4.43	4.35	N/A	4.69	4.45					± 0.60
	%	59.0	58.0	N/A	62.5	58.5					± 6.0

Ref: TEMP-0830 Rev.36 FRONT / RECTO 1219930084-B

<b>LOT</b>		<b>MX 413 H</b>	<b>CONTROL H</b>				(Exp.)				<b>2018-12-05</b> (YYYY - MM - DD)			
Rev 1		<b>ABX Lyse</b>												
PARAMETRES PARAMETERS	UNITES UNITS	MICROS		MICROS		MICROS		MICROS		ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.3</b>	<b>20.3</b>	<b>20.0</b>	<b>20.1</b>	<b>20.6</b>							± 1.6	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.60</b>	<b>5.56</b>	<b>5.59</b>	<b>5.63</b>	<b>5.55</b>							± 0.20	
HB HGB	g/dl	<b>17.7</b>	<b>17.5</b>	<b>17.7</b>	<b>17.9</b>	<b>17.8</b>							± 0.6	
	g/l	<b>177</b>	<b>175</b>	<b>177</b>	<b>179</b>	<b>178</b>							± 6	
	mmol/l	<b>10.99</b>	<b>10.87</b>	<b>10.99</b>	<b>11.12</b>	<b>11.05</b>							± 0.37	
HT HCT	%	<b>49.3</b>	<b>49.5</b>	<b>49.8</b>	<b>49.5</b>	<b>49.4</b>							± 3.0	
	l/l	<b>0.493</b>	<b>0.495</b>	<b>0.498</b>	<b>0.495</b>	<b>0.494</b>							± 0.030	
VGM MCV	µm <sup>3</sup> ; fl	<b>88</b>	<b>89</b>	<b>89</b>	<b>88</b>	<b>89</b>							± 4	
TGMH MCH	pg	<b>31.6</b>	<b>31.5</b>	<b>31.7</b>	<b>31.8</b>	<b>32.1</b>							± 2.0	
	fmol	<b>1.96</b>	<b>1.95</b>	<b>1.97</b>	<b>1.97</b>	<b>1.99</b>							± 0.12	
CCMH MCHC	g/dl	<b>35.9</b>	<b>35.4</b>	<b>35.6</b>	<b>36.1</b>	<b>36.0</b>							± 3.0	
	g/l	<b>359</b>	<b>354</b>	<b>356</b>	<b>361</b>	<b>360</b>							± 30	
	mmol/l	<b>22.30</b>	<b>21.96</b>	<b>22.09</b>	<b>22.44</b>	<b>22.38</b>							± 1.86	
IDR RDW	%	<b>13.5</b>	<b>13.0</b>	N/A	<b>13.5</b>	<b>12.5</b>							± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>507</b>	<b>513</b>	<b>500</b>	<b>515</b>	<b>510</b>							± 55	
VPM MPV	µm <sup>3</sup> ; fl	<b>8.3</b>	<b>8.2</b>	N/A	<b>8.8</b>	<b>7.7</b>							± 2.0	
LYMPHO	#	<b>3.45</b>	<b>3.35</b>	N/A	<b>2.61</b>	<b>3.30</b>							± 1.00	
	%	<b>17.0</b>	<b>16.5</b>	N/A	<b>13.0</b>	<b>16.0</b>							± 5.0	
MONO	#	<b>1.22</b>	<b>1.32</b>	N/A	<b>1.01</b>	<b>1.24</b>							± 0.80	
	%	<b>6.0</b>	<b>6.5</b>	N/A	<b>5.0</b>	<b>6.0</b>							± 4.0	
GRANULO	#	<b>15.63</b>	<b>15.63</b>	N/A	<b>16.48</b>	<b>16.07</b>							± 1.40	
	%	<b>77.0</b>	<b>77.0</b>	N/A	<b>82.0</b>	<b>78.0</b>							± 7.0	