

LOT MX 428 L
Rev 1

CONTROL L

 (Exp.)

2021-06-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-----------------|---------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP 200 | | | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.1 | 2.0 | | | 2.0 | | | | | ± 0.4 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.38 | 2.34 | | | 2.38 | | | | | ± 0.15 |
| HB HGB | g/dl | 6.0 | 5.9 | | | 6.1 | | | | | ± 0.4 |
| | g/l | 69 | 59 | | | 61 | | | | | ± 4 |
| | mmol/l | 3.73 | 3.66 | | | 3.79 | | | | | ± 0.25 |
| HT HCT | % | 16.4 | 15.9 | | | 16.4 | | | | | ± 2.0 |
| | l/l | 0.164 | 0.159 | | | 0.164 | | | | | ± 0.020 |
| VGM MCV | µm ³ ; fl | 69 | 68 | | | 69 | | | | | ± 4 |
| TGMH MCH | pg | 25.2 | 25.2 | | | 25.6 | | | | | ± 2.0 |
| | fmol | 1.57 | 1.57 | | | 1.59 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.5 | 37.1 | | | 37.1 | | | | | ± 3.0 |
| | g/l | 365 | 371 | | | 371 | | | | | ± 30 |
| | mmol/l | 22.69 | 23.03 | | | 23.07 | | | | | ± 1.86 |
| IDR RDW | % | 14.5 | 14.5 | | | 13.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 77 | 81 | | | 77 | | | | | ± 20 |
| VPM MPV | µm ³ ; fl | 9.1 | 8.6 | | | 8.1 | | | | | ± 2.0 |
| LYMPHO | # | 1.40 | 1.40 | | | 1.34 | | | | | ± 0.40 |
| | % | 66.5 | 70.0 | | | 67.0 | | | | | ± 8.0 |
| MONO | # | 0.20 | 0.20 | | | 0.20 | | | | | ± 0.20 |
| | % | 8.5 | 9.0 | | | 8.5 | | | | | ± 6.0 |
| GRANULO | # | 0.53 | 0.42 | | | 0.49 | | | | | ± 0.40 |
| | % | 25.0 | 21.0 | | | 24.5 | | | | | ± 7.0 |

LOT MX 428 N
Rev 1

CONTROL N

 (Exp.)

2021-06-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-----------------|---------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP 200 | | | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 7.7 | 7.5 | | | 7.7 | | | | | ± 0.8 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 4.65 | 4.60 | | | 4.61 | | | | | ± 0.18 |
| HB HGB | g/dl | 13.4 | 13.2 | | | 13.5 | | | | | ± 0.5 |
| | g/l | 134 | 132 | | | 135 | | | | | ± 5 |
| | mmol/l | 8.32 | 8.20 | | | 8.38 | | | | | ± 0.31 |
| HT HCT | % | 37.2 | 36.8 | | | 37.3 | | | | | ± 2.5 |
| | l/l | 0.372 | 0.368 | | | 0.373 | | | | | ± 0.025 |
| VGM MCV | µm ³ ; fl | 80 | 80 | | | 81 | | | | | ± 4 |
| TGMH MCH | pg | 28.8 | 28.7 | | | 29.3 | | | | | ± 2.0 |
| | fmol | 1.79 | 1.78 | | | 1.82 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.0 | 35.9 | | | 36.2 | | | | | ± 3.0 |
| | g/l | 360 | 359 | | | 362 | | | | | ± 30 |
| | mmol/l | 22.37 | 22.28 | | | 22.45 | | | | | ± 1.86 |
| IDR RDW | % | 13.7 | 13.8 | | | 13.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 252 | 248 | | | 250 | | | | | ± 40 |
| VPM MPV | µm ³ ; fl | 8.2 | 8.1 | | | 7.5 | | | | | ± 2.0 |
| LYMPHO | # | 2.66 | 2.63 | | | 2.62 | | | | | ± 0.40 |
| | % | 34.5 | 35.0 | | | 34.0 | | | | | ± 5.0 |
| MONO | # | 0.54 | 0.60 | | | 0.58 | | | | | ± 0.30 |
| | % | 7.0 | 8.0 | | | 7.5 | | | | | ± 4.0 |
| GRANULO | # | 4.50 | 4.28 | | | 4.50 | | | | | ± 0.60 |
| | % | 58.5 | 57.0 | | | 58.5 | | | | | ± 6.0 |

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060513



LOT MX 428 H
 Rev 1

CONTROL H

(Exp.)

2021-06-05
 (YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-----------------|--------------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP 200 | | | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 20.8 | 19.8 | | | 20.5 | | | | | ± 1.6 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 5.68 | 5.55 | | | 5.63 | | | | | ± 0.20 |
| HB HGB | g/dl | 17.8 | 17.5 | | | 17.9 | | | | | ± 0.6 |
| | g/l | 178 | 175 | | | 179 | | | | | ± 6 |
| | mmol/l | 11.05 | 10.87 | | | 11.12 | | | | | ± 0.37 |
| HT HCT | % | 50.0 | 48.8 | | | 50.1 | | | | | ± 3.0 |
| | l/l | 0.500 | 0.488 | | | 0.501 | | | | | ± 0.030 |
| VGM MCV | µm ³ ; fl | 88 | 88 | | | 89 | | | | | ± 4 |
| TGMH MCH | pg | 31.3 | 31.5 | | | 31.8 | | | | | ± 2.0 |
| | fmol | 1.95 | 1.96 | | | 1.97 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 35.6 | 35.8 | | | 35.7 | | | | | ± 3.0 |
| | g/l | 356 | 358 | | | 357 | | | | | ± 30 |
| | mmol/l | 22.11 | 22.25 | | | 22.18 | | | | | ± 1.86 |
| IDR RDW | % | 13.2 | 13.2 | | | 12.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 498 | 490 | | | 493 | | | | | ± 45 |
| VPM MPV | µm ³ ; fl | 8.1 | 8.1 | | | 7.5 | | | | | ± 2.0 |
| LYMPHO | # | 3.43 | 3.17 | | | 3.28 | | | | | ± 1.00 |
| | % | 16.5 | 16.0 | | | 16.0 | | | | | ± 5.0 |
| MONO | # | 0.94 | 1.09 | | | 0.92 | | | | | ± 0.80 |
| | % | 4.5 | 5.5 | | | 4.5 | | | | | ± 4.0 |
| GRANULO | # | 16.43 | 15.54 | | | 16.30 | | | | | ± 1.40 |
| | % | 79.0 | 78.5 | | | 79.5 | | | | | ± 7.0 |