

ABX Minotrol 16



(Exp.) 2012-02-05
 (YYYY - MM - DD)

LOT		MX 111 L		CONTROL		L							
Rev		1											
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE		
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA			
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX 120			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	1.9	1.9	2.0	1.9	1.9		1.9	1.9	1.9	± 0.4		
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.42	2.42	2.44	2.45	2.42		2.43	2.36	2.42	± 0.15		
HB HGB	g/dl	6.2	6.1	6.0	6.0	6.1		6.0	6.1	6.0	± 0.4		
	g/l	62	61	60	60	61		60	61	60	± 4		
	mmol/l	3.86	3.79	3.73	3.73	3.79		3.73	3.79	3.73	± 0.25		
HT HCT	%	17.1	16.7	17.3	17.2	16.9		16.8	17.2	17.2	± 2.0		
	l/l	0.171	0.167	0.173	0.172	0.169		0.168	0.172	0.172	± 0.020		
VGM MCV	µm ³ ; fl	71	69	71	70	70		69	73	71	± 4		
TGMH MCH	pg	25.7	25.2	24.6	24.5	25.2		24.7	25.8	24.8	± 2.0		
	f mol	1.60	1.57	1.53	1.52	1.57		1.53	1.61	1.54	± 0.12		
CCMH MCHC	g/dl	36.4	36.5	34.6	35.0	36.0		35.8	35.4	34.9	± 3.0		
	g/l	364	365	346	350	360		358	354	349	± 30		
	mmol/l	22.58	22.69	21.51	21.73	22.36		22.22	21.99	21.69	± 1.86		
IDR RDW	%	14.3	13.5	N/A	13.8	14.0		14.5	14.5	15.8	± 3.0		
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	82	78	75	77	80		* 80	* 75	78	± 20 *± 22		
VPM MPV	µm ³ ; fl	7.9	8.5	N/A	8.6	8.2		9.4	9.5	9.2	± 2.0		
LYMPHO	#	1.19	1.2	N/A	0.9	1.2		N/A	N/A	N/A	± 0.40		
	%	61.6	63.0	N/A	47.0	63.5		N/A	N/A	N/A	± 9.0		
MONO	#	0.21	0.2	N/A	0.3	0.2		N/A	N/A	N/A	± 0.20		
	%	11.0	12.5	N/A	17.0	10.0		N/A	N/A	N/A	± 7.0		
GRANULO	#	0.53	0.5	N/A	0.7	0.5		N/A	N/A	N/A	± 0.40		
	%	27.3	24.5	N/A	36.0	26.5		N/A	N/A	N/A	± 8.0		

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.16 FRONT / RECTO 9930054-A



ABX Minotrol 16



LOT MX 111 N
 Rev 1

CONTROL N

(Exp.) 2012-02-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 CareST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.7	7.6	7.5	7.6		7.7	7.7	7.9	± 0.8
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	4.68	4.72	4.72	4.72	4.73		4.71	4.70	4.73	± 0.18
HB HGB	g/dl	13.5	13.4	13.5	13.5	13.5		13.5	13.5	13.6	± 0.5
	g/l	135	134	135	135	135		135	135	136	± 5
	mmol/l	8.38	8.32	8.38	8.38	8.38		8.38	8.38	8.45	± 0.31
HT HCT	%	38.3	38.2	38.2	38.2	38.3		37.7	38.1	37.8	± 2.5
	l/l	0.383	0.382	0.382	0.382	0.383		0.377	0.381	0.378	± 0.025
VGM MCV	µm ³ ; fl	82	81	81	81	81		80	81	80	± 4
TGMH MCH	pg	28.8	28.4	28.6	28.6	28.5		28.7	28.7	28.8	± 2.0
	f mol	1.79	1.76	1.78	1.78	1.77		1.78	1.78	1.79	± 0.12
CCMH MCHC	g/dl	35.2	35.0	35.3	35.3	35.2		35.8	35.5	35.9	± 3.0
	g/l	352	350	353	353	352		358	355	359	± 30
	mmol/l	21.88	21.77	21.93	21.93	21.88		22.25	22.02	22.32	± 1.86
IDR RDW	%	13.6	13.0	N/A	13.3	13.5		13.3	14.0	15.7	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	253	255	254	252	256		* 253	* 254	255	± 40 ± 40
VPM MPV	µm ³ ; fl	7.7	7.4	N/A	7.8	7.8		8.2	8.3	8.5	± 2.0
LYMPHO	#	2.48	2.5	N/A	1.8	2.5		N/A	N/A	N/A	± 0.60
	%	32.3	32.0	N/A	24.5	33.0		N/A	N/A	N/A	± 6.0
MONO	#	0.53	0.6	N/A	0.7	0.5		N/A	N/A	N/A	± 0.40
	%	6.9	8.0	N/A	9.5	7.0		N/A	N/A	N/A	± 5.0
GRANULO	#	4.65	4.6	N/A	5.0	4.6		N/A	N/A	N/A	± 0.60
	%	60.6	60.0	N/A	66.0	60.0		N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

9930055-A
 Ref: FORM-0830 Rev.16 FRONT / RECTO



ABX Minotrol 16



LOT MX 111 H
Rev 1

CONTROL H

(Exp.) 2012-02-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.4	20.0	19.7	19.6	20.0		20.8	20.5	21.0	± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.67	5.65	5.64	5.60	5.67		5.57	5.60	5.67	± 0.20
HB HGB	g/dl	18.0	17.7	17.8	17.8	18.0		17.8	18.1	17.8	± 0.6
	g/l	180	177	178	178	180		178	181	178	± 6
	mmol/l	11.18	10.99	11.05	11.05	11.18		11.05	11.24	11.05	± 0.37
HT HCT	%	51.0	50.3	50.8	50.4	51.0		49.0	49.3	49.3	± 3.0
	l/l	0.510	0.503	0.508	0.504	0.510		0.490	0.493	0.493	± 0.030
VGM MCV	µm ³ ; fl	90	89	90	90	90		88	88	87	± 4
TGMH MCH	pg	31.7	31.3	31.6	31.8	31.7		32.0	32.3	31.4	± 2.0
	f mol	1.97	1.95	1.96	1.97	1.97		1.98	2.01	1.95	± 0.12
CCMH MCHC	g/dl	35.3	35.2	35.1	35.3	35.3		36.3	36.7	36.1	± 3.0
	g/l	353	352	351	353	353		363	367	361	± 30
	mmol/l	21.90	21.86	21.78	21.93	21.90		22.55	22.81	22.41	± 1.86
IDR RDW	%	13.0	12.5	N/A	12.9	13.0		12.5	13.5	14.8	± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	495	500	522	510	500		* 504	* 511	510	± 45 *± 55
VPM MPV	µm ³ ; fl	7.8	7.6	N/A	7.8	8.0		8.2	8.5	8.7	± 2.0
LYMPHO	#	3.20	3.0	N/A	2.2	3.2		N/A	N/A	N/A	± 1.00
	%	15.7	15.0	N/A	11.0	16.0		N/A	N/A	N/A	± 5.0
MONO	#	0.73	0.9	N/A	0.8	0.8		N/A	N/A	N/A	± 0.80
	%	3.6	4.5	N/A	4.0	4.0		N/A	N/A	N/A	± 4.0
GRANULO	#	16.48	16.1	N/A	16.7	16.0		N/A	N/A	N/A	± 1.40
	%	80.8	80.5	N/A	85.0	80.0		N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.16 FRONT / RECTO 9930056-A