

LOT

MX 055 L
Rev 1

CONTROL L

 (Exp.)

2015-08-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.1	2.1	2.1	2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.40	2.44	2.39	2.40					± 0.15
HB HGB	g/dl	6.1	6.1	6.1	6.0	6.1					± 0.4
	g/l	61	61	61	60	61					± 4
	mmol/l	3.79	3.79	3.79	3.73	3.79					± 0.25
HT HCT	%	16.6	16.3	17.1	16.7	16.6					± 2.0
	l/l	0.166	0.163	0.171	0.167	0.166					± 0.020
VGM MCV	µm ³ ; fl	69	68	70	70	69					± 4
TGMH MCH	pg	25.4	25.4	25.0	25.1	25.4					± 2.0
	fmol	1.58	1.58	1.55	1.56	1.58					± 0.12
CCMH MCHC	g/dl	36.8	37.4	35.7	35.9	36.8					± 3.0
	g/l	368	374	357	359	368					± 30
	mmol/l	22.88	23.21	22.18	22.27	22.88					± 1.86
IDR RDW	%	14.4	14.0	N/A	14.3	13.9					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	74	72	72	76					± 20
VPM MPV	µm ³ ; fl	9.4	9.2	N/A	9.3	8.5					± 2.0
LYMPHO	#	1.34	1.32	N/A	1.09	1.32					± 0.40
	%	64.0	63.0	N/A	52.0	63.0					± 8.0
MONO	#	0.25	0.26	N/A	0.32	0.25					± 0.20
	%	12.0	12.5	N/A	15.0	12.0					± 6.0
GRANULO	#	0.50	0.51	N/A	0.69	0.53					± 0.40
	%	24.0	24.5	N/A	33.0	25.0					± 7.0

LOT		MX 055	N	CONTROL			N	(Exp.)				2015-08-05 (YYYY-MM-DD)
Rev 1		ABX Lyse										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.8	7.8	7.8	7.6	7.9					± 0.8	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.66	4.62	4.65	4.57	4.64					± 0.18	
HB HGB	g/dl	13.6	13.5	13.6	13.4	13.5					± 0.5	
	g/l	136	135	136	134	135					± 5	
	mmol/l	8.45	8.38	8.45	8.32	8.38					± 0.31	
HT HCT	%	37.7	37.9	38.1	37.5	38.0					± 2.5	
	l/l	0.377	0.379	0.381	0.375	0.380					± 0.025	
VGM MCV	µm ³ ; fl	81	82	82	82	82					± 4	
TGMH MCH	pg	29.2	29.2	29.2	29.3	29.1					± 2.0	
	fmol	1.81	1.81	1.82	1.82	1.81					± 0.12	
CCMH MCHC	g/dl	36.0	35.6	35.7	35.8	35.5					± 3.0	
	g/l	360	356	357	358	355					± 30	
	mmol/l	22.37	22.13	22.15	22.21	22.03					± 1.86	
IDR RDW	%	13.9	13.2	N/A	13.7	13.6					± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	265	260	265	262	265					± 40	
VPM MPV	µm ³ ; fl	8.3	7.9	N/A	8.4	7.7					± 2.0	
LYMPHO	#	2.57	2.54	N/A	2.05	2.49					± 0.40	
	%	33.0	32.5	N/A	27.0	31.5					± 5.0	
MONO	#	0.62	0.62	N/A	0.61	0.63					± 0.30	
	%	8.0	8.0	N/A	8.0	8.0					± 4.0	
GRANULO	#	4.60	4.64	N/A	4.94	4.78					± 0.60	
	%	59.0	59.5	N/A	65.0	60.5					± 6.0	

Ret: TEMP-0830 Rev.36 FRONT / RECTO 9950084-B

LOT		MX 055 H		CONTROL H						(Exp.)	2015-08-05 (YYYY-MM-DD)
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.5	20.4	20.0	20.1	20.6					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.76	5.73	5.82	5.72	5.79					± 0.20
HB HGB	g/dl	18.4	18.1	18.2	18.3	18.3					± 0.6
	g/l	184	181	182	183	183					± 6
	mmol/l	11.43	11.24	11.30	11.36	11.36					± 0.37
HT HCT	%	51.3	51.0	51.8	50.9	51.5					± 3.0
	l/l	0.513	0.510	0.518	0.509	0.515					± 0.030
VGM MCV	µm ³ ; fl	89	89	89	89	89					± 4
TGMH MCH	pg	31.9	31.6	31.3	32.0	31.6					± 2.0
	fmol	1.98	1.96	1.94	1.99	1.96					± 0.12
CCMH MCHC	g/dl	35.9	35.5	35.1	35.9	35.5					± 3.0
	g/l	359	355	351	359	355					± 30
	mmol/l	22.29	22.04	21.82	22.32	22.05					± 1.86
IDR RDW	%	13.5	12.7	N/A	13.0	13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	512	500	530	518	515					± 45
VPM MPV	µm ³ ; fl	8.6	8.2	N/A	8.8	8.0					± 2.0
LYMPHO	#	3.59	3.37	N/A	2.81	3.30					± 1.00
	%	17.5	16.5	N/A	14.0	16.0					± 5.0
MONO	#	1.03	1.12	N/A	0.80	1.13					± 0.80
	%	5.0	5.5	N/A	4.0	5.5					± 4.0
GRANULO	#	15.89	15.91	N/A	16.48	16.17					± 1.40
	%	77.5	78.0	N/A	82.0	78.5					± 7.0