

**LOT** MX 075 L  
Rev 1

**CONTROL** L

 (Exp.)

2015-10-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.1	2.1	2.0	2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.41	2.40	2.44	2.37	2.40					± 0.15
HB HGB	g/dl	6.1	6.0	6.1	6.0	6.1					± 0.4
	g/l	61	60	61	60	61					± 4
	mmol/l	3.79	3.73	3.79	3.73	3.79					± 0.25
HT HCT	%	16.6	16.3	17.1	16.6	16.6					± 2.0
	l/l	0.166	0.163	0.171	0.166	0.166					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	69	68	70	70	69					± 4
TGMH MCH	pg	25.3	25.0	25.0	25.3	25.4					± 2.0
	fmol	1.57	1.55	1.55	1.57	1.58					± 0.12
CCMH MCHC	g/dl	36.7	36.8	35.7	36.2	36.8					± 3.0
	g/l	367	368	357	362	368					± 30
	mmol/l	22.78	22.83	22.18	22.46	22.88					± 1.86
IDR RDW	%	14.4	14.0	N/A	14.0	14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	79	70	73	68	75					± 20
VPM MPV	µm <sup>3</sup> ; fl	9.5	9.0	N/A	9.1	8.2					± 2.0
LYMPHO	#	1.26	1.34	N/A	1.02	1.25					± 0.40
	%	63.0	64.0	N/A	51.0	62.5					± 8.0
MONO	#	0.25	0.26	N/A	0.34	0.26					± 0.20
	%	12.5	12.5	N/A	17.0	13.0					± 6.0
GRANULO	#	0.49	0.49	N/A	0.64	0.49					± 0.40
	%	24.5	23.5	N/A	32.0	24.5					± 7.0

**LOT** MX 075 N  
 Rev 1

**CONTROL** N

(Exp.)

**2015-10-05**  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.5	7.5	7.4	7.4	7.5					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.60	4.60	4.60	4.54	4.60					± 0.18
HB HGB	g/dl	13.5	13.4	13.5	13.4	13.5					± 0.5
	g/l	135	134	135	134	135					± 5
	mmol/l	8.38	8.32	8.38	8.32	8.38					± 0.31
HT HCT	%	36.8	36.8	36.8	36.8	36.8					± 2.5
	l/l	0.368	0.368	0.368	0.368	0.368					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	80	80	80	81	80					± 4
TGMH MCH	pg	29.3	29.1	29.3	29.5	29.3					± 2.0
	fmol	1.82	1.81	1.82	1.83	1.82					± 0.12
CCMH MCHC	g/dl	36.7	36.4	36.7	36.4	36.7					± 3.0
	g/l	367	364	367	364	367					± 30
	mmol/l	22.78	22.61	22.78	22.63	22.78					± 1.86
IDR RDW	%	13.8	13.5	N/A	13.7	13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	260	250	258	245	252					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.9	8.5	N/A	8.8	8.0					± 2.0
LYMPHO	#	2.51	2.51	N/A	2.00	2.48					± 0.40
	%	33.5	33.5	N/A	27.0	33.0					± 5.0
MONO	#	0.60	0.64	N/A	0.63	0.60					± 0.30
	%	8.0	8.5	N/A	8.5	8.0					± 4.0
GRANULO	#	4.39	4.35	N/A	4.77	4.43					± 0.60
	%	58.5	58.0	N/A	64.5	59.0					± 6.0

Ref: TEMP-0830 Rev.37 FRONT / RECTO 1219930084-B

**LOT** MX 075 H  
 Rev 1

**CONTROL** H

(Exp.)

**2015-10-05**  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	19.8	20.1	19.8	19.9	20.0					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.74	5.70	5.70	5.65	5.70					± 0.20
HB HGB	g/dl	17.9	17.8	17.8	18.0	17.8					± 0.6
	g/l	179	178	178	180	178					± 6
	mmol/l	11.12	11.05	11.05	11.18	11.05					± 0.37
HT HCT	%	50.5	50.2	50.2	49.7	50.2					± 3.0
	l/l	0.505	0.502	0.502	0.497	0.502					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	88	88	88	88	88					± 4
TGMH MCH	pg	31.2	31.2	31.2	31.9	31.2					± 2.0
	fmol	1.94	1.94	1.94	1.98	1.94					± 0.12
CCMH MCHC	g/dl	35.4	35.5	35.5	36.2	35.5					± 3.0
	g/l	354	355	355	362	355					± 30
	mmol/l	22.01	22.04	22.04	22.48	22.04					± 1.86
IDR RDW	%	13.2	13.0	N/A	13.2	13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	530	520	542	515	520					± 45
VPM MPV	µm <sup>3</sup> ; fl	8.5	8.0	N/A	8.4	7.7					± 2.0
LYMPHO	#	2.97	2.91	N/A	2.39	2.80					± 1.00
	%	15.0	14.5	N/A	12.0	14.0					± 5.0
MONO	#	0.99	1.11	N/A	0.80	1.00					± 0.80
	%	5.0	5.5	N/A	4.0	5.0					± 4.0
GRANULO	#	15.84	16.08	N/A	16.72	16.20					± 1.40
	%	80.0	80.0	N/A	84.0	81.0					± 7.0