

LOT

MX 429 L
Rev 1

CONTROL L

(Exp.)

2021-08-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.27	2.24			2.28					± 0.15
HB HGB	g/dl	6.1	6.0			6.1					± 0.4
	g/l	61	60			61					± 4
	mmol/l	3.79	3.73			3.79					± 0.25
HT HCT	%	16.6	16.4			16.9					± 2.0
	l/l	0.166	0.164			0.169					± 0.020
VGM MCV	µm ³ ; fl	73	73			74					± 4
TGMH MCH	pg	26.9	26.8			26.8					± 2.0
	fmol	1.67	1.66			1.66					± 0.12
CCMH MCHC	g/dl	36.8	36.7			36.2					± 3.0
	g/l	368	367			362					± 30
	mmol/l	22.86	22.79			22.45					± 1.86
IDR RDW	%	15.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	70			73					± 20
VPM MPV	µm ³ ; fl	8.2	8.1			7.6					± 2.0
LYMPHO	#	1.34	1.38			1.34					± 0.40
	%	67.0	69.0			67.0					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	8.0	8.0			7.5					± 6.0
GRANULO	#	0.50	0.46			0.51					± 0.40
	%	25.0	23.0			25.5					± 7.0

LOT		MX 429 N		CONTROL N						2021-08-05 (YYYY-MM-DD)	
Rev 1											
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.8	7.6			7.8					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.55	4.50			4.55					± 0.18
HB HGB	g/dl	13.4	13.3			13.5					± 0.5
	g/l	134	133			135					± 5
	mmol/l	8.32	8.26			8.38					± 0.31
HT HCT	%	37.8	37.4			37.8					± 2.5
	l/l	0.378	0.374			0.378					± 0.025
VGM MCV	µm ³ ; fl	83	83			83					± 4
TGMH MCH	pg	29.5	29.6			29.7					± 2.0
	fmol	1.83	1.84			1.84					± 0.12
CCMH MCHC	g/dl	35.5	35.6			35.7					± 3.0
	g/l	355	356			357					± 30
	mmol/l	22.03	22.11			22.20					± 1.86
IDR RDW	%	14.0	13.5			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	251	243			246					± 40
VPM MPV	µm ³ ; fl	8.1	7.9			7.5					± 2.0
LYMPHO	#	2.73	2.70			2.73					± 0.40
	%	35.0	35.5			35.0					± 5.0
MONO	#	0.62	0.65			0.62					± 0.30
	%	8.0	8.5			8.0					± 4.0
GRANULO	#	4.45	4.26			4.45					± 0.60
	%	57.0	56.0			57.0					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060513

LOT

MX 429 H
Rev 1

CONTROL H

 (Exp.)

2021-08-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.6	20.2			20.7					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.64	5.55			5.63					± 0.20
HB HGB	g/dl	18.0	17.9			18.2					± 0.6
	g/l	180	179			182					± 6
	mmol/l	11.18	11.12			11.30					± 0.37
HT HCT	%	50.8	50.5			51.2					± 3.0
	l/l	0.508	0.505			0.512					± 0.030
VGM MCV	µm ³ ; fl	90	91			91					± 4
TGMH MCH	pg	31.9	32.3			32.3					± 2.0
	fmol	1.98	2.00			2.01					± 0.12
CCMH MCHC	g/dl	35.5	35.4			35.5					± 3.0
	g/l	355	354			355					± 30
	mmol/l	22.02	22.01			22.06					± 1.86
IDR RDW	%	13.7	13.5			12.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	499	480			494					± 45
VPM MPV	µm ³ ; fl	7.8	7.8			7.4					± 2.0
LYMPHO	#	3.19	3.33			3.21					± 1.00
	%	15.5	16.5			15.5					± 5.0
MONO	#	1.03	1.11			0.93					± 0.80
	%	5.0	5.5			4.5					± 4.0
GRANULO	#	16.38	15.76			16.56					± 1.40
	%	79.5	78.0			80.0					± 7.0