



HORIBA ABX SAS

HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17



LOT MX 074 L
 Rev 1

CONTROL L



2014-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		2.0	2.0		2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l		2.36	2.40		2.35					± 0.15
HB HGB	g/dl		5.8	6.0		5.5					± 0.4
	g/l		58	60		55					± 4
	mmol/l		3.59	3.73		3.41					± 0.25
HT HCT	%		16.0	16.3		16.2					± 2.0
	l/l		0.160	0.163		0.162					± 0.020
VGM MCV	µm ³ ; fl		68	68		69					± 5
TGMH MCH	pg		24.5	25.0		23.3					± 2.0
	fmol		1.52	1.55		1.45					± 0.12
CCMH MCHC	g/dl		36.0	36.8		33.8					± 3.0
	g/l		360	368		338					± 30
	mmol/l		22.37	22.83		21.01					± 1.86
IDR RDW	%		13.8	13.5		13.6					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		80	81		81					± 20
VPM MPV	µm ³ ; fl		8.7	8.6		8.2					± 2.0
LYMPHO	#		1.26	1.2		1.24					± 0.40
	%		63.1	61.0		62.0					± 7.0
MONO	#		0.20	0.2		0.19					± 0.20
	%		10.0	9.0		9.5					± 6.0
GRANULO	#		0.54	0.6		0.57					± 0.40
	%		27.0	30.0		28.5					± 7.0

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT MX 074 N
 Rev 1

CONTROL N

(Exp.)

2014-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		7.7	7.9		7.6					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l		4.62	4.66		4.66					± 0.18
HB HGB	g/dl		13.4	13.5		13.1					± 0.5
	g/l		134	135		131					± 5
	mmol/l		8.32	8.38		8.15					± 0.31
HT HCT	%		37.0	36.8		37.3					± 2.5
	l/l		0.370	0.368		0.373					± 0.025
VGM MCV	µm ³ ; fl		80	79		80					± 5
TGMH MCH	pg		29.0	29.0		28.2					± 2.0
	fmol		1.80	1.80		1.75					± 0.12
CCMH MCHC	g/dl		36.3	36.7		35.2					± 3.0
	g/l		363	367		352					± 30
	mmol/l		22.51	22.77		21.87					± 1.86
IDR RDW	%		13.0	12.7		13.4					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		262	260		265					± 40
VPM MPV	µm ³ ; fl		8.0	8.2		7.6					± 2.0
LYMPHO	#		2.73	2.7		2.62					± 0.40
	%		35.5	34.0		34.7					± 5.0
MONO	#		0.55	0.6		0.53					± 0.30
	%		7.1	7.5		7.0					± 4.0
GRANULO	#		4.41	4.6		4.43					± 0.60
	%		57.3	58.5		58.5					± 6.0

Ref: TEMP-0630 Rev.35 9930086-B



HORIBA ABX SAS

HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17



LOT MX 074 H
 Rev 1

CONTROL H



2014-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		20.0	20.1		19.7					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l		5.50	5.53		5.58					± 0.20
HB HGB	g/dl		17.6	17.6		17.5					± 0.6
	g/l		176	176		175					± 6
	mmol/l		10.93	10.93		10.88					± 0.37
HT HCT	%		49.0	48.7		49.7					± 3.0
	l/l		0.490	0.487		0.497					± 0.030
VGM MCV	µm ³ ; fl		89	88		89					± 5
TGMH MCH	pg		32.0	31.8		31.4					± 2.0
	fmol		1.99	1.98		1.95					± 0.12
CCMH MCHC	g/dl		36.0	36.2		35.3					± 3.0
	g/l		360	362		353					± 30
	mmol/l		22.33	22.46		21.91					± 1.86
IDR RDW	%		12.5	12.2		13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		507	500		515					± 45
VPM MPV	µm ³ ; fl		8.0	8.2		7.5					± 2.0
LYMPHO	#		3.18	3.4		3.04					± 1.00
	%		15.9	17.0		15.5					± 5.0
MONO	#		1.19	1.0		0.98					± 0.80
	%		5.9	5.0		5.0					± 4.0
GRANULO	#		15.58	15.7		15.60					± 1.40
	%		77.9	78.0		79.2					± 7.0