

ABX Minotrol CRP



LOT

MC 413

CONTROL



(Exp.) **2020-01-05**
(YYYY - MM - DD)

Rev 2

		ABX Lyse														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0			± 0.4	7.1	7.6			± 1.0	19.1	20.1			± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.31	2.33			± 0.15	4.50	4.54			± 0.18	5.62	5.62			± 0.25
HB HGB	g/dl	6.1	6.1			± 0.4	13.3	13.4			± 0.5	17.5	17.6			± 0.7
	g/l	61	61			± 4	133	134			± 5	175	176			± 7
HT HCT	mmol/l	3.79	3.79			± 0.25	8.26	8.32			± 0.31	10.87	10.93			± 0.44
	%	16.4	16.5			± 2.0	36.5	36.8			± 2.5	49.5	49.5			± 3.0
VGM MCV	l/l	0.164	0.165			± 0.020	0.365	0.368			± 0.025	0.495	0.495			± 0.030
	µm ³ ; fl	71	71			± 4	81	81			± 4	88	88			± 4
TGMH MCH	pg	26.4	26.2			± 2.4	29.6	29.5			± 2.4	31.1	31.3			± 2.8
	fmol	1.64	1.63			± 0.15	1.84	1.83			± 0.15	1.93	1.94			± 0.18
CCMH MCHC	g/dl	37.2	36.9			± 3.0	36.5	36.4			± 3.0	35.4	35.6			± 3.0
	g/l	372	369			± 30	365	364			± 30	354	356			± 30
	mmol/l	23.10	22.90			± 1.86	22.66	22.63			± 1.86	21.97	22.10			± 1.86
IDR RDW	%	14.0	14.0			± 3.0	13.5	13.5			± 3.0	13.0	12.5			± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	77	77			± 20	250	260			± 40	490	500			± 60
VPM MPV	µm ³ ; fl	7.9	7.4			± 2.0	7.5	7.3			± 2.0	7.3	7.0			± 2.0
LYMPHO	#	1.4	1.3			± 0.2	2.5	2.5			± 0.5	3.5	3.4			± 1.3
	%	68.5	66.0			± 7.0	35.0	33.5			± 6.0	18.5	17.0			± 6.0
MONO	#	0.2	0.2			± 0.2	0.6	0.6			± 0.4	1.1	1.0			± 0.8
	%	8.0	8.0			± 6.0	9.0	8.0			± 5.0	5.5	5.0			± 4.0
GRANULO	#	0.5	0.5			± 0.2	4.0	4.4			± 0.5	14.5	15.7			± 1.5
	%	23.5	26.0			± 7.0	56.0	58.5			± 6.0	76.0	78.0			± 7.0
CRP	mg/dl	0.39	N/A			± 0.30	2.07	N/A			± 0.70	3.95	N/A			± 1.00
	mg/l	3.9	N/A			± 3.0	20.7	N/A			± 7.0	39.5	N/A			± 10.0

1300060514 FRONT/RECTO Ref: TEMP-1070 Rev.38