

HORIBA ABX SAS
HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée - BP 7290
34184 MONTPELLIER Cedex 4
FRANCE
Tél.: 33 (0) 4 67 14 15 16
Fax.: 33 (0) 4 67 14 15 17

LOT

MC 400

CONTROL

(Exp.) **2016-10-05**
(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | | UNITES UNITS | | ABX Lyse | | | | | | | | | | | | | | | | |
|--------------------------|------|---|--------------|--------------|--------------|-----------------|--|-------------------------|--------------|--------------|-----------------|--|-------------------------|---------|--------------|-----------------|--------------|-------------------------|---|-------------------------|
| | | | | CONTROL | | | | 1 | CONTROL | | | | 2 | | CONTROL | | | | 3 | TOLERANCES TOLERANCE |
| | | | | MICROS | | | | TOLERANCES TOLERANCE | MICROS | | | | TOLERANCES TOLERANCE | MICROS | | | | TOLERANCES TOLERANCE | | |
| | | | | CRP | CRP 200 | 60 8P 60 18P | | | CRP | CRP 200 | 60 8P 60 18P | | | CRP | CRP 200 | 60 8P 60 18P | | | | |
| GB | WBC | 10 ³ /mm ³ ;10 ⁹ /l | 2.0 | 2.0 | 2.0 | | | ± 0.4 | 7.3 | 7.3 | 7.5 | | | ± 1.0 | 19.5 | 19.5 | 19.8 | | | ± 2.2 |
| GR | RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 2.35 | 2.35 | 2.36 | | | ± 0.15 | 4.60 | 4.60 | 4.63 | | | ± 0.18 | 5.76 | 5.76 | 5.80 | | | ± 0.25 |
| HB | HGB | g/dl | 6.2 | 6.2 | 6.1 | | | ± 0.4 | 13.6 | 13.6 | 13.6 | | | ± 0.5 | 18.1 | 18.1 | 18.1 | | | ± 0.7 |
| | | g/l | 62 | 62 | 61 | | | ± 4 | 136 | 136 | 136 | | | ± 5 | 181 | 181 | 181 | | | ± 7 |
| | | mmol/l | 3.85 | 3.85 | 3.79 | | | ± 0.25 | 8.45 | 8.45 | 8.45 | | | ± 0.31 | 11.24 | 11.24 | 11.24 | | | ± 0.44 |
| HT | HCT | % | 16.5 | 16.5 | 16.5 | | | ± 2.0 | 37.3 | 37.3 | 37.5 | | | ± 2.5 | 50.1 | 50.1 | 50.5 | | | ± 3.0 |
| | | l/l | 0.165 | 0.165 | 0.165 | | | ± 0.020 | 0.373 | 0.373 | 0.375 | | | ± 0.025 | 0.501 | 0.501 | 0.505 | | | ± 0.030 |
| VGM | MCV | µm ³ ; fl | 70 | 70 | 70 | | | ± 4 | 81 | 81 | 81 | | | ± 4 | 87 | 87 | 87 | | | ± 4 |
| TGMH | MCH | pg | 26.4 | 26.4 | 25.8 | | | ± 2.4 | 29.6 | 29.6 | 29.4 | | | ± 2.4 | 31.4 | 31.4 | 31.2 | | | ± 2.8 |
| | | fmol | 1.64 | 1.64 | 1.61 | | | ± 0.15 | 1.84 | 1.84 | 1.82 | | | ± 0.15 | 1.95 | 1.95 | 1.94 | | | ± 0.18 |
| CCMH | MCHC | g/dl | 37.7 | 37.7 | 36.9 | | | ± 3.0 | 36.5 | 36.5 | 36.3 | | | ± 3.0 | 36.1 | 36.1 | 35.9 | | | ± 3.0 |
| | | g/l | 377 | 377 | 369 | | | ± 30 | 365 | 365 | 363 | | | ± 30 | 361 | 361 | 359 | | | ± 30 |
| | | mmol/l | 23.41 | 23.41 | 22.93 | | | ± 1.86 | 22.67 | 22.67 | 22.52 | | | ± 1.86 | 22.43 | 22.43 | 22.28 | | | ± 1.86 |
| IDR | RDW | % | 13.6 | 13.6 | 13.4 | | | ± 3.0 | 13.2 | 13.2 | 13.3 | | | ± 3.0 | 12.8 | 12.8 | 13.0 | | | ± 3.0 |
| PLAQ. | PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 75 | 75 | 72 | | | ± 20 | 254 | 254 | 250 | | | ± 40 | 510 | 510 | 505 | | | ± 60 |
| VPM | MPV | µm ³ ; fl | 8.7 | 8.7 | 8.3 | | | ± 2.0 | 8.0 | 8.0 | 7.6 | | | ± 2.0 | 7.9 | 7.9 | 7.5 | | | ± 2.0 |
| LYMPHO | | # | 1.3 | 1.3 | 1.3 | | | ± 0.3 | 2.5 | 2.5 | 2.6 | | | ± 0.5 | 3.0 | 3.0 | 3.2 | | | ± 1.2 |
| | | % | 64.0 | 64.0 | 62.8 | | | ± 8.0 | 34.0 | 34.0 | 35.0 | | | ± 6.0 | 15.5 | 15.5 | 16.0 | | | ± 6.0 |
| MONO | | # | 0.2 | 0.2 | 0.2 | | | ± 0.2 | 0.6 | 0.6 | 0.5 | | | ± 0.4 | 1.2 | 1.2 | 1.0 | | | ± 1.0 |
| | | % | 10.0 | 10.0 | 10.0 | | | ± 7.0 | 8.0 | 8.0 | 7.0 | | | ± 5.0 | 6.0 | 6.0 | 5.0 | | | ± 5.0 |
| GRANULO | | # | 0.5 | 0.5 | 0.5 | | | ± 0.3 | 4.2 | 4.2 | 4.4 | | | ± 0.5 | 15.3 | 15.3 | 15.6 | | | ± 1.4 |
| | | % | 26.0 | 26.0 | 27.2 | | | ± 8.0 | 58.0 | 58.0 | 58.0 | | | ± 6.0 | 78.5 | 78.5 | 79.0 | | | ± 7.0 |
| CRP | | mg/dl | 0.54 | 0.49 | N/A | | | ± 0.30 | 2.65 | 2.65 | N/A | | | ± 0.70 | 4.73 | 4.75 | N/A | | | ± 1.00 |
| | | mg/l | 5.4 | 4.9 | N/A | | | ± 3.0 | 26.5 | 26.5 | N/A | | | ± 7.0 | 47.3 | 47.5 | N/A | | | ± 10.0 |

Ref: TEMP-1070 Rev.36 FRONT/RECTO 9930089-A