

# ABX Minotrol CRP



**LOT**

**MC 410**

**CONTROL**



**(Exp.) 2019-04-05**  
(YYYY - MM - DD)

Rev 1

		<b>ABX Lyse</b>															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>			± 0.4	<b>7.6</b>	<b>7.7</b>			± 1.0	<b>19.9</b>	<b>20.4</b>			± 2.2	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.37</b>	<b>2.36</b>			± 0.15	<b>4.75</b>	<b>4.74</b>			± 0.18	<b>5.73</b>	<b>5.70</b>			± 0.25	
HB HGB	g/dl	<b>6.1</b>	<b>6.1</b>			± 0.4	<b>13.5</b>	<b>13.6</b>			± 0.5	<b>18.0</b>	<b>18.0</b>			± 0.7	
	g/l	<b>61</b>	<b>61</b>			± 4	<b>135</b>	<b>136</b>			± 5	<b>180</b>	<b>180</b>			± 7	
	mmol/l	<b>3.79</b>	<b>3.79</b>			± 0.25	<b>8.38</b>	<b>8.45</b>			± 0.31	<b>11.18</b>	<b>11.18</b>			± 0.44	
HT HCT	%	<b>16.6</b>	<b>16.5</b>			± 2.0	<b>38.0</b>	<b>37.9</b>			± 2.5	<b>51.0</b>	<b>50.7</b>			± 3.0	
	l/l	<b>0.166</b>	<b>0.165</b>			± 0.020	<b>0.380</b>	<b>0.379</b>			± 0.025	<b>0.510</b>	<b>0.507</b>			± 0.030	
VGM MCV	µm <sup>3</sup> ; fl	<b>70</b>	<b>70</b>			± 4	<b>80</b>	<b>80</b>			± 4	<b>89</b>	<b>89</b>			± 4	
TGMH MCH	pg	<b>25.7</b>	<b>25.8</b>			± 2.4	<b>28.4</b>	<b>28.7</b>			± 2.4	<b>31.4</b>	<b>31.6</b>			± 2.8	
	fmol	<b>1.60</b>	<b>1.61</b>			± 0.15	<b>1.76</b>	<b>1.78</b>			± 0.15	<b>1.95</b>	<b>1.96</b>			± 0.18	
CCMH MCHC	g/dl	<b>36.8</b>	<b>36.9</b>			± 3.0	<b>35.5</b>	<b>35.9</b>			± 3.0	<b>35.3</b>	<b>35.5</b>			± 3.0	
	g/l	<b>368</b>	<b>369</b>			± 30	<b>355</b>	<b>359</b>			± 30	<b>353</b>	<b>355</b>			± 30	
	mmol/l	<b>22.83</b>	<b>22.93</b>			± 1.86	<b>22.06</b>	<b>22.27</b>			± 1.86	<b>21.92</b>	<b>22.03</b>			± 1.86	
IDR RDW	%	<b>13.0</b>	<b>13.0</b>			± 3.0	<b>13.5</b>	<b>13.0</b>			± 3.0	<b>13.0</b>	<b>12.5</b>			± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>72</b>	<b>73</b>			± 20	<b>265</b>	<b>265</b>			± 40	<b>515</b>	<b>516</b>			± 60	
VPM MPV	µm <sup>3</sup> ; fl	<b>8.4</b>	<b>7.6</b>			± 2.0	<b>8.4</b>	<b>7.7</b>			± 2.0	<b>8.0</b>	<b>7.5</b>			± 2.0	
LYMPHO	#	<b>1.3</b>	<b>1.3</b>			± 0.2	<b>2.6</b>	<b>2.6</b>			± 0.5	<b>3.0</b>	<b>3.1</b>			± 1.2	
	%	<b>65.0</b>	<b>63.0</b>			± 8.0	<b>34.0</b>	<b>34.0</b>			± 6.0	<b>15.0</b>	<b>15.0</b>			± 6.0	
MONO	#	<b>0.2</b>	<b>0.2</b>			± 0.2	<b>0.6</b>	<b>0.5</b>			± 0.4	<b>1.2</b>	<b>1.0</b>			± 1.0	
	%	<b>9.5</b>	<b>10.0</b>			± 7.0	<b>8.0</b>	<b>7.0</b>			± 5.0	<b>6.0</b>	<b>5.0</b>			± 5.0	
GRANULO	#	<b>0.5</b>	<b>0.5</b>			± 0.2	<b>4.4</b>	<b>4.5</b>			± 0.5	<b>15.7</b>	<b>16.3</b>			± 1.4	
	%	<b>25.5</b>	<b>27.0</b>			± 7.0	<b>58.0</b>	<b>59.0</b>			± 6.0	<b>79.0</b>	<b>80.0</b>			± 7.0	
CRP	mg/dl	<b>0.52</b>	N/A			± 0.30	<b>1.95</b>	N/A			± 0.70	<b>3.57</b>	N/A			± 1.00	
	mg/l	<b>5.2</b>	N/A			± 3.0	<b>19.5</b>	N/A			± 7.0	<b>35.7</b>	N/A			± 10.0	

Ref: TEMP-1070 Rev.38 FRONT/RECTO 1300060514