

ABX Minotrol CRP



LOT

MC 401

CONTROL



(Exp.) 2017-01-05

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse													
				CONTROL			1	CONTROL			2	CONTROL			3	TOLERANCES TOLERANCE	
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P			
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.1	2.1	± 0.4	7.5	7.5	7.6	± 1.0	20.0	20.0	20.2	± 2.2			
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.43	2.43	2.40	± 0.15	4.60	4.60	4.60	± 0.18	5.67	5.67	5.70	± 0.25			
HB	HGB	g/dl	6.3	6.3	6.2	± 0.4	13.6	13.6	13.5	± 0.5	18.0	18.0	17.9	± 0.7			
		g/l	63	63	62	± 4	136	136	135	± 5	180	180	179	± 7			
HT	HCT	mmol/l	3.91	3.91	3.85	± 0.25	8.45	8.45	8.38	± 0.31	11.18	11.18	11.12	± 0.44			
		%	17.0	17.0	16.8	± 2.0	37.7	37.7	37.3	± 2.5	49.9	49.9	50.2	± 3.0			
VGM	MCV	l/l	0.170	0.170	0.168	± 0.020	0.377	0.377	0.373	± 0.025	0.499	0.499	0.502	± 0.030			
		µm ³ ; fl	70	70	70	± 4	82	82	81	± 4	88	88	88	± 4			
TGMH	MCH	pg	25.9	25.9	25.8	± 2.4	29.6	29.6	29.3	± 2.4	31.7	31.7	31.4	± 2.8			
		fmol	1.61	1.61	1.60	± 0.15	1.84	1.84	1.82	± 0.15	1.97	1.97	1.95	± 0.18			
CCMH	MCHC	g/dl	37.0	37.0	36.9	± 3.0	36.1	36.1	36.2	± 3.0	36.1	36.1	35.7	± 3.0			
		g/l	370	370	369	± 30	361	361	362	± 30	361	361	357	± 30			
		mmol/l	23.00	23.00	22.92	± 1.86	22.39	22.39	22.50	± 1.86	22.40	22.40	22.16	± 1.86			
IDR	RDW	%	14.0	14.0	13.5	± 3.0	14.0	14.0	14.0	± 3.0	13.0	13.0	13.0	± 3.0			
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	75	75	74	± 20	246	246	247	± 40	480	480	484	± 60			
VPM	MPV	µm ³ ; fl	8.0	8.0	7.0	± 2.0	7.8	7.8	7.0	± 2.0	7.5	7.5	7.0	± 2.0			
LYMPHO		#	1.3	1.3	1.3	± 0.3	2.4	2.4	2.4	± 0.5	3.1	3.1	3.1	± 1.3			
		%	63.0	63.0	62.0	± 8.0	32.5	32.5	32.0	± 6.0	15.5	15.5	15.5	± 6.0			
MONO		#	0.3	0.3	0.3	± 0.2	0.6	0.6	0.5	± 0.4	1.1	1.1	1.1	± 1.0			
		%	12.0	12.0	12.0	± 7.0	7.5	7.5	7.0	± 5.0	5.5	5.5	5.5	± 5.0			
GRANULO		#	0.5	0.5	0.5	± 0.3	4.5	4.5	4.6	± 0.5	15.8	15.8	16.0	± 1.5			
		%	25.0	25.0	26.0	± 8.0	60.0	60.0	61.0	± 6.0	79.0	79.0	79.0	± 7.0			
CRP		mg/dl	0.65	0.48	N/A	± 0.30	2.40	2.36	N/A	± 0.70	4.09	4.20	N/A	± 1.00			
		mg/l	6.5	4.8	N/A	± 3.0	24.0	23.6	N/A	± 7.0	40.9	42.0	N/A	± 10.0			

Ref: TEMP-1070 Rev.36 FRONT/RECTO 9930089-A