

ABX Minotrol CRP



LOT

MC 411

CONTROL



(Exp.) **2019-07-05**

(YYYY - MM - DD)

Rev 2

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS						MICROS						MICROS					
		CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.2			± 0.4	7.4	7.6			± 1.0	19.4	20.2			± 2.2			
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.36	2.38			± 0.15	4.55	4.58			± 0.18	5.59	5.62			± 0.25			
HB HGB	g/dl	6.0	6.0			± 0.4	13.3	13.4			± 0.5	17.7	17.8			± 0.7			
	g/l	60	60			± 4	133	134			± 5	177	178			± 7			
HT HCT	mmol/l	3.73	3.73			± 0.25	8.26	8.32			± 0.31	10.99	11.05			± 0.44			
	%	16.5	16.4			± 2.0	36.9	37.1			± 2.5	49.8	50.0			± 3.0			
VGM MCV	l/l	0.165	0.164			± 0.020	0.369	0.371			± 0.025	0.498	0.500			± 0.030			
	µm ³ ; fl	70	69			± 4	81	81			± 4	89	89			± 4			
TGMH MCH	pg	25.4	25.2			± 2.4	29.2	29.3			± 2.4	31.7	31.7			± 2.8			
	fmol	1.58	1.57			± 0.15	1.82	1.82			± 0.15	1.97	1.97			± 0.18			
CCMH MCHC	g/dl	36.3	36.5			± 3.0	36.1	36.1			± 3.0	35.6	35.6			± 3.0			
	g/l	363	365			± 30	361	361			± 30	356	356			± 30			
	mmol/l	22.55	22.69			± 1.86	22.41	22.43			± 1.86	22.09	22.10			± 1.86			
IDR RDW	%	14.5	14.0			± 3.0	13.5	13.2			± 3.0	12.8	12.7			± 3.0			
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	75	75			± 20	265	265			± 40	505	515			± 60			
VPM MPV	µm ³ ; fl	8.6	8.3			± 2.0	8.0	7.8			± 2.0	8.0	7.8			± 2.0			
LYMPHO	#	1.4	1.4			± 0.2	2.6	2.6			± 0.5	3.2	3.1			± 1.2			
	%	65.0	64.0			± 7.0	35.5	34.0			± 6.0	16.5	15.5			± 6.0			
MONO	#	0.2	0.2			± 0.2	0.5	0.5			± 0.4	0.9	0.8			± 0.8			
	%	8.5	8.0			± 6.0	7.0	6.5			± 5.0	4.5	4.0			± 4.0			
GRANULO	#	0.6	0.6			± 0.2	4.3	4.5			± 0.5	15.3	16.3			± 1.4			
	%	26.5	28.0			± 7.0	57.5	59.5			± 6.0	79.0	80.5			± 7.0			
CRP	mg/dl	0.58	N/A			± 0.30	2.58	N/A			± 0.70	4.36	N/A			± 1.00			
	mg/l	5.8	N/A			± 3.0	25.8	N/A			± 7.0	43.6	N/A			± 10.0			

Ref: TEMP-1070 Rev.38 FRONT/RECTO 1300060514