

# ABX Minotrol CRP



**LOT**

**MC 412**

**CONTROL**



(Exp.) **2019-10-05**  
(YYYY - MM - DD)

Rev 1

		<b>ABX Lyse</b>															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	1.9	2.0			± 0.4	7.4	7.8			± 1.0	19.7	20.8			± 2.2	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.31	2.32			± 0.15	4.59	4.62			± 0.18	5.55	5.61			± 0.25	
HB HGB	g/dl	6.0	6.0			± 0.4	13.3	13.4			± 0.5	17.6	17.8			± 0.7	
	g/l	60	60			± 4	133	134			± 5	176	178			± 7	
HT HCT	mmol/l	3.73	3.73			± 0.25	8.26	8.32			± 0.31	10.93	11.05			± 0.44	
	%	16.2	16.2			± 2.0	36.7	37.0			± 2.5	49.4	49.9			± 3.0	
VGM MCV	l/l	0.162	0.162			± 0.020	0.367	0.370			± 0.025	0.494	0.499			± 0.030	
	µm <sup>3</sup> ; fl	70	70			± 4	80	80			± 4	89	89			± 4	
TGMH MCH	pg	26.0	25.9			± 2.4	29.0	29.0			± 2.4	31.7	31.7			± 2.8	
	fmol	1.61	1.61			± 0.15	1.80	1.80			± 0.15	1.97	1.97			± 0.18	
CCMH MCHC	g/dl	37.1	36.9			± 3.0	36.2	36.3			± 3.0	35.6	35.7			± 3.0	
	g/l	371	369			± 30	362	363			± 30	356	357			± 30	
	mmol/l	23.04	22.94			± 1.86	22.49	22.51			± 1.86	22.13	22.14			± 1.86	
IDR RDW	%	14.4	13.5			± 3.0	13.8	13.3			± 3.0	12.8	12.7			± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	76	75			± 20	258	260			± 40	510	515			± 60	
VPM MPV	µm <sup>3</sup> ; fl	8.1	7.6			± 2.0	7.9	7.5			± 2.0	7.7	7.4			± 2.0	
LYMPHO	#	1.3	1.3			± 0.2	2.8	2.8			± 0.5	3.4	3.5			± 1.3	
	%	66.0	64.5			± 7.0	37.5	36.0			± 6.0	17.5	17.0			± 6.0	
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.2	1.0			± 0.8	
	%	9.5	9.2			± 6.0	7.5	6.7			± 5.0	6.0	5.0			± 4.0	
GRANULO	#	0.5	0.5			± 0.2	4.1	4.5			± 0.5	15.1	16.2			± 1.5	
	%	24.5	26.3			± 7.0	55.0	57.3			± 6.0	76.5	78.0			± 7.0	
CRP	mg/dl	0.48	N/A			± 0.30	2.10	N/A			± 0.70	4.12	N/A			± 1.00	
	mg/l	4.8	N/A			± 3.0	21.0	N/A			± 7.0	41.2	N/A			± 10.0	

Ref: TEMP-1070 Rev.38 FRONT/RECTO 1300060514