

ABX Minotrol CRP



LOT

MC 415

CONTROL



(Exp.) **2020-07-05**
(YYYY - MM - DD)

Rev 1

		ABX Lyse															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0			± 0.4	7.5	7.6			± 1.0	20.4	20.8			± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.43	2.43			± 0.15	4.61	4.68			± 0.18	5.62	5.68			± 0.25	
HB HGB	g/dl	6.2	6.2			± 0.4	13.3	13.5			± 0.5	18.0	18.3			± 0.7	
	g/l	62	62			± 4	133	135			± 5	180	183			± 7	
	mmol/l	3.85	3.85			± 0.25	8.26	8.38			± 0.31	11.18	11.36			± 0.44	
HT HCT	%	16.8	16.8			± 2.0	36.9	37.4			± 2.5	50.6	51.1			± 3.0	
	l/l	0.168	0.168			± 0.020	0.369	0.374			± 0.025	0.506	0.511			± 0.030	
VGM MCV	µm ³ ; fl	69	69			± 4	80	80			± 4	90	90			± 4	
TGMH MCH	pg	25.5	25.5			± 2.4	28.9	28.8			± 2.4	32.0	32.2			± 2.8	
	fmol	1.58	1.58			± 0.15	1.79	1.79			± 0.15	1.99	2.00			± 0.18	
CCMH MCHC	g/dl	37.0	37.0			± 3.0	36.1	36.1			± 3.0	35.6	35.8			± 3.0	
	g/l	370	370			± 30	361	361			± 30	356	358			± 30	
	mmol/l	22.96	22.96			± 1.86	22.40	22.39			± 1.86	22.10	22.23			± 1.86	
IDR RDW	%	13.7	13.6			± 3.0	14.6	13.8			± 3.0	13.5	13.0			± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	65	70			± 20	245	255			± 40	500	510			± 60	
VPM MPV	µm ³ ; fl	8.8	7.9			± 2.0	8.1	7.4			± 2.0	8.0	7.4			± 2.0	
LYMPHO	#	1.3	1.2			± 0.2	2.6	2.6			± 0.5	3.5	3.4			± 1.3	
	%	64.5	62.0			± 8.0	34.0	34.5			± 6.0	17.0	16.5			± 6.0	
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.2	0.9			± 0.8	
	%	10.5	11.0			± 7.0	7.5	6.5			± 5.0	6.0	4.5			± 4.0	
GRANULO	#	0.5	0.5			± 0.2	4.4	4.5			± 0.5	15.7	16.4			± 1.5	
	%	25.0	27.0			± 7.0	58.5	59.0			± 6.0	77.0	79.0			± 7.0	
CRP	mg/dl	0.59	N/A			± 0.30	2.11	N/A			± 0.70	4.68	N/A			± 1.00	
	mg/l	5.9	N/A			± 3.0	21.1	N/A			± 7.0	46.8	N/A			± 10.0	

Ref: TEMP-1070 Rev.38 FRONT/RECTO 1300060514