

ABX Minotrol CRP



LOT

MC 416

CONTROL



(Exp.) **2020-10-05**
(YYYY - MM - DD)

Rev 2

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					1	CONTROL					2	CONTROL					3
		MICROS					TOLERANCES TOLERANCE	MICROS					TOLERANCES TOLERANCE	MICROS					TOLERANCES TOLERANCE
		CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	1.9	2.0			± 0.4	7.6	7.8			± 1.0	20.0	20.5			± 2.2			
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.32	2.35			± 0.15	4.50	4.54			± 0.18	5.58	5.60			± 0.25			
HB HGB	g/dl	6.0	6.1			± 0.4	13.2	13.4			± 0.5	17.7	17.9			± 0.7			
	g/l	60	61			± 4	132	134			± 5	177	179			± 7			
	mmol/l	3.73	3.79			± 0.25	8.20	8.32			± 0.31	10.99	11.12			± 0.44			
HT HCT	%	16.2	16.5			± 2.0	36.9	36.8			± 2.5	49.7	49.8			± 3.0			
	l/l	0.162	0.165			± 0.020	0.369	0.368			± 0.025	0.497	0.498			± 0.030			
VGM MCV	µm ³ ; fl	70	70			± 4	82	81			± 4	89	89			± 4			
TGMH MCH	pg	25.9	26.0			± 2.4	29.3	29.5			± 2.4	31.7	32.0			± 2.8			
	fmol	1.61	1.61			± 0.15	1.82	1.83			± 0.15	1.97	1.98			± 0.18			
CCMH MCHC	g/dl	36.9	37.1			± 3.0	35.8	36.4			± 3.0	35.6	35.9			± 3.0			
	g/l	369	371			± 30	358	364			± 30	356	359			± 30			
	mmol/l	22.94	23.03			± 1.86	22.21	22.63			± 1.86	22.13	22.30			± 1.86			
IDR RDW	%	14.0	13.5			± 3.0	13.5	13.3			± 3.0	13.0	12.5			± 3.0			
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	72	68			± 20	265	270			± 40	505	509			± 60			
VPM MPV	µm ³ ; fl	8.0	7.4			± 2.0	7.5	7.1			± 2.0	7.3	7.0			± 2.0			
LYMPHO	#	1.2	1.2			± 0.2	2.6	2.6			± 0.5	3.4	3.3			± 1.3			
	%	63.5	62.0			± 8.0	34.0	33.0			± 6.0	17.0	16.0			± 6.0			
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.0	0.8			± 0.8			
	%	10.5	10.5			± 7.0	7.5	7.0			± 5.0	5.0	4.0			± 4.0			
GRANULO	#	0.5	0.6			± 0.2	4.4	4.7			± 0.5	15.6	16.4			± 1.5			
	%	26.0	27.5			± 7.0	58.5	60.0			± 6.0	78.0	80.0			± 7.0			
CRP	mg/dl	0.42	N/A			± 0.30	2.05	N/A			± 0.70	3.55	N/A			± 1.00			
	mg/l	4.2	N/A			± 3.0	20.5	N/A			± 7.0	35.5	N/A			± 10.0			

Ref: TEMP-1070 Rev.38 FRONT/RECTO 1300060514