

HORIBA ABX SAS
HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée - BP 7290
34184 MONTPELLIER Cedex 4
FRANCE
Tél.: 33 (0) 4 67 14 15 16
Fax.: 33 (0) 4 67 14 15 17

LOT

MC 407

CONTROL



(Exp.) 2018-07-05
(YYYY - MM - DD)

Rev 1

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				MICROS		MICROS				MICROS		MICROS					
		CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P				CRP	CRP 200	60 8P 60 18P			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0	2.0		± 0.4	7.5	7.5	7.7		± 1.0	20.1	20.1	20.5		± 2.2			
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.32	2.32	2.32		± 0.15	4.62	4.62	4.62		± 0.18	5.83	5.83	5.86		± 0.25			
HB HGB	g/dl	6.1	6.1	6.1		± 0.4	13.3	13.3	13.4		± 0.5	17.7	17.7	17.8		± 0.7			
	g/l	61	61	61		± 4	133	133	134		± 5	177	177	178		± 7			
HT HCT	mmol/l	3.79	3.79	3.79		± 0.25	8.26	8.26	8.32		± 0.31	10.99	10.99	11.05		± 0.44			
	%	16.2	16.2	16.5		± 2.0	37.4	37.4	37.4		± 2.5	50.1	50.1	50.4		± 3.0			
VGM MCV	l/l	0.162	0.162	0.165		± 0.020	0.374	0.374	0.374		± 0.025	0.501	0.501	0.504		± 0.030			
	µm ³ ; fl	70	70	71		± 4	81	81	81		± 4	86	86	86		± 4			
TGMH MCH	pg	26.3	26.3	26.3		± 2.4	28.8	28.8	29.0		± 2.4	30.4	30.4	30.4		± 2.8			
	fmol	1.63	1.63	1.63		± 0.15	1.79	1.79	1.80		± 0.15	1.89	1.89	1.89		± 0.18			
CCMH MCHC	g/dl	37.6	37.6	37.0		± 3.0	35.5	35.5	35.8		± 3.0	35.3	35.3	35.3		± 3.0			
	g/l	376	376	370		± 30	355	355	358		± 30	353	353	353		± 30			
	mmol/l	23.33	23.33	23.00		± 1.86	22.07	22.07	22.24		± 1.86	21.92	21.92	21.93		± 1.86			
IDR RDW	%	14.0	14.0	13.0		± 3.0	14.0	14.0	13.5		± 3.0	13.0	13.0	12.0		± 3.0			
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	79	79	81		± 20	256	256	260		± 40	509	509	510		± 60			
VPM MPV	µm ³ ; fl	8.1	8.1	7.6		± 2.0	8.5	8.5	8.1		± 2.0	8.0	8.0	7.7		± 2.0			
LYMPHO	#	1.3	1.3	1.3		± 0.2	2.5	2.5	2.5		± 0.5	3.0	3.0	3.1		± 1.2			
	%	66.0	66.0	64.5		± 8.0	33.5	33.5	32.5		± 6.0	15.0	15.0	15.0		± 6.0			
MONO	#	0.2	0.2	0.2		± 0.2	0.5	0.5	0.6		± 0.4	0.9	0.9	0.8		± 0.8			
	%	11.0	11.0	11.0		± 7.0	7.0	7.0	7.5		± 5.0	4.5	4.5	4.0		± 4.0			
GRANULO	#	0.5	0.5	0.5		± 0.2	4.5	4.5	4.6		± 0.5	16.2	16.2	16.6		± 1.5			
	%	23.0	23.0	24.5		± 7.0	59.5	59.5	60.0		± 6.0	80.5	80.5	81.0		± 7.0			
CRP	mg/dl	0.69	0.50	N/A		± 0.30	3.14	2.65	N/A		± 0.70	4.48	3.88	N/A		± 1.00			
	mg/l	6.9	5.0	N/A		± 3.0	31.4	26.5	N/A		± 7.0	44.8	38.8	N/A		± 10.0			

Ref: TEMP-1070 Rev.37 FRONT/RECTO 9930089-A