

ABX Minotrol CRP



LOT

MC 417

CONTROL



(Exp.) **2021-01-05**
(YYYY-MM-DD)

Rev 1

		ABX Lyse														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0			± 0.4	7.6	7.8			± 1.0	19.8	20.1			± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.42	2.40			± 0.15	4.56	4.57			± 0.18	5.50	5.48			± 0.25
HB HGB	g/dl	6.1	6.1			± 0.4	13.5	13.5			± 0.5	17.7	17.7			± 0.7
	g/l	61	61			± 4	135	135			± 5	177	177			± 7
	mmol/l	3.79	3.79			± 0.25	8.38	8.38			± 0.31	10.99	10.99			± 0.44
HT HCT	%	16.9	17.0			± 2.0	37.8	37.9			± 2.5	50.1	49.9			± 3.0
	l/l	0.169	0.170			± 0.020	0.378	0.379			± 0.025	0.501	0.499			± 0.030
VGM MCV	µm ³ ; fl	70	71			± 4	83	83			± 4	91	91			± 4
TGMH MCH	pg	25.2	25.4			± 2.4	29.6	29.5			± 2.4	32.2	32.3			± 2.8
	fmol	1.57	1.58			± 0.15	1.84	1.83			± 0.15	2.00	2.01			± 0.18
CCMH MCHC	g/dl	36.0	35.8			± 3.0	35.7	35.6			± 3.0	35.4	35.5			± 3.0
	g/l	360	358			± 30	357	356			± 30	354	355			± 30
	mmol/l	22.36	22.23			± 1.86	22.15	22.10			± 1.86	21.96	22.04			± 1.86
IDR RDW	%	14.0	14.0			± 3.0	13.5	13.5			± 3.0	12.5	12.5			± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	74	75			± 20	252	257			± 40	507	503			± 60
VPM MPV	µm ³ ; fl	8.9	7.9			± 2.0	7.9	7.3			± 2.0	7.3	6.8			± 2.0
LYMPHO	#	1.3	1.3			± 0.2	2.7	2.8			± 0.5	3.4	3.3			± 1.2
	%	66.0	65.0			± 8.0	35.5	36.0			± 6.0	17.0	16.5			± 6.0
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.0	0.9			± 0.9
	%	10.5	9.7			± 7.0	7.8	6.5			± 5.0	5.0	4.5			± 4.5
GRANULO	#	0.5	0.5			± 0.2	4.3	4.5			± 0.5	15.4	15.9			± 1.5
	%	23.5	25.3			± 7.0	56.7	57.5			± 6.0	78.0	79.0			± 7.0
CRP	mg/dl	0.46	N/A			± 0.30	1.92	N/A			± 0.70	3.94	N/A			± 1.00
	mg/l	4.6	N/A			± 3.0	19.2	N/A			± 7.0	39.4	N/A			± 10.0

Ref: TEMP-1070 Rev.38 FRONT/RECTO 1300060514