

ABX Minotrol CRP



LOT

MC 418

CONTROL



(Exp.) **2021-04-05**
(YYYY - MM - DD)

Rev 1

		ABX Lyse															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.1			± 0.4	7.6	7.8			± 1.0	19.9	20.9			± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.37	2.37			± 0.15	4.66	4.71			± 0.18	5.55	5.65			± 0.25	
HB HGB	g/dl	6.1	6.1			± 0.4	13.7	13.7			± 0.5	17.6	18.0			± 0.7	
	g/l	61	61			± 4	137	137			± 5	176	180			± 7	
	mmol/l	3.79	3.79			± 0.25	8.51	8.51			± 0.31	10.93	11.18			± 0.44	
HT HCT	%	16.8	16.6			± 2.0	38.2	38.2			± 2.5	50.0	50.3			± 3.0	
	l/l	0.168	0.166			± 0.020	0.382	0.382			± 0.025	0.500	0.503			± 0.030	
VGM MCV	µm ³ ; fl	71	70			± 4	82	81			± 4	90	89			± 4	
TGMH MCH	pg	25.7	25.7			± 2.4	29.4	29.1			± 2.4	31.7	31.9			± 2.8	
	fmol	1.60	1.60			± 0.15	1.83	1.81			± 0.15	1.97	1.98			± 0.18	
CCMH MCHC	g/dl	36.3	36.8			± 3.0	35.9	35.9			± 3.0	35.2	35.8			± 3.0	
	g/l	363	368			± 30	359	359			± 30	352	358			± 30	
	mmol/l	22.51	22.83			± 1.86	22.26	22.30			± 1.86	21.88	22.23			± 1.86	
IDR RDW	%	14.0	14.0			± 3.0	13.5	13.0			± 3.0	13.0	13.0			± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	82	83			± 20	274	272			± 40	530	533			± 60	
VPM MPV	µm ³ ; fl	8.2	7.5			± 2.0	8.1	7.2			± 2.0	7.8	7.0			± 2.0	
LYMPHO	#	1.4	1.4			± 0.3	2.8	2.7			± 0.5	3.4	3.2			± 1.3	
	%	66.0	64.5			± 8.0	37.0	35.0			± 6.0	17.0	15.5			± 6.0	
MONO	#	0.2	0.2			± 0.2	0.6	0.6			± 0.4	1.1	1.0			± 1.0	
	%	9.5	9.0			± 7.0	8.0	7.5			± 5.0	5.5	5.0			± 5.0	
GRANULO	#	0.5	0.6			± 0.2	4.2	4.5			± 0.6	15.4	16.6			± 1.5	
	%	24.5	26.5			± 7.0	55.0	57.5			± 6.0	77.5	79.5			± 7.0	
CRP	mg/dl	0.46	N/A			± 0.30	2.09	N/A			± 0.70	3.89	N/A			± 1.00	
	mg/l	4.6	N/A			± 3.0	20.9	N/A			± 7.0	38.9	N/A			± 10.0	

Ref: TEMP-1070 Rev.38 FRONT/RECTO 1300060514