

LOT MC 070
 Rev 4

CONTROL

(Exp.) 2010-10-05
 (YYYY - MM - DD)

| PARAMETRES PARAMETERS | | UNITES UNITS | ABX Lyse | | | | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|--|-----------------|--------------|--------------|-----------------|-------------------------|--------------|--------------|-----------------|-------------------------|--------------|--------------|-----------------|-------------------------|-------------------------|
| | | | CONTROL | | | 1 | CONTROL | | | 2 | CONTROL | | | 3 | |
| | | | MICROS | | | TOLERANCES TOLERANCE | MICROS | | | TOLERANCES TOLERANCE | MICROS | | | TOLERANCES TOLERANCE | |
| | | | CRP | CRP 200 | 60 8P 60 18P | | CRP | CRP 200 | 60 8P 60 18P | | CRP | CRP 200 | 60 8P 60 18P | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.0 | 2.0 | 2.1 | ± 0.4 | 7.7 | 7.7 | 7.7 | ± 1.0 | 20.4 | 20.4 | 20.4 | ± 2.2 | | |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.39 | 2.39 | 2.36 | ± 0.15 | 4.62 | 4.62 | 4.59 | ± 0.18 | 5.60 | 5.60 | 5.57 | ± 0.25 | | |
| HB HGB | g/dl | 6.2 | 6.2 | 6.0 | ± 0.4 | 13.6 | 13.6 | 13.5 | ± 0.5 | 17.9 | 17.9 | 17.7 | ± 0.7 | | |
| | g/l | 62 | 62 | 60 | ± 4 | 136 | 136 | 135 | ± 5 | 179 | 179 | 177 | ± 7 | | |
| HT HCT | mmol/l | 3.85 | 3.85 | 3.73 | ± 0.25 | 8.45 | 8.45 | 8.38 | ± 0.31 | 11.12 | 11.12 | 10.99 | ± 0.44 | | |
| | % | 17.2 | 17.2 | 17.0 | ± 2.0 | 37.9 | 37.9 | 38.1 | ± 2.5 | 50.4 | 50.4 | 50.1 | ± 3.0 | | |
| VGM MCV | l/l | 0.172 | 0.172 | 0.170 | ± 0.020 | 0.379 | 0.379 | 0.381 | ± 0.025 | 0.504 | 0.504 | 0.501 | ± 0.030 | | |
| | µm ³ ; fl | 72 | 72 | 72 | ± 4 | 82 | 82 | 83 | ± 4 | 90 | 90 | 90 | ± 4 | | |
| TGMH MCH | pg | 25.9 | 25.9 | 25.4 | ± 2.4 | 29.4 | 29.4 | 29.4 | ± 2.4 | 32.0 | 32.0 | 31.8 | ± 2.8 | | |
| | fmol | 1.61 | 1.61 | 1.58 | ± 0.15 | 1.83 | 1.83 | 1.83 | ± 0.15 | 1.98 | 1.98 | 1.97 | ± 0.18 | | |
| CCMH MCHC | g/dl | 36.0 | 36.0 | 35.3 | ± 3.0 | 35.9 | 35.9 | 35.4 | ± 3.0 | 35.5 | 35.5 | 35.3 | ± 3.0 | | |
| | g/l | 360 | 360 | 353 | ± 30 | 359 | 359 | 354 | ± 30 | 355 | 355 | 353 | ± 30 | | |
| IDR RDW | mmol/l | 22.37 | 22.37 | 21.93 | ± 1.86 | 22.29 | 22.29 | 22.01 | ± 1.86 | 22.06 | 22.06 | 21.93 | ± 1.86 | | |
| | % | 13.8 | 13.8 | 13.5 | ± 3.0 | 12.8 | 12.8 | 13.4 | ± 3.0 | 12.3 | 12.3 | 12.5 | ± 3.0 | | |
| PLAQ. PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 80 | 80 | 78 | ± 20 | 260 | 260 | 258 | ± 40 | 510 | 510 | 500 | ± 60 | | |
| VPM MPV | µm ³ ; fl | 8.7 | 8.7 | 8.8 | ± 2.0 | 8.0 | 8.0 | 8.0 | ± 2.0 | 8.3 | 8.3 | 8.3 | ± 2.0 | | |
| LYMPHO | # | 1.2 | 1.2 | 1.3 | ± 0.2 | 2.5 | 2.5 | 2.6 | ± 0.5 | 3.2 | 3.2 | 3.2 | ± 1.1 | | |
| | % | 61.5 | 61.5 | 61.0 | ± 7.0 | 33.0 | 33.0 | 34.0 | ± 6.0 | 15.8 | 15.8 | 15.8 | ± 5.0 | | |
| MONO | # | 0.2 | 0.2 | 0.2 | ± 0.2 | 0.5 | 0.5 | 0.5 | ± 0.5 | 1.0 | 1.0 | 1.0 | ± 1.0 | | |
| | % | 12.3 | 12.3 | 11.8 | ± 7.0 | 7.0 | 7.0 | 6.0 | ± 5.0 | 4.8 | 4.8 | 4.8 | ± 4.8 | | |
| GRANULO | # | 0.5 | 0.5 | 0.6 | ± 0.2 | 4.6 | 4.6 | 4.6 | ± 0.5 | 16.2 | 16.2 | 16.2 | ± 1.5 | | |
| | % | 26.2 | 26.2 | 27.2 | ± 7.0 | 60.0 | 60.0 | 60.0 | ± 6.0 | 79.4 | 79.4 | 79.4 | ± 7.0 | | |
| CRP | mg/dl | 0.70 | 0.50 | N/A | ± 0.30 | 2.55 | 2.45 | N/A | ± 0.70 | 3.95 | 3.95 | N/A | ± 1.00 | | |
| | mg/l | 7.0 | 5.0 | N/A | ± 3.0 | 25.5 | 24.5 | N/A | ± 7.0 | 39.5 | 39.5 | N/A | ± 10.0 | | |

Ref: FORM-1070 Rev.12 FRONT / RECTO 9930045-B