

ABX Minotrol CRP



LOT MC 012
Rev 1

CONTROL

(Exp.) 2012-04-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lyse													TOLERANCES TOLERANCE
			CONTROL			1	CONTROL			2	CONTROL			3		
			MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
			CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0	2.0	± 0.4	7.5	7.5	7.5	± 1.0	19.9	19.9	19.8	± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.33	2.33	2.30	± 0.15	4.79	4.79	4.76	± 0.18	5.72	5.72	5.74	± 0.25			
HB HGB	g/dl	6.0	6.0	6.0	± 0.4	13.6	13.6	13.5	± 0.5	17.9	17.9	18.0	± 0.7			
	g/l	60	60	60	± 4	136	136	135	± 5	179	179	180	± 7			
HT HCT	mmol/l	3.73	3.73	3.73	± 0.25	8.45	8.45	8.38	± 0.31	11.12	11.12	11.18	± 0.44			
	%	16.5	16.5	16.6	± 2.0	38.3	38.3	38.6	± 2.5	50.9	50.9	51.7	± 3.0			
VGM MCV	l/l	0.165	0.165	0.166	± 0.020	0.383	0.383	0.386	± 0.025	0.509	0.509	0.517	± 0.030			
	µm ³ ; fl	71	71	72	± 4	80	80	81	± 4	89	89	90	± 4			
TGMH MCH	pg	25.8	25.8	26.1	± 2.4	28.4	28.4	28.4	± 2.4	31.3	31.3	31.4	± 2.8			
	fmol	1.60	1.60	1.62	± 0.15	1.76	1.76	1.76	± 0.15	1.94	1.94	1.95	± 0.18			
CCMH MCHC	g/dl	36.3	36.3	36.2	± 3.0	35.5	35.5	35.0	± 3.0	35.2	35.2	34.8	± 3.0			
	g/l	363	363	362	± 30	355	355	350	± 30	352	352	348	± 30			
	mmol/l	22.52	22.52	22.50	± 1.86	22.04	22.04	21.74	± 1.86	21.84	21.84	21.64	± 1.86			
IDR RDW	%	13.5	13.5	14.0	± 3.0	13.0	13.0	14.0	± 3.0	13.0	13.0	13.8	± 3.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	76	76	80	± 20	250	250	254	± 40	500	500	500	± 60			
VPM MPV	µm ³ ; fl	8.2	8.2	8.0	± 2.0	7.6	7.6	7.8	± 2.0	7.4	7.4	7.7	± 2.0			
LYMPHO	#	1.3	1.3	1.3	± 0.2	2.6	2.6	2.6	± 0.5	2.9	2.9	3.1	± 1.1			
	%	63.5	63.5	64.0	± 8.0	34.5	34.5	35.0	± 6.0	14.5	14.5	15.5	± 5.0			
MONO	#	0.3	0.3	0.2	± 0.2	0.6	0.6	0.5	± 0.4	1.1	1.1	1.0	± 1.0			
	%	12.5	12.5	11.5	± 7.0	7.5	7.5	7.0	± 5.0	5.5	5.5	5.0	± 5.0			
GRANULO	#	0.5	0.5	0.5	± 0.2	4.4	4.4	4.4	± 0.5	15.9	15.9	15.7	± 1.4			
	%	24.0	24.0	24.5	± 7.0	58.0	58.0	58.0	± 6.0	80.0	80.0	79.5	± 7.0			
CRP	mg/dl	0.50	0.43	N/A	± 0.30	2.50	2.35	N/A	± 0.70	4.60	4.30	N/A	± 1.00			
	mg/l	5.0	4.3	N/A	± 3.0	25.0	23.5	N/A	± 7.0	46.0	43.0	N/A	± 10.0			

Ref: FORM-1070 Rev.14 FRONT / RECTO 9930045-C