

# ABX Minotrol CRP



**LOT**

**MC 074**

**CONTROL**



**(Exp.) 2014-10-05**

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse													TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3			
				MICROS				MICROS				MICROS						
				CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P		CRP	CRP 200	60 8P 60 18P				
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>	<b>2.1</b>		± 0.4	<b>7.4</b>	<b>7.4</b>	<b>7.6</b>		± 1.0	<b>19.7</b>	<b>19.7</b>	<b>20.1</b>		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.34</b>	<b>2.34</b>	<b>2.36</b>		± 0.15	<b>4.59</b>	<b>4.59</b>	<b>4.64</b>		± 0.18	<b>5.47</b>	<b>5.47</b>	<b>5.53</b>		± 0.25	
HB	HGB	g/dl	<b>6.0</b>	<b>6.0</b>	<b>6.0</b>		± 0.4	<b>13.4</b>	<b>13.4</b>	<b>13.4</b>		± 0.5	<b>17.7</b>	<b>17.7</b>	<b>17.7</b>		± 0.7	
		g/l	<b>60</b>	<b>60</b>	<b>60</b>		± 4	<b>134</b>	<b>134</b>	<b>134</b>		± 5	<b>177</b>	<b>177</b>	<b>177</b>		± 7	
		mmol/l	<b>3.73</b>	<b>3.73</b>	<b>3.73</b>		± 0.25	<b>8.32</b>	<b>8.32</b>	<b>8.32</b>		± 0.31	<b>10.99</b>	<b>10.99</b>	<b>10.99</b>		± 0.44	
HT	HCT	%	<b>16.4</b>	<b>16.4</b>	<b>16.3</b>		± 2.0	<b>37.2</b>	<b>37.2</b>	<b>37.6</b>		± 2.5	<b>49.2</b>	<b>49.2</b>	<b>49.8</b>		± 3.0	
		l/l	<b>0.164</b>	<b>0.164</b>	<b>0.163</b>		± 0.020	<b>0.372</b>	<b>0.372</b>	<b>0.376</b>		± 0.025	<b>0.492</b>	<b>0.492</b>	<b>0.498</b>		± 0.030	
VGM	MCV	µm <sup>3</sup> ; fl	<b>70</b>	<b>70</b>	<b>69</b>		± 4	<b>81</b>	<b>81</b>	<b>81</b>		± 4	<b>90</b>	<b>90</b>	<b>90</b>		± 4	
TGMH	MCH	pg	<b>25.6</b>	<b>25.6</b>	<b>25.4</b>		± 2.4	<b>29.2</b>	<b>29.2</b>	<b>28.9</b>		± 2.4	<b>32.4</b>	<b>32.4</b>	<b>32.0</b>		± 2.8	
		fmol	<b>1.59</b>	<b>1.59</b>	<b>1.58</b>		± 0.15	<b>1.81</b>	<b>1.81</b>	<b>1.79</b>		± 0.15	<b>2.01</b>	<b>2.01</b>	<b>1.99</b>		± 0.18	
CCMH	MCHC	g/dl	<b>36.6</b>	<b>36.6</b>	<b>36.8</b>		± 3.0	<b>36.0</b>	<b>36.0</b>	<b>35.7</b>		± 3.0	<b>36.0</b>	<b>36.0</b>	<b>35.6</b>		± 3.0	
		g/l	<b>366</b>	<b>366</b>	<b>368</b>		± 30	<b>360</b>	<b>360</b>	<b>357</b>		± 30	<b>360</b>	<b>360</b>	<b>356</b>		± 30	
		mmol/l	<b>22.75</b>	<b>22.75</b>	<b>22.88</b>		± 1.86	<b>22.38</b>	<b>22.38</b>	<b>22.14</b>		± 1.86	<b>22.33</b>	<b>22.33</b>	<b>22.08</b>		± 1.86	
IDR	RDW	%	<b>14.0</b>	<b>14.0</b>	<b>14.0</b>		± 3.0	<b>13.7</b>	<b>13.7</b>	<b>13.7</b>		± 3.0	<b>13.0</b>	<b>13.0</b>	<b>13.2</b>		± 3.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>80</b>	<b>80</b>	<b>80</b>		± 20	<b>257</b>	<b>257</b>	<b>258</b>		± 40	<b>495</b>	<b>495</b>	<b>500</b>		± 60	
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.6</b>	<b>8.6</b>	<b>8.0</b>		± 2.0	<b>7.7</b>	<b>7.7</b>	<b>7.4</b>		± 2.0	<b>7.6</b>	<b>7.6</b>	<b>7.3</b>		± 2.0	
LYMPHO		#	<b>1.3</b>	<b>1.3</b>	<b>1.3</b>		± 0.2	<b>2.5</b>	<b>2.5</b>	<b>2.5</b>		± 0.5	<b>3.1</b>	<b>3.1</b>	<b>3.1</b>		± 1.1	
		%	<b>63.5</b>	<b>63.5</b>	<b>63.0</b>		± 8.0	<b>33.5</b>	<b>33.5</b>	<b>33.5</b>		± 6.0	<b>15.5</b>	<b>15.5</b>	<b>15.5</b>		± 5.0	
MONO		#	<b>0.2</b>	<b>0.2</b>	<b>0.2</b>		± 0.2	<b>0.6</b>	<b>0.6</b>	<b>0.6</b>		± 0.4	<b>1.0</b>	<b>1.0</b>	<b>1.0</b>		± 1.0	
		%	<b>8.5</b>	<b>8.5</b>	<b>8.5</b>		± 7.0	<b>8.0</b>	<b>8.0</b>	<b>7.5</b>		± 5.0	<b>5.0</b>	<b>5.0</b>	<b>5.0</b>		± 5.0	
GRANULO		#	<b>0.6</b>	<b>0.6</b>	<b>0.6</b>		± 0.2	<b>4.3</b>	<b>4.3</b>	<b>4.5</b>		± 0.5	<b>15.7</b>	<b>15.7</b>	<b>16.0</b>		± 1.5	
		%	<b>28.0</b>	<b>28.0</b>	<b>28.5</b>		± 7.0	<b>58.5</b>	<b>58.5</b>	<b>59.0</b>		± 6.0	<b>79.5</b>	<b>79.5</b>	<b>79.5</b>		± 7.0	
CRP		mg/dl	<b>0.81</b>	<b>0.46</b>	N/A		± 0.30	<b>2.84</b>	<b>2.32</b>	N/A		± 0.70	<b>4.46</b>	<b>3.85</b>	N/A		± 1.00	
		mg/l	<b>8.1</b>	<b>4.6</b>	N/A		± 3.0	<b>28.4</b>	<b>23.2</b>	N/A		± 7.0	<b>44.6</b>	<b>38.5</b>	N/A		± 10.0	

9930089-A  
FRONT/RECTO  
Ref: TEMP-1070 Rev.35