

HORIBA ABX SAS
HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée - BP 7290
34184 MONTPELLIER Cedex 4
FRANCE
Tél.: 33 (0) 4 67 14 15 16
Fax.: 33 (0) 4 67 14 15 17

LOT

MC 015

CONTROL



(Exp.) **2015-04-05**

(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | | UNITES UNITS | | ABX Lyse | | | | | | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|------|---|--------------|--------------|--------------|-----------------|---------|--------------|--------------|-----------------|---|---------|--------------|-----------------|--------------|--|-------------------------|--|
| | | | | CONTROL | | | 1 | CONTROL | | | 2 | CONTROL | | | 3 | | | |
| | | | | MICROS | | | | MICROS | | | | MICROS | | | | | | |
| | | | | CRP | CRP 200 | 60 8P 60 18P | | CRP | CRP 200 | 60 8P 60 18P | | CRP | CRP 200 | 60 8P 60 18P | | | | |
| GB | WBC | 10 ³ /mm ³ ;10 ⁹ /l | 2.1 | 2.1 | 2.1 | | ± 0.4 | 7.6 | 7.6 | 7.6 | | ± 1.0 | 20.3 | 20.3 | 20.5 | | ± 2.2 | |
| GR | RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 2.34 | 2.34 | 2.35 | | ± 0.15 | 4.60 | 4.60 | 4.61 | | ± 0.18 | 5.60 | 5.60 | 5.66 | | ± 0.25 | |
| HB | HGB | g/dl | 6.1 | 6.1 | 6.1 | | ± 0.4 | 13.7 | 13.7 | 13.6 | | ± 0.5 | 17.9 | 17.9 | 17.8 | | ± 0.7 | |
| | | g/l | 61 | 61 | 61 | | ± 4 | 137 | 137 | 136 | | ± 5 | 179 | 179 | 178 | | ± 7 | |
| | | mmol/l | 3.79 | 3.79 | 3.79 | | ± 0.25 | 8.51 | 8.51 | 8.45 | | ± 0.31 | 11.12 | 11.12 | 11.05 | | ± 0.44 | |
| HT | HCT | % | 16.6 | 16.6 | 16.5 | | ± 2.0 | 37.7 | 37.7 | 37.8 | | ± 2.5 | 49.8 | 49.8 | 50.4 | | ± 3.0 | |
| | | l/l | 0.166 | 0.166 | 0.165 | | ± 0.020 | 0.377 | 0.377 | 0.378 | | ± 0.025 | 0.498 | 0.498 | 0.504 | | ± 0.030 | |
| VGM | MCV | µm ³ , fl | 71 | 71 | 70 | | ± 4 | 82 | 82 | 82 | | ± 4 | 89 | 89 | 89 | | ± 4 | |
| TGMH | MCH | pg | 26.1 | 26.1 | 26.0 | | ± 2.4 | 29.8 | 29.8 | 29.5 | | ± 2.4 | 32.0 | 32.0 | 31.4 | | ± 2.8 | |
| | | fmol | 1.62 | 1.62 | 1.61 | | ± 0.15 | 1.85 | 1.85 | 1.83 | | ± 0.15 | 1.98 | 1.98 | 1.95 | | ± 0.18 | |
| CCMH | MCHC | g/dl | 36.7 | 36.7 | 37.1 | | ± 3.0 | 36.3 | 36.3 | 36.0 | | ± 3.0 | 35.9 | 35.9 | 35.3 | | ± 3.0 | |
| | | g/l | 367 | 367 | 371 | | ± 30 | 363 | 363 | 360 | | ± 30 | 359 | 359 | 353 | | ± 30 | |
| | | mmol/l | 22.80 | 22.80 | 23.03 | | ± 1.86 | 22.55 | 22.55 | 22.34 | | ± 1.86 | 22.30 | 22.30 | 21.94 | | ± 1.86 | |
| IDR | RDW | % | 14.0 | 14.0 | 13.5 | | ± 3.0 | 13.5 | 13.5 | 13.0 | | ± 3.0 | 13.0 | 13.0 | 12.5 | | ± 3.0 | |
| PLAQ. | PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 74 | 74 | 77 | | ± 20 | 250 | 250 | 255 | | ± 40 | 500 | 500 | 500 | | ± 60 | |
| VPM | MPV | µm ³ , fl | 8.5 | 8.5 | 7.7 | | ± 2.0 | 8.0 | 8.0 | 7.5 | | ± 2.0 | 7.7 | 7.7 | 7.1 | | ± 2.0 | |
| LYMPHO | | # | 1.4 | 1.4 | 1.4 | | ± 0.2 | 2.6 | 2.6 | 2.5 | | ± 0.5 | 2.9 | 2.9 | 2.9 | | ± 1.3 | |
| | | % | 67.0 | 67.0 | 65.5 | | ± 8.0 | 34.0 | 34.0 | 33.5 | | ± 6.0 | 14.5 | 14.5 | 14.0 | | ± 6.0 | |
| MONO | | # | 0.2 | 0.2 | 0.2 | | ± 0.2 | 0.6 | 0.6 | 0.5 | | ± 0.4 | 1.2 | 1.2 | 1.1 | | ± 1.0 | |
| | | % | 9.0 | 9.0 | 9.0 | | ± 7.0 | 7.5 | 7.5 | 7.0 | | ± 5.0 | 6.0 | 6.0 | 5.5 | | ± 5.0 | |
| GRANULO | | # | 0.5 | 0.5 | 0.5 | | ± 0.2 | 4.4 | 4.4 | 4.5 | | ± 0.5 | 16.1 | 16.1 | 16.5 | | ± 1.5 | |
| | | % | 24.0 | 24.0 | 25.5 | | ± 7.0 | 58.5 | 58.5 | 59.5 | | ± 6.0 | 79.5 | 79.5 | 80.5 | | ± 7.0 | |
| CRP | | mg/dl | 0.69 | 0.41 | N/A | | ± 0.30 | 2.56 | 2.12 | N/A | | ± 0.70 | 4.77 | 4.21 | N/A | | ± 1.00 | |
| | | mg/l | 6.9 | 4.1 | N/A | | ± 3.0 | 25.6 | 21.2 | N/A | | ± 7.0 | 47.7 | 42.1 | N/A | | ± 10.0 | |

9930089-A
FRONT/RECTO
Ref: TEMP-1070 Rev.36