

# ABX Minotrol CRP

**LOT**

**MC 105**

Rev 2

**CONTROL**



(Exp.) **2016-01-05**

(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse																
				CONTROL				1	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
				MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE		
				CRP	CRP 200	60 8P 60 18P			CRP	CRP 200	60 8P 60 18P			CRP	CRP 200	60 8P 60 18P				
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.1</b>	<b>2.1</b>			± 0.4	<b>7.9</b>	<b>7.9</b>	<b>7.9</b>			± 1.0	<b>19.9</b>	<b>19.9</b>	<b>19.9</b>			± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.40</b>	<b>2.40</b>	<b>2.43</b>			± 0.15	<b>4.53</b>	<b>4.53</b>	<b>4.55</b>			± 0.18	<b>5.62</b>	<b>5.62</b>	<b>5.65</b>			± 0.25
HB	HGB	g/dl	<b>6.1</b>	<b>6.1</b>	<b>6.1</b>			± 0.4	<b>13.3</b>	<b>13.3</b>	<b>13.4</b>			± 0.5	<b>17.5</b>	<b>17.5</b>	<b>17.8</b>			± 0.7
		g/l	<b>61</b>	<b>61</b>	<b>61</b>			± 4	<b>133</b>	<b>133</b>	<b>134</b>			± 5	<b>175</b>	<b>175</b>	<b>178</b>			± 7
HT	HCT	mmol/l	<b>3.79</b>	<b>3.79</b>	<b>3.79</b>			± 0.25	<b>8.26</b>	<b>8.26</b>	<b>8.32</b>			± 0.31	<b>10.87</b>	<b>10.87</b>	<b>11.05</b>			± 0.44
		%	<b>16.8</b>	<b>16.8</b>	<b>16.8</b>			± 2.0	<b>37.1</b>	<b>37.1</b>	<b>37.3</b>			± 2.5	<b>50.0</b>	<b>50.0</b>	<b>50.3</b>			± 3.0
		l/l	<b>0.168</b>	<b>0.168</b>	<b>0.168</b>			± 0.020	<b>0.371</b>	<b>0.371</b>	<b>0.373</b>			± 0.025	<b>0.500</b>	<b>0.500</b>	<b>0.503</b>			± 0.030
VGM	MCV	µm <sup>3</sup> ; fl	<b>70</b>	<b>70</b>	<b>69</b>			± 4	<b>82</b>	<b>82</b>	<b>82</b>			± 4	<b>89</b>	<b>89</b>	<b>89</b>			± 4
TGMH	MCH	pg	<b>25.4</b>	<b>25.4</b>	<b>25.1</b>			± 2.4	<b>29.4</b>	<b>29.4</b>	<b>29.5</b>			± 2.4	<b>31.1</b>	<b>31.1</b>	<b>31.5</b>			± 2.8
		fmol	<b>1.58</b>	<b>1.58</b>	<b>1.56</b>			± 0.15	<b>1.82</b>	<b>1.82</b>	<b>1.83</b>			± 0.15	<b>1.93</b>	<b>1.93</b>	<b>1.96</b>			± 0.18
CCMH	MCHC	g/dl	<b>36.3</b>	<b>36.3</b>	<b>36.4</b>			± 3.0	<b>35.8</b>	<b>35.8</b>	<b>35.9</b>			± 3.0	<b>35.0</b>	<b>35.0</b>	<b>35.4</b>			± 3.0
		g/l	<b>363</b>	<b>363</b>	<b>364</b>			± 30	<b>358</b>	<b>358</b>	<b>359</b>			± 30	<b>350</b>	<b>350</b>	<b>354</b>			± 30
		mmol/l	<b>22.55</b>	<b>22.55</b>	<b>22.59</b>			± 1.86	<b>22.23</b>	<b>22.23</b>	<b>22.30</b>			± 1.86	<b>21.73</b>	<b>21.73</b>	<b>21.98</b>			± 1.86
IDR	RDW	%	<b>13.8</b>	<b>13.8</b>	<b>13.8</b>			± 3.0	<b>13.5</b>	<b>13.5</b>	<b>13.3</b>			± 3.0	<b>12.7</b>	<b>12.7</b>	<b>13.0</b>			± 3.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>66</b>	<b>66</b>	<b>71</b>			± 20	<b>248</b>	<b>248</b>	<b>250</b>			± 40	<b>480</b>	<b>480</b>	<b>480</b>			± 60
VPM	MPV	µm <sup>3</sup> ; fl	<b>9.2</b>	<b>9.2</b>	<b>8.5</b>			± 2.0	<b>8.0</b>	<b>8.0</b>	<b>7.7</b>			± 2.0	<b>7.8</b>	<b>7.8</b>	<b>7.3</b>			± 2.0
LYMPHO		#	<b>1.3</b>	<b>1.3</b>	<b>1.3</b>			± 0.3	<b>2.8</b>	<b>2.8</b>	<b>2.7</b>			± 0.5	<b>3.2</b>	<b>3.2</b>	<b>3.1</b>			± 1.3
		%	<b>64.0</b>	<b>64.0</b>	<b>62.5</b>			± 8.0	<b>35.0</b>	<b>35.0</b>	<b>34.0</b>			± 6.0	<b>16.0</b>	<b>16.0</b>	<b>15.5</b>			± 6.0
MONO		#	<b>0.2</b>	<b>0.2</b>	<b>0.2</b>			± 0.2	<b>0.7</b>	<b>0.7</b>	<b>0.7</b>			± 0.5	<b>1.2</b>	<b>1.2</b>	<b>1.1</b>			± 1.0
		%	<b>11.5</b>	<b>11.5</b>	<b>11.0</b>			± 7.0	<b>9.0</b>	<b>9.0</b>	<b>8.5</b>			± 5.0	<b>6.0</b>	<b>6.0</b>	<b>5.5</b>			± 5.0
GRANULO		#	<b>0.5</b>	<b>0.5</b>	<b>0.6</b>			± 0.2	<b>4.4</b>	<b>4.4</b>	<b>4.5</b>			± 0.6	<b>15.5</b>	<b>15.5</b>	<b>15.7</b>			± 1.5
		%	<b>24.5</b>	<b>24.5</b>	<b>26.5</b>			± 7.0	<b>56.0</b>	<b>56.0</b>	<b>57.5</b>			± 6.0	<b>78.0</b>	<b>78.0</b>	<b>79.0</b>			± 7.0
CRP		mg/dl	<b>0.78</b>	<b>0.62</b>	N/A			± 0.30	<b>2.40</b>	<b>2.21</b>	N/A			± 0.70	<b>4.13</b>	<b>4.01</b>	N/A			± 1.00
		mg/l	<b>7.8</b>	<b>6.2</b>	N/A			± 3.0	<b>24.0</b>	<b>22.1</b>	N/A			± 7.0	<b>41.3</b>	<b>40.1</b>	N/A			± 10.0

Ref: TEMP-1070 Rev.36 FRONT/RECTO 9930089-A