

ABX Minotrol CRP



LOT

MC 420

CONTROL



(Exp.) **2021-10-05**
(YYYY-MM-DD)

Rev 1

		ABX Lyse														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0			± 0.4	7.6	8.0			± 1.0	19.8	20.7			± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.33	2.36			± 0.15	4.55	4.59			± 0.18	5.48	5.56			± 0.25
HB HGB	g/dl	6.0	6.1			± 0.4	13.4	13.6			± 0.5	17.6	17.7			± 0.7
	g/l	60	61			± 4	134	136			± 5	176	177			± 7
HT HCT	mmol/l	3.73	3.79			± 0.25	8.32	8.45			± 0.31	10.93	10.99			± 0.44
	%	16.5	16.8			± 2.0	37.8	37.6			± 2.5	49.9	50.0			± 3.0
VGM MCV	l/l	0.165	0.168			± 0.020	0.378	0.376			± 0.025	0.499	0.500			± 0.030
	µm ³ ; fl	71	71			± 4	83	82			± 4	91	90			± 4
TGMH MCH	pg	25.8	25.8			± 2.4	29.5	29.6			± 2.4	32.1	31.8			± 2.8
	fmol	1.60	1.61			± 0.15	1.83	1.84			± 0.15	1.99	1.98			± 0.18
CCMH MCHC	g/dl	36.3	36.4			± 3.0	35.5	36.1			± 3.0	35.3	35.4			± 3.0
	g/l	363	364			± 30	355	361			± 30	353	354			± 30
	mmol/l	22.52	22.61			± 1.86	22.03	22.44			± 1.86	21.92	21.97			± 1.86
IDR RDW	%	14.0	14.0			± 3.0	13.5	13.5			± 3.0	13.0	13.0			± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	71	78			± 20	256	257			± 40	481	500			± 60
VPM MPV	µm ³ ; fl	9.3	8.2			± 2.0	8.5	7.9			± 2.0	8.1	7.6			± 2.0
LYMPHO	#	1.4	1.3			± 0.2	2.9	2.9			± 0.5	3.4	3.2			± 1.1
	%	69.5	66.5			± 8.0	38.5	36.5			± 6.0	17.0	15.5			± 5.0
MONO	#	0.2	0.2			± 0.2	0.5	0.5			± 0.4	1.2	1.0			± 1.0
	%	7.5	8.0			± 7.0	6.5	6.0			± 5.0	6.0	5.0			± 5.0
GRANULO	#	0.5	0.5			± 0.2	4.2	4.6			± 0.5	15.2	16.5			± 1.5
	%	23.0	25.5			± 7.0	55.0	57.5			± 6.0	77.0	79.5			± 7.0
CRP	mg/dl	0.62	N/A			± 0.30	2.38	N/A			± 0.70	3.95	N/A			± 1.00
	mg/l	6.2	N/A			± 3.0	23.8	N/A			± 7.0	39.5	N/A			± 10.0

Ref: TEMP-1070 Rev.39 FRONT/RECTO 1300092170