

ABX Minotrol CRP



LOT

MC 414

CONTROL



(Exp.) **2020-04-05**

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	1.8	1.9		± 0.4	7.3	7.8		± 1.0	19.0	20.5		± 2.2				
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.40	2.41		± 0.15	4.55	4.60		± 0.18	5.60	5.65		± 0.25				
HB HGB	g/dl	5.8	5.9		± 0.4	13.0	13.4		± 0.5	17.5	18.0		± 0.7				
	g/l	58	59		± 4	130	134		± 5	175	180		± 7				
HT HCT	mmol/l	3.60	3.66		± 0.25	8.09	8.32		± 0.31	10.87	11.18		± 0.44				
	%	16.8	16.9		± 2.0	36.9	37.3		± 2.5	49.8	50.3		± 3.0				
VGM MCV	l/l	0.168	0.169		± 0.020	0.369	0.373		± 0.025	0.498	0.503		± 0.030				
	µm ³ ; fl	70	70		± 4	81	81		± 4	89	89		± 4				
TGMH MCH	pg	24.1	24.5		± 2.4	28.6	29.1		± 2.4	31.3	31.9		± 2.8				
	fmol	1.50	1.52		± 0.15	1.78	1.81		± 0.15	1.94	1.98		± 0.18				
CCMH MCHC	g/dl	34.5	35.0		± 3.0	35.4	36.0		± 3.0	35.1	35.8		± 3.0				
	g/l	345	350		± 30	354	360		± 30	351	358		± 30				
	mmol/l	21.42	21.72		± 1.86	21.96	22.33		± 1.86	21.80	22.23		± 1.86				
IDR RDW	%	14.5	14.5		± 3.0	14.0	14.5		± 3.0	13.5	14.0		± 3.0				
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	77	75		± 20	263	265		± 40	550	565		± 60				
VPM MPV	µm ³ ; fl	8.3	9.0		± 2.0	7.9	8.5		± 2.0	7.9	8.6		± 2.0				
LYMPHO	#	1.2	1.2		± 0.2	2.6	2.8		± 0.5	3.4	3.5		± 1.3				
	%	68.6	65.5		± 8.0	35.7	36.0		± 6.0	17.7	17.0		± 6.0				
MONO	#	0.1	0.2		± 0.2	0.5	0.5		± 0.4	1.1	0.8		± 0.8				
	%	7.6	8.0		± 7.0	6.7	6.0		± 5.0	6.0	4.0		± 4.0				
GRANULO	#	0.4	0.5		± 0.2	4.2	4.5		± 0.5	14.6	16.2		± 1.5				
	%	24.0	26.5		± 7.0	57.8	58.0		± 6.0	76.7	79.0		± 7.0				
CRP	mg/dl	0.50	N/A		± 0.30	2.62	N/A		± 0.70	4.09	N/A		± 1.00				
	mg/l	5.0	N/A		± 3.0	26.2	N/A		± 7.0	40.9	N/A		± 10.0				

1300027267-A
Ref: TEMP-1070 Rev.38