

ABX Minotrol CRP



LOT

MC 419

CONTROL



(Exp.) 2021-07-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.1		± 0.4	7.5	7.6		± 1.0	19.8	20.7		± 2.2			
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.32	2.38		± 0.15	4.51	4.54		± 0.18	5.51	5.55		± 0.25			
HB	HGB	g/dl	5.6	5.9		± 0.4	12.8	13.3		± 0.5	17.8	18.3		± 0.7			
		g/l	56	59		± 4	128	133		± 5	178	183		± 7			
		mmol/l	3.48	3.66		± 0.25	7.97	8.26		± 0.31	11.05	11.36		± 0.44			
HT	HCT	%	16.2	16.9		± 2.0	36.5	37.2		± 2.5	50.1	51.1		± 3.0			
		l/l	0.162	0.169		± 0.020	0.365	0.372		± 0.025	0.501	0.511		± 0.030			
VGM	MCV	µm ³ ; fl	70	71		± 4	81	82		± 4	91	92		± 4			
TGMH	MCH	pg	24.2	24.8		± 2.4	28.5	29.3		± 2.4	32.3	33.0		± 2.8			
		fmol	1.50	1.54		± 0.15	1.77	1.82		± 0.15	2.01	2.05		± 0.18			
CCMH	MCHC	g/dl	34.5	34.9		± 3.0	35.1	35.7		± 3.0	35.5	35.8		± 3.0			
		g/l	345	349		± 30	351	357		± 30	355	358		± 30			
		mmol/l	21.43	21.68		± 1.86	21.82	22.19		± 1.86	22.05	22.26		± 1.86			
IDR	RDW	%	13.6	14.5		± 3.0	13.8	14.5		± 3.0	13.0	14.0		± 3.0			
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	67	75		± 20	254	258		± 40	510	515		± 60			
VPM	MPV	µm ³ ; fl	8.1	8.6		± 2.0	7.4	8.4		± 2.0	7.2	8.0		± 2.0			
LYMPHO		#	1.4	1.4		± 0.2	2.8	2.7		± 0.5	3.9	3.5		± 1.1			
		%	72.3	68.0		± 8.0	37.3	35.0		± 6.0	19.8	17.0		± 5.0			
MONO		#	0.2	0.2		± 0.2	0.5	0.5		± 0.4	1.0	1.0		± 1.0			
		%	5.6	6.5		± 5.6	7.1	6.5		± 5.0	5.0	5.0		± 5.0			
GRANULO		#	0.5	0.5		± 0.2	4.2	4.4		± 0.5	15.0	16.1		± 1.5			
		%	23.0	25.5		± 7.0	55.9	58.5		± 6.0	75.7	78.0		± 7.0			
CRP		mg/dl	0.55	N/A		± 0.30	2.20	N/A		± 0.70	4.10	N/A		± 1.00			
		mg/l	5.5	N/A		± 3.0	22.0	N/A		± 7.0	41.0	N/A		± 10.0			

1300027267-A
Ref: TEMP-1070 Rev.38