

ABX Minotrol CRP



LOT

MC 406

CONTROL



(Exp.) 2018-04-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.1		± 0.4	7.5	7.7		± 1.0	19.7	20.6		± 2.2			
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.28	2.27		± 0.15	4.72	4.74		± 0.18	5.67	5.72		± 0.25			
HB	HGB	g/dl	5.6	5.7		± 0.4	13.3	13.5		± 0.5	17.7	18.1		± 0.7			
		g/l	56	57		± 4	133	135		± 5	177	181		± 7			
HT	HCT	mmol/l	3.48	3.54		± 0.25	8.28	8.38		± 0.31	10.99	11.24		± 0.44			
		%	16.0	15.9		± 2.0	38.2	38.4		± 2.5	49.9	50.3		± 3.0			
		l/l	0.160	0.159		± 0.020	0.382	0.384		± 0.025	0.499	0.503		± 0.030			
VGM	MCV	µm ³ ; fl	70	70		± 4	81	81		± 4	88	88		± 4			
TGMH	MCH	pg	24.6	25.1		± 2.4	28.2	28.5		± 2.4	31.2	31.6		± 2.8			
		fmol	1.53	1.56		± 0.15	1.75	1.77		± 0.15	1.94	1.97		± 0.18			
CCMH	MCHC	g/dl	35.1	35.9		± 3.0	34.9	35.2		± 3.0	35.5	36.0		± 3.0			
		g/l	351	359		± 30	349	352		± 30	355	360		± 30			
		mmol/l	21.81	22.28		± 1.86	21.65	21.84		± 1.86	22.03	22.33		± 1.86			
IDR	RDW	%	14.0	14.0		± 3.0	13.5	13.5		± 3.0	13.5	13.5		± 3.0			
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	76	83		± 20	255	260		± 40	495	505		± 60			
VPM	MPV	µm ³ ; fl	8.3	8.5		± 2.0	8.5	8.8		± 2.0	7.8	8.0		± 2.0			
LYMPHO		#	1.4	1.4		± 0.2	2.8	2.8		± 0.5	3.5	3.5		± 1.3			
		%	66.6	64.5		± 8.0	37.8	37.0		± 6.0	17.7	17.0		± 6.0			
MONO		#	0.1	0.1		± 0.1	0.4	0.4		± 0.4	1.0	0.8		± 0.8			
		%	6.8	7.0		± 7.0	5.7	5.0		± 5.0	5.0	4.0		± 4.0			
GRANULO		#	0.6	0.6		± 0.2	4.3	4.5		± 0.5	15.3	16.3		± 1.5			
		%	27.0	28.5		± 7.0	56.8	58.0		± 6.0	77.7	79.0		± 7.0			
CRP		mg/dl	0.57	N/A		± 0.30	2.14	N/A		± 0.70	4.28	N/A		± 1.00			
		mg/l	5.7	N/A		± 3.0	21.4	N/A		± 7.0	42.8	N/A		± 10.0			

Ref: TEMP-1070 Rev.37 1300027267-A