

# ABX Minotrol CRP



**LOT**

**MC 404**

Rev 2

**CONTROL**



(Exp.) **2017-10-05**

(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Minilysebio												TOLERANCES TOLERANCE
			CONTROL			1	CONTROL			2	CONTROL			3	
			MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	
			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.1		± 0.4	7.4	7.7		± 1.0	19.4	20.2		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.43	2.43		± 0.15	4.77	4.78		± 0.18	5.76	5.78		± 0.25	
HB	HGB	g/dl	5.9	6.1		± 0.4	13.1	13.5		± 0.5	17.9	18.4		± 0.7	
		g/l	59	61		± 4	131	135		± 5	179	184		± 7	
HT	HCT	mmol/l	3.66	3.79		± 0.25	8.15	8.38		± 0.31	11.12	11.43		± 0.44	
		%	17.0	17.0		± 2.0	38.2	38.2		± 2.5	51.3	52.0		± 3.0	
VGM	MCV	l/l	0.170	0.170		± 0.020	0.382	0.382		± 0.025	0.513	0.520		± 0.030	
		μm <sup>3</sup> ; fl	70	70		± 4	80	80		± 4	89	90		± 4	
TGMH	MCH	pg	24.2	25.1		± 2.4	27.5	28.2		± 2.4	31.1	31.8		± 2.8	
		fmol	1.51	1.56		± 0.15	1.71	1.75		± 0.15	1.93	1.98		± 0.18	
CCMH	MCHC	g/dl	34.6	35.9		± 3.0	34.4	35.3		± 3.0	34.9	35.4		± 3.0	
		g/l	346	359		± 30	344	353		± 30	349	354		± 30	
IDR	RDW	mmol/l	21.50	22.27		± 1.86	21.37	21.92		± 1.86	21.68	21.97		± 1.86	
		%	14.0	13.2		± 3.0	14.5	14.0		± 3.0	13.5	13.2		± 3.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	79		± 20	250	253		± 40	485	500		± 60	
VPM	MPV	μm <sup>3</sup> ; fl	9.0	8.6		± 2.0	8.3	7.8		± 2.0	7.6	7.0		± 2.0	
LYMPHO		#	1.4	1.4		± 0.2	2.6	2.7		± 0.5	3.2	3.4		± 1.2	
		%	68.1	65.5		± 7.0	35.2	35.5		± 6.0	16.6	17.0		± 6.0	
MONO		#	0.1	0.1		± 0.1	0.6	0.5		± 0.5	1.1	0.9		± 0.9	
		%	6.8	7.0		± 6.0	8.1	6.0		± 5.0	5.5	4.5		± 4.5	
GRANULO		#	0.5	0.6		± 0.2	4.2	4.5		± 0.5	15.2	15.9		± 1.5	
		%	25.5	27.5		± 8.0	56.8	58.5		± 6.0	78.2	78.5		± 7.0	
CRP		mg/dl	0.54	N/A		± 0.30	2.32	N/A		± 0.70	4.78	N/A		± 1.00	
		mg/l	5.4	N/A		± 3.0	23.2	N/A		± 7.0	47.8	N/A		± 10.0	

Ref: TEMP-1070 Rev.37 1300027267-A