

ABX Minotrol CRP



LOT

MC 409

CONTROL



(Exp.) 2019-01-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.2		± 0.4	7.5	7.7		± 1.0	19.5	20.2		± 2.2				
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.29	2.31		± 0.15	4.63	4.65		± 0.18	5.85	5.83		± 0.25				
HB HGB	g/dl	5.8	5.9		± 0.4	13.1	13.5		± 0.5	18.1	18.5		± 0.7				
	g/l	58	59		± 4	131	135		± 5	181	185		± 7				
HT HCT	mmol/l	3.60	3.66		± 0.25	8.15	8.38		± 0.31	11.24	11.49		± 0.44				
	%	16.3	16.4		± 2.0	37.5	37.7		± 2.5	51.5	51.3		± 3.0				
VGM MCV	l/l	0.163	0.164		± 0.020	0.375	0.377		± 0.025	0.515	0.513		± 0.030				
	µm ³ ; fl	71	71		± 4	81	81		± 4	88	88		± 4				
TGMH MCH	pg	25.3	25.5		± 2.4	28.4	29.0		± 2.4	30.9	31.7		± 2.8				
	fmol	1.57	1.59		± 0.15	1.76	1.80		± 0.15	1.92	1.97		± 0.18				
CCMH MCHC	g/dl	35.6	36.0		± 3.0	35.0	35.8		± 3.0	35.2	36.1		± 3.0				
	g/l	356	360		± 30	350	358		± 30	352	361		± 30				
	mmol/l	22.13	22.34		± 1.86	21.74	22.26		± 1.86	21.83	22.39		± 1.86				
IDR RDW	%	13.5	14.0		± 3.0	13.5	14.5		± 3.0	13.0	13.5		± 3.0				
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	76	80		± 20	255	262		± 40	500	505		± 60				
VPM MPV	µm ³ ; fl	8.3	8.3		± 2.0	8.1	8.0		± 2.0	7.9	8.0		± 2.0				
LYMPHO	#	1.4	1.5		± 0.2	2.7	2.7		± 0.5	3.3	3.4		± 1.2				
	%	67.6	66.0		± 8.0	35.7	35.5		± 6.0	17.1	17.0		± 6.0				
MONO	#	0.2	0.2		± 0.2	0.6	0.5		± 0.4	1.1	1.0		± 1.0				
	%	7.2	8.0		± 7.0	7.6	6.5		± 5.0	5.5	5.0		± 5.0				
GRANULO	#	0.5	0.6		± 0.2	4.3	4.5		± 0.6	15.2	15.8		± 1.5				
	%	25.5	26.0		± 7.0	56.8	58.0		± 6.0	77.7	78.0		± 7.0				
CRP	mg/dl	0.55	N/A		± 0.30	2.70	N/A		± 0.70	4.89	N/A		± 1.00				
	mg/l	5.5	N/A		± 3.0	27.0	N/A		± 7.0	48.9	N/A		± 10.0				

Ref: TEMP-1070 Rev.37 1300027267-A