

ABX Minotrol CRP



LOT

MC 423

CONTROL



(Exp.) 2022-07-05
(YYYY-MM-DD)

Rev 2

		Minilysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS						MICROS						MICROS					
		CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.2			± 0.4	7.5	7.6			± 1.0	20.0	20.8			± 2.2			
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.22	2.19			± 0.15	4.53	4.50			± 0.18	5.67	5.64			± 0.25			
HB HGB	g/dl	5.6	5.7			± 0.4	13.0	13.3			± 0.5	18.0	18.5			± 0.7			
	g/l	56	57			± 4	130	133			± 5	180	185			± 7			
HT HCT	mmol/l	3.48	3.54			± 0.25	8.09	8.26			± 0.31	11.18	11.49			± 0.44			
	%	16.0	16.2			± 2.0	36.7	36.9			± 2.5	51.0	51.3			± 3.0			
VGM MCV	l/l	0.160	0.162			± 0.020	0.367	0.369			± 0.025	0.510	0.513			± 0.030			
	µm ³ ; fl	72	74			± 4	81	82			± 4	90	91			± 4			
TGMH MCH	pg	25.2	26.0			± 2.4	28.8	29.6			± 2.4	31.7	32.8			± 2.8			
	fmol	1.57	1.62			± 0.15	1.79	1.84			± 0.15	1.97	2.04			± 0.18			
CCMH MCHC	g/dl	35.1	35.2			± 3.0	35.5	36.0			± 3.0	35.3	36.0			± 3.0			
	g/l	351	352			± 30	355	360			± 30	353	360			± 30			
	mmol/l	21.78	21.84			± 1.86	22.06	22.38			± 1.86	21.90	22.38			± 1.86			
IDR RDW	%	13.5	14.0			± 3.0	13.0	14.0			± 3.0	12.5	13.0			± 3.0			
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	80	80			± 20	253	251			± 40	495	502			± 60			
VPM MPV	µm ³ ; fl	8.2	9.2			± 2.0	8.1	9.3			± 2.0	7.9	9.0			± 2.0			
LYMPHO	#	1.3	1.3			± 0.2	2.6	2.5			± 0.5	3.7	3.7			± 1.1			
	%	61.9	60.0			± 8.0	34.7	33.5			± 6.0	18.7	18.0			± 5.0			
MONO	#	0.2	0.2			± 0.2	0.7	0.6			± 0.4	1.4	1.1			± 1.0			
	%	10.8	11.0			± 7.0	9.0	8.0			± 5.0	7.0	5.5			± 5.0			
GRANULO	#	0.6	0.6			± 0.2	4.2	4.4			± 0.5	14.9	15.9			± 1.5			
	%	26.5	29.0			± 7.0	56.4	58.5			± 6.0	74.7	76.5			± 7.0			
CRP	mg/dl	0.42	N/A			± 0.30	2.66	N/A			± 0.70	4.16	N/A			± 1.00			
	mg/l	4.2	N/A			± 3.0	26.6	N/A			± 7.0	41.6	N/A			± 10.0			

1300027267-A
Ref: TEMP-1070 Rev.39