

ABX Minotrol CRP



LOT

MC 422

CONTROL



(Exp.) **2022-04-05**
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.2	2.1		± 0.4	8.6	8.7		± 1.0	20.6	21.6		± 2.2				
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.33	2.28		± 0.15	4.53	4.45		± 0.18	5.56	5.46		± 0.25				
HB HGB	g/dl	5.9	5.9		± 0.4	13.1	13.5		± 0.5	17.4	17.9		± 0.7				
	g/l	59	59		± 4	131	135		± 5	174	179		± 7				
	mmol/l	3.66	3.66		± 0.25	8.15	8.38		± 0.31	10.81	11.12		± 0.44				
HT HCT	%	16.8	16.9		± 2.0	37.1	37.4		± 2.5	49.5	50.2		± 3.0				
	l/l	0.168	0.169		± 0.020	0.371	0.374		± 0.025	0.495	0.502		± 0.030				
VGM MCV	µm ³ ; fl	72	74		± 4	82	84		± 4	89	92		± 4				
TGMH MCH	pg	25.3	25.9		± 2.4	29.0	30.3		± 2.4	31.3	32.8		± 2.8				
	fmol	1.57	1.61		± 0.15	1.80	1.88		± 0.15	1.94	2.04		± 0.18				
CCMH MCHC	g/dl	35.1	35.0		± 3.0	35.4	36.1		± 3.0	35.2	35.6		± 3.0				
	g/l	351	350		± 30	354	361		± 30	352	356		± 30				
	mmol/l	21.80	21.72		± 1.86	21.95	22.43		± 1.86	21.84	22.13		± 1.86				
IDR RDW	%	13.5	14.0		± 3.0	13.0	13.5		± 3.0	12.0	12.5		± 3.0				
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	79	72		± 20	268	269		± 40	499	495		± 60				
VPM MPV	µm ³ ; fl	8.5	9.6		± 2.0	8.0	8.9		± 2.0	8.1	9.0		± 2.0				
LYMPHO	#	1.5	1.4		± 0.2	3.0	2.9		± 0.5	3.7	3.8		± 1.1				
	%	69.2	66.5		± 8.0	35.2	33.5		± 6.0	18.2	17.5		± 5.0				
MONO	#	0.2	0.2		± 0.2	0.7	0.5		± 0.4	1.0	0.9		± 0.9				
	%	8.0	7.5		± 7.0	7.6	6.0		± 5.0	5.0	4.0		± 4.0				
GRANULO	#	0.5	0.5		± 0.2	4.9	5.3		± 0.5	15.9	17.0		± 1.5				
	%	23.0	26.0		± 7.0	57.3	60.5		± 6.0	77.2	78.5		± 7.0				
CRP	mg/dl	0.51	N/A		± 0.30	2.23	N/A		± 0.70	4.23	N/A		± 1.00				
	mg/l	5.1	N/A		± 3.0	22.3	N/A		± 7.0	42.3	N/A		± 10.0				

Ref: TEMP-1070 Rev.39 1300027267-A