

# ABX Minotrol CRP



**LOT**

**MC 412**

**CONTROL**



(Exp.) **2019-10-05**  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE													
				CONTROL				1				CONTROL						2				CONTROL				3			
				MICROS								MICROS								MICROS									
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P										
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l		<b>1.9</b>	<b>2.0</b>			± 0.4		<b>7.4</b>	<b>7.7</b>			± 1.0		<b>19.3</b>	<b>21.0</b>			± 2.2									
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l		<b>2.31</b>	<b>2.31</b>			± 0.15		<b>4.59</b>	<b>4.59</b>			± 0.18		<b>5.55</b>	<b>5.58</b>			± 0.25									
HB	HGB	g/dl		<b>5.7</b>	<b>5.8</b>			± 0.4		<b>13.0</b>	<b>13.2</b>			± 0.5		<b>17.6</b>	<b>18.2</b>			± 0.7									
		g/l		<b>57</b>	<b>58</b>			± 4		<b>130</b>	<b>132</b>			± 5		<b>176</b>	<b>182</b>			± 7									
		mmol/l		<b>3.54</b>	<b>3.60</b>			± 0.25		<b>8.09</b>	<b>8.20</b>			± 0.31		<b>10.93</b>	<b>11.30</b>			± 0.44									
HT	HCT	%		<b>16.2</b>	<b>16.2</b>			± 2.0		<b>36.7</b>	<b>36.7</b>			± 2.5		<b>49.4</b>	<b>50.2</b>			± 3.0									
		l/l		<b>0.162</b>	<b>0.162</b>			± 0.020		<b>0.367</b>	<b>0.367</b>			± 0.025		<b>0.494</b>	<b>0.502</b>			± 0.030									
VGM	MCV	µm <sup>3</sup> ; fl		<b>70</b>	<b>70</b>			± 4		<b>80</b>	<b>80</b>			± 4		<b>89</b>	<b>90</b>			± 4									
TGMH	MCH	pg		<b>24.7</b>	<b>25.1</b>			± 2.4		<b>28.4</b>	<b>28.8</b>			± 2.4		<b>31.7</b>	<b>32.6</b>			± 2.8									
		fmol		<b>1.53</b>	<b>1.56</b>			± 0.15		<b>1.76</b>	<b>1.79</b>			± 0.15		<b>1.97</b>	<b>2.03</b>			± 0.18									
CCMH	MCHC	g/dl		<b>35.3</b>	<b>35.9</b>			± 3.0		<b>35.5</b>	<b>35.9</b>			± 3.0		<b>35.6</b>	<b>36.2</b>			± 3.0									
		g/l		<b>353</b>	<b>359</b>			± 30		<b>355</b>	<b>359</b>			± 30		<b>356</b>	<b>362</b>			± 30									
		mmol/l		<b>21.89</b>	<b>22.27</b>			± 1.86		<b>22.04</b>	<b>22.32</b>			± 1.86		<b>22.13</b>	<b>22.51</b>			± 1.86									
IDR	RDW	%		<b>14.4</b>	<b>14.0</b>			± 3.0		<b>13.8</b>	<b>14.0</b>			± 3.0		<b>12.8</b>	<b>13.5</b>			± 3.0									
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l		<b>76</b>	<b>75</b>			± 20		<b>258</b>	<b>260</b>			± 40		<b>510</b>	<b>520</b>			± 60									
VPM	MPV	µm <sup>3</sup> ; fl		<b>8.1</b>	<b>8.9</b>			± 2.0		<b>7.9</b>	<b>8.5</b>			± 2.0		<b>7.7</b>	<b>8.4</b>			± 2.0									
LYMPHO		#		<b>1.3</b>	<b>1.3</b>			± 0.2		<b>2.9</b>	<b>2.8</b>			± 0.5		<b>3.6</b>	<b>3.8</b>			± 1.3									
		%		<b>68.6</b>	<b>65.5</b>			± 7.0		<b>39.4</b>	<b>36.5</b>			± 6.0		<b>18.7</b>	<b>18.0</b>			± 6.0									
MONO		#		<b>0.1</b>	<b>0.2</b>			± 0.2		<b>0.5</b>	<b>0.5</b>			± 0.4		<b>1.2</b>	<b>1.1</b>			± 0.8									
		%		<b>7.6</b>	<b>7.5</b>			± 6.0		<b>7.1</b>	<b>6.0</b>			± 5.0		<b>6.0</b>	<b>5.0</b>			± 4.0									
GRANULO		#		<b>0.5</b>	<b>0.5</b>			± 0.2		<b>4.0</b>	<b>4.4</b>			± 0.5		<b>14.6</b>	<b>16.2</b>			± 1.5									
		%		<b>24.0</b>	<b>27.0</b>			± 7.0		<b>53.9</b>	<b>57.5</b>			± 6.0		<b>75.7</b>	<b>77.0</b>			± 7.0									
CRP		mg/dl		<b>0.48</b>	N/A			± 0.30		<b>2.10</b>	N/A			± 0.70		<b>4.12</b>	N/A			± 1.00									
		mg/l		<b>4.8</b>	N/A			± 3.0		<b>21.0</b>	N/A			± 7.0		<b>41.2</b>	N/A			± 10.0									

Ref: TEMP-1070 Rev.38 1300027267-A