

ABX Minotrol CRP



LOT

MC 402

CONTROL



(Exp.) **2017-04-05**

(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	Minilysebio														
			CONTROL				1	CONTROL				2	CONTROL				3
			MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
			CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0			± 0.4	7.6	7.5			± 1.0	19.7	20.1			± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.42	2.40			± 0.15	4.72	4.68			± 0.18	5.65	5.58			± 0.25	
HB HGB	g/dl	6.0	5.9			± 0.4	13.4	13.6			± 0.5	17.8	17.8			± 0.7	
	g/l	60	59			± 4	134	136			± 5	178	178			± 7	
HT HCT	mmol/l	3.72	3.66			± 0.25	8.34	8.45			± 0.31	11.05	11.05			± 0.44	
	%	16.9	16.8			± 2.0	37.8	37.9			± 2.5	50.3	49.7			± 3.0	
VGM MCV	l/l	0.169	0.168			± 0.020	0.378	0.379			± 0.025	0.503	0.497			± 0.030	
	µm ³ ; fl	70	70			± 4	80	81			± 4	89	89			± 4	
TGMH MCH	pg	24.7	24.6			± 2.4	28.4	29.1			± 2.4	31.5	31.9			± 2.8	
	fmol	1.54	1.53			± 0.15	1.77	1.80			± 0.15	1.96	1.98			± 0.18	
CCMH MCHC	g/dl	35.3	35.1			± 3.0	35.6	35.9			± 3.0	35.4	35.8			± 3.0	
	g/l	353	351			± 30	356	359			± 30	354	358			± 30	
	mmol/l	21.94	21.81			± 1.86	22.08	22.28			± 1.86	21.98	22.26			± 1.86	
IDR RDW	%	13.8	13.7			± 3.0	13.2	14.4			± 3.0	12.8	12.9			± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	70	73			± 20	255	255			± 40	510	520			± 60	
VPM MPV	µm ³ ; fl	9.0	9.0			± 2.0	8.7	8.6			± 2.0	8.6	8.2			± 2.0	
LYMPHO	#	1.3	1.3			± 0.3	2.7	2.7			± 0.6	3.5	3.6			± 1.3	
	%	66.6	66.6			± 8.0	35.7	36.0			± 6.0	17.7	17.9			± 6.0	
MONO	#	0.2	0.2			± 0.2	0.5	0.5			± 0.5	1.1	0.9			± 0.9	
	%	8.0	8.4			± 7.0	6.7	6.0			± 5.0	5.5	4.6			± 4.6	
GRANULO	#	0.5	0.5			± 0.3	4.4	4.4			± 0.6	15.2	15.6			± 1.5	
	%	25.5	25.0			± 8.0	57.8	58.0			± 6.0	77.2	77.5			± 7.0	
CRP	mg/dl	0.54	N/A			± 0.30	2.27	N/A			± 0.70	4.28	N/A			± 1.00	
	mg/l	5.4	N/A			± 3.0	22.7	N/A			± 7.0	42.8	N/A			± 10.0	

Ref: TEMP-1070 Rev.37 1300027267-A