

ABX Minotrol CRP



LOT

MC 421

CONTROL



(Exp.) **2022-01-05**
(YYYY-MM-DD)

Rev 1

		Minilysebio															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.2	2.2			± 0.4	7.3	7.7			± 1.0	18.8	20.0			± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.29	2.26			± 0.15	4.46	4.48			± 0.18	5.58	5.61			± 0.25	
HB HGB	g/dl	5.7	5.9			± 0.4	13.0	13.5			± 0.5	17.8	18.5			± 0.7	
	g/l	57	59			± 4	130	135			± 5	178	185			± 7	
	mmol/l	3.54	3.66			± 0.25	8.09	8.38			± 0.31	11.05	11.49			± 0.44	
HT HCT	%	16.5	16.7			± 2.0	37.0	37.6			± 2.5	50.8	51.1			± 3.0	
	l/l	0.165	0.167			± 0.020	0.370	0.376			± 0.025	0.508	0.511			± 0.030	
VGM MCV	µm ³ ; fl	72	74			± 4	83	84			± 4	91	91			± 4	
TGMH MCH	pg	24.9	26.1			± 2.4	29.2	30.1			± 2.4	31.9	33.0			± 2.8	
	fmol	1.55	1.62			± 0.15	1.81	1.87			± 0.15	1.98	2.05			± 0.18	
CCMH MCHC	g/dl	34.6	35.3			± 3.0	35.2	35.9			± 3.0	35.1	36.2			± 3.0	
	g/l	346	353			± 30	352	359			± 30	351	362			± 30	
	mmol/l	21.47	21.91			± 1.86	21.87	22.28			± 1.86	21.77	22.50			± 1.86	
IDR RDW	%	13.5	14.0			± 3.0	14.0	14.0			± 3.0	13.5	13.5			± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	74	76			± 20	256	256			± 40	510	509			± 60	
VPM MPV	µm ³ ; fl	8.8	9.3			± 2.0	8.0	8.6			± 2.0	7.8	8.3			± 2.0	
LYMPHO	#	1.5	1.4			± 0.2	2.6	2.5			± 0.5	3.2	3.2			± 1.1	
	%	67.1	63.5			± 8.0	35.2	33.0			± 6.0	17.1	16.0			± 5.0	
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	0.9	0.9			± 0.9	
	%	8.0	8.5			± 7.0	8.1	7.0			± 5.0	5.0	4.5			± 4.5	
GRANULO	#	0.5	0.6			± 0.2	4.1	4.6			± 0.5	14.7	15.9			± 1.5	
	%	25.0	28.0			± 7.0	56.8	60.0			± 6.0	78.2	79.5			± 7.0	
CRP	mg/dl	0.56	N/A			± 0.30	2.15	N/A			± 0.70	3.40	N/A			± 1.00	
	mg/l	5.6	N/A			± 3.0	21.5	N/A			± 7.0	34.0	N/A			± 10.0	

1300027267-A
Ref: TEMP-1070 Rev.39