

ABX Minotrol CRP



LOT

MC 410

CONTROL



(Exp.) 2019-04-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio																
				CONTROL				1	CONTROL				2	CONTROL				3		
				MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P					
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l		2.0	2.0			± 0.4		7.6	7.8			± 1.0		19.5	20.5			± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l		2.37	2.33			± 0.15		4.75	4.72			± 0.18		5.73	5.70			± 0.25
HB	HGB	g/dl		5.8	5.8			± 0.4		13.2	13.5			± 0.5		18.0	18.5			± 0.7
		g/l		58	58			± 4		132	135			± 5		180	185			± 7
HT	HCT	mmol/l		3.60	3.60			± 0.25		8.22	8.38			± 0.31		11.18	11.49			± 0.44
		%		16.6	16.3			± 2.0		38.0	37.8			± 2.5		51.0	50.7			± 3.0
		l/l		0.166	0.163			± 0.020		0.380	0.378			± 0.025		0.510	0.507			± 0.030
VGM	MCV	µm ³ ; fl		70	70			± 4		80	80			± 4		89	89			± 4
TGMH	MCH	pg		24.5	24.9			± 2.4		27.9	28.6			± 2.4		31.4	32.5			± 2.8
		fmol		1.52	1.55			± 0.15		1.73	1.78			± 0.15		1.95	2.02			± 0.18
CCMH	MCHC	g/dl		34.9	35.6			± 3.0		34.8	35.8			± 3.0		35.3	36.5			± 3.0
		g/l		349	356			± 30		348	358			± 30		353	365			± 30
		mmol/l		21.69	22.08			± 1.86		21.62	22.20			± 1.86		21.92	22.65			± 1.86
IDR	RDW	%		13.0	14.0			± 3.0		13.5	14.0			± 3.0		13.0	13.5			± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l		72	76			± 20		265	271			± 40		515	523			± 60
VPM	MPV	µm ³ ; fl		8.4	8.3			± 2.0		8.4	8.4			± 2.0		8.0	8.0			± 2.0
LYMPHO		#		1.4	1.3			± 0.2		2.7	2.7			± 0.5		3.1	3.3			± 1.2
		%		67.6	65.5			± 8.0		35.7	35.0			± 6.0		16.1	16.0			± 6.0
MONO		#		0.2	0.2			± 0.2		0.6	0.5			± 0.4		1.2	1.0			± 1.0
		%		7.6	8.0			± 7.0		7.6	7.0			± 5.0		6.0	5.0			± 5.0
GRANULO		#		0.5	0.5			± 0.2		4.3	4.5			± 0.5		15.3	16.2			± 1.4
		%		25.0	26.5			± 7.0		56.8	58.0			± 6.0		78.2	79.0			± 7.0
CRP		mg/dl		0.52	N/A			± 0.30		1.95	N/A			± 0.70		3.57	N/A			± 1.00
		mg/l		5.2	N/A			± 3.0		19.5	N/A			± 7.0		35.7	N/A			± 10.0

1300027267-A
Ref: TEMP-1070 Rev.38