

ABX Minotrol CRP



LOT

MC 420

CONTROL



(Exp.) **2021-10-05**

(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE		
				CONTROL			1	CONTROL			2		CONTROL					3
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE			
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0		± 0.4	7.6	7.9			± 1.0	19.4	20.7			± 2.2			
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.33	2.32		± 0.15	4.55	4.55			± 0.18	5.48	5.52			± 0.25			
HB HGB	g/dl	5.7	5.8		± 0.4	13.1	13.5			± 0.5	17.6	17.9			± 0.7			
	g/l	57	58		± 4	131	135			± 5	176	179			± 7			
	mmol/l	3.54	3.60		± 0.25	8.15	8.38			± 0.31	10.93	11.12			± 0.44			
HT HCT	%	16.5	16.5		± 2.0	37.8	37.8			± 2.5	49.9	49.7			± 3.0			
	l/l	0.165	0.165		± 0.020	0.378	0.378			± 0.025	0.499	0.497			± 0.030			
VGM MCV	µm ³ ; fl	71	71		± 4	83	83			± 4	91	90			± 4			
TGMH MCH	pg	24.5	25.0		± 2.4	28.9	29.7			± 2.4	32.1	32.4			± 2.8			
	fmol	1.52	1.55		± 0.15	1.79	1.84			± 0.15	1.99	2.01			± 0.18			
CCMH MCHC	g/dl	34.5	35.2		± 3.0	34.8	35.7			± 3.0	35.3	36.0			± 3.0			
	g/l	345	352		± 30	348	357			± 30	353	360			± 30			
	mmol/l	21.40	21.87		± 1.86	21.59	22.20			± 1.86	21.92	22.38			± 1.86			
IDR RDW	%	14.0	14.5		± 3.0	13.5	14.0			± 3.0	13.0	13.5			± 3.0			
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	71	76		± 20	256	257			± 40	481	505			± 60			
VPM MPV	µm ³ ; fl	9.3	10.0		± 2.0	8.5	8.9			± 2.0	8.1	9.0			± 2.0			
LYMPHO	#	1.4	1.4		± 0.2	3.1	3.0			± 0.5	3.5	3.5			± 1.1			
	%	72.3	68.5		± 8.0	40.4	37.5			± 6.0	18.2	17.0			± 5.0			
MONO	#	0.1	0.2		± 0.2	0.5	0.4			± 0.4	1.2	0.9			± 0.9			
	%	6.0	5.5		± 5.5	6.2	5.0			± 5.0	6.0	4.5			± 4.5			
GRANULO	#	0.5	0.5		± 0.2	4.1	4.5			± 0.5	14.8	16.2			± 1.5			
	%	22.5	26.0		± 7.0	53.9	57.5			± 6.0	76.2	78.5			± 7.0			
CRP	mg/dl	0.62	N/A		± 0.30	2.38	N/A			± 0.70	3.95	N/A			± 1.00			
	mg/l	6.2	N/A		± 3.0	23.8	N/A			± 7.0	39.5	N/A			± 10.0			

1300027267-A
Ref: TEMP-1070 Rev.39